



PRODUCT DESCRIPTION

UR .53

Revision

1

Date

1.09.2021

DESCRIPTION

GENERAL INFORMATION

Product :	STANDART SURGICAL SET	Standard :	DIR 93/42/CE
Art. N° :	INT-3010-X1	Sterile :	Yes
Class :	1S	All items	LATEX-FREE

In accordance with EN 13795 ISO 811:2018

Products

Table Cover x 1

RAW MATEREAL

Weight	Color	Material	Use
56g/m2 (+/-2)	Blue	SMS & reinforced	Drape

FEATURES

Designation	Value	Tolerance
A Total width	190cm	+/-10cm
B Total width	150cm	+/-10cm



Mayo Cover x 1

RAW MATEREAL			
Weight	Color	Material	Use
50g/m2 (+/-2)	Blue	Reinforced & SMS	Mayo Cover


FEATURES		
Designation	Value	Tolerance
A Total width	80cm	+/-10cm
B Total width	140cm	+/-10cm



Absorbent Towel x2

RAW MATEREAL			
Weight	Color	Material	Use
52g/m2 (+/-2)	White	Absrb. Nonwoven	Towel


FEATURES		
Designation	Value	Tolerance
A Total width	30cm	+/-10cm
B Total width	35cm	+/-10cm



Bottom Drape x1

RAW MATEREAL			
Weight	Color	Material	Use
56g/m2 (+/-2)	Blue	SMS	Drape
70g/m2(+/-2)	White	Med. Adhesive	Adhesive

FEATURES		
Designation	Value	Tolerance
A Total width	190cm	+/-10cm
B Total width	195cm	+/-10cm
C Total width	130cm	+/-2cm




A=190 cm

T0p Drape x1

RAW MATEREAL			
Weight	Color	Material	Use
56g/m2 (+/-2)	Blue	SMMS	Drape
70g/m2(+/-2)	White	Med. Adhesive	Adhesive

FEATURES		
Designation	Value	Tolerance
A Total width	150cm	+/-10cm
B Total width	270cm	+/-10cm
C Total width	150cm	+/-10cm




Optape (Tape Strip) x1

RAW MATEREAL

Weight	Color	Material	Use
40g/m2 (+/-2)	Blue	SMS	Adhesive
35g/m2(+/-2)	White	Med. Adhesive	Adhesive

FEATURES

Designation	Value	Tolerance
A Total width	10cm	+/-5cm
B Total width	50cm	+/-5cm



- All the dimensions of the plain drape can be customized according to the desir of the customer. Available - custom dimensions, Reinforcement area, wrap.
- Intersect set is manufactured in accordance with the EN ISO 13485 standard. This ensures that we deliver products and services that consistently meet customer requirements, guaranteeing full customer satisfaction.
- All our products have a CE label and comply with the requirements of the Medical Device Directive 93/42/EEC

Manufacturer: **TOTAL İth. İhr. ve Paz. Ltd. Şti**