

## **Certificate of Compliance**

The following product has been tested by us with the listed standards and found in conformity with the council LVD directive 2006/95/EC. It is possible to use CE marking to demonstrate the conformity with this LVD Directive.

Certificate No. : 20150202

Applicant : Nanjing BW Optics and Instrument Co., Ltd

Address : No.219, Heyan Road, Nanjing, Jiangsu Province

Manufacturer : Nanjing BW Optics and Instrument Co., Ltd

Address : No.219, Heyan Road, Nanjing, Jiangsu Province

Trade Mark : N/A

**EUT** : BIOLOGICAL MICROSCOPE

M/N : MF606, MF608, MF406Y, MXD-400, MXD-300, MB220, MB270

MB510, MB520, MB522, MB540, MB620NS, MB802,

Technical Data : Input: 100-240V~, 50/60Hz, 2.0W

Test Standards : EN 61010-1: 2010



Approved by:\_

Laboratory Manager Issued date:2015-02-09

The certificate is based on a single evaluation of one sample of above-mentioned products. It does not imply an assessment of the whole production and does not permit the use of the test lab. logo.

