

Quality Assurance Certificate

Product: RIDASCREEN® Clostridium difficile GDH

REF	Art. No.:	C0701
LOT	Lot:	11087
	Expiry:	2018-11

Control		OD _{450 nm}	OD _{450/620 nm}	QC specification	
				OD _{450 nm}	OD _{450/620 nm}
Control	+	2.600	2.514	> 0.8	> 0.8
Control	-	0.062	0.001	< 0.200	< 0.160

Content	LOT	
Control	+	12037
Control	-	12336
Conjugate 1		14037
Conjugate 2		14037
Plate		13506
Stop		14067
Diluent		12047
Substrate		12066
Wash		11057
IFU (and other accompanying documents where applicable)		

R-Biopharm AG, Darmstadt, Germany, certifies that this batch has been approved by the Quality Control and conforms to specifications.

R-Biopharm AG
Quality Control

sign. N. Stork-Heininger
Head of Quality Control

Date: 2017-02-20

Remark: This document is created electronically and therefore valid without a signature.

Quality
Management
System certified
ISO 9001
ISO 13485
DQS, Germany

R-Biopharm AG
Postal address: An der neuen Bergstraße 17
64297 Darmstadt, Germany
Corporate seat: Pfungstadt
Phone: +49 (0) 61 51 - 81 02-0
Fax: +49 (0) 61 51 - 81 02-40
www.r-biopharm.com

**Chairman of supervisory
board:** Dietrich Mollat
Board of directors:
Dr. Ralf M. Dreher (CEO),
Dr. Carsten Bruns,
Jochen Hirsch,
Dr. Peter Schubert

Reg.No.:
Amtsgericht Darmstadt,
HRB 8321
Tax No.:
07242 / 02 888
VAT Reg. No.:
DE 111 657 409

