

CERTIFICATE OF ANALYSIS**Certificado de Análisis / Certificat de contrôle / Chargen-Prüfprotokoll
RealPCR CSFV RNA Mix [CSFV RNA Mix]****Product and lot number:** 99-56022 - 230196**Expiration date:** 11-FEB-2025

Number of reactions: 100

Manufacturing date: OCT-2023**The RealPCR CSFV RNA Mix needs to be used with the insert 06-56022-00.****The PC compatibility is ≥ 1.5 and IPC compatibility is ≥ 1.1 .**

Each RealPCR CSFV RNA Mix is tested with RealPCR RNA Master Mix on release panel (2 replicates of each panel member) and must meet the specifications detailed below.

Functional Test	Specification (mean Ct)	Result (Pass/ Fail)
CSFV 100 copy / reaction	< 35	Pass
CSFV 1000 copy / reaction	< 32	Pass
Internal Control	< 35	Pass
No template control	no Ct value	Pass

**This product was performance tested and has met all quality control specifications required for release.
This information is released by:**

Name: Rachel Mauran, Quality Specialist

Signature: