


Specific details and minimal technical standards

LOT I. Gloves, examination/ surgical, n/s:

Item / Description	Minimal requirements (Technical specifications and listed standards)		Offered specifications, manufacturer (name & country)
<p>Gloves, examination, n/s, pwd free, latex, size S, M, L piece</p> 	<p>Single use, ambidextrous examination gloves, made from natural latex, powder-free and non-sterile. Disposable, single-use. Available in sizes small (S), Medium (M) and large(L). Can come in several colors.</p> <p>Do not apply force when putting on, keep fingernails short when using.</p> <p>Dispose carefully after use. Packed in boxes of 100 gloves.</p>	<p>Minimum thickness 0.05mm, minimum 230mm total length, tested standard: EN 455.</p>	<p>Proposed model: DLOFBOG, manuf. SRI Trang Gloves (Thailand).</p> <p>Single use, ambidextrous examination gloves, made from natural latex, powder-free and non-sterile. – Compliance; Disposable, single-use. Available in sizes small (S), Medium (M) and large(L). – Compliance; Can come in several colors. – <i>latex gloves cannot be in several colors, only white.</i></p> <p>*Do not apply force when putting on, keep fingernails short when using. *Dispose carefully after use.</p> <p>Packed in boxes of 100 gloves. – Compliance.</p> <p>Minimum thickness 0.05mm, minimum 230mm total length – thickness 0.08 mm; length med. 240 mm. Tested standard: EN 455. – Compliance.</p>
<ul style="list-style-type: none"> • Manufacturer’s quality compliance ISO 9001 or ISO13485; • All presented certificates should be valid; • Original manufacturer packing for all items; • The Shelf life - not less than 70% of the total shelf life of the product on delivery; 			

- **All goods will be tested by the beneficiary.**
- **The goods will be accepted and the acts of receipt will be signed upon positive lab testing performed by beneficiary.**

NAME OF SUPPLIER S.C. IMUNOTEHNOMED S.R.L.

Authorized Signature and Stamp: General Manager, Sergiu RATA _____

Place: Chisinau

Date: 12.10.2022