#### *Prerequisite*

Action

- ☑ When using drapes from third-party manufacturers without RFID detection, the "AutoDrape" function (air extraction for easier draping) is not available. The magnetic fastening function cannot be used if the drape does not have a corresponding magnetic holder.
- ☑ ZEISS SMARTDRAPES (order number 306028-0000-000) are tested for use with navigation systems.
- 1. Unpack the SMARTDRAPE.
- 2. Pull the SMARTDRAPE over the surgical microscope.
- 3. Fasten the drape ring of the SMARTDRAPE to the lens below the surgical microscope.
  - ⇒ The drape ring with the magnet engages in the correct position (proper orientation of the beveled cover glass to prevent reflections).
- 4. Pull the SMARTDRAPE over the horizontal arm until it covers the marking.
  - When attaching the SMARTDRAPE, make sure that there is sufficient free space for swiveling, tilting and rotating movements of the surgical microscope.
  - Make sure the control elements can be safely operated through the SMARTDRAPE.
- 5. Seal the SMARTDRAPE airtight at the marking using one of the fastening straps.
- 6. Remove the RFID label from the SMARTDRAPE.
- 7. Stick the adhesive label on the RFID reader.



- ⇒ It can take up to 10 seconds to successfully read in the code.
- ⇒ A beep sounds as soon as the code has been successfully read in.
- ⇒ The AudoDrape function is available.
- 8. Open the additional menu "Extended operation" on the touch-screen.

#### 11.9 Fluorescence

| Designation                      | Order no. |
|----------------------------------|-----------|
| Fluorescence target INFRARED 800 | 30IR800T  |
| INFRARED 800 module              | 30INFCUC  |
| INFRARED 800 complete            | 30INCPUC  |
| Fluorescence target YELLOW 560   | 30YE560T  |
| YELLOW 560 module                | 30YELLUC  |
| YELLOW 560 complete              | 30YEXCPL  |
| YELLOW 560 LED module            | 30YLEDUC  |
| YELLOW 560 LED complete          | 30YELCPL  |

### 11.10 Foot control panel

| Designation                            | Order no. |
|--|-----------|
| Foot control panel, wireless (FCP WL)* | 10HFWLUC  |
| Foot control panel, wired (FCP)*       | 10HFWRUC  |

## 11.11 Sterile covers/drapes

| Designation                                      | Order no.       |
|--|-----------------|
| SMARTDRAPE, no. 28 (5 pcs), sterile              | 306028-0000-000 |
| VisionGuard replacement lenses (20 pcs), sterile | 306001-0000-000 |

# 11.12 Sterilizable caps

| Designation   | Order no. |
|---|-----------|
| Asepsis caps 12 mm (6 pcs), sterile                 | 60SETKAP  |
| Asepsis caps 22 mm (6 pcs), sterile                 |           |
| Asepsis caps PD 180° tiltable tube (6 pcs), sterile |           |