

Certificate of Analysis

Product name: BioPro RABIES ELISA Ab kit
Catalogue number: RAB 01-02
Kit batch number: R2308
Expiry date: 31-01- 2025
Kit components:

Quantity	Component	Batch
2 x 1	Rabies coated microplates	1R220219
1 x 0,6 ml	Rabies Positive control serum	2R230731
1 x 0,6 ml	Rabies Negative control serum	3R230731
1 x 0,3 ml	Rabies control serum 1	12R230728
1 x 0,3 ml	Rabies control serum 2	13R230728
1 x 0,3 ml	Rabies control serum 3	14R230728
1 x 15 ml	Sample diluent	4R230707
1 x 30 ml	Diluent for biotinylated antibody	5R230707
1 x 0,3 ml	Biotinylated anti-Rabies antibody (100x)	6R230808
1 x 30 ml	Diluent for Streptavidin peroxidase conjugate	7R230707
1 x 0,3 ml	Streptavidin peroxidase conjugate (100x)	8R230808
1 x 30 ml	TMB Substrate	9R230707
1 x 15 ml	Stop solution	10R230707
1 x 250 ml	Washing solution (10x)	11R230707
6 x 1	Adhesive foils	-

Kit control

Negative control serum OD : 1,390
 Positive control serum OD: 0,119
 Rabies control serum 1 PB: 60,6%
 Rabies control serum 2 PB: 36,1%
 Rabies control serum 3 PB: 8,7%
 Inter plate: 2,2%
 Intra plate: 2,0%

<u>Rabies Panel</u>	<u>result</u>	<u>expected result</u>
1	positive	positive
2	positive	positive
3	positive	positive
4	negative	negative / positive
5	negative	negative

Publication date: 15.9.2023

QM Signature: 