

Weighing scale RGZ-Series



Specifications:

	Model	RGZ-120	RGZ-160	RGZ-200
1	Max. weighing	120kg	160kg	200kg
2	Resolution	0.1kg/1lb	0.1kg/1lb	0.1kg/1lb
3	Height range	70cm-190cm	70cm-190cm	70cm-190cm
4	Min. value of height per division	0.5cm	0.5cm	0.5cm
5	Platform size(L*W)	375*270mm	375*270mm	375*270mm
6	Overall dim.(L*W*H)	695*286*930mm	695*286*930mm	695*286*930mm

Packing:

Dimension	91cm×34cm×29cm
Gross / Net Weight	12 / 11 KG
Package Type	1 Unit/ 1 CTN

1PC/CTN,
Carton size (cm) : 91 * 34 * 29;
GW. /NW.; 12kg/11kg



江苏苏宏医疗器械有限公司
JIANGSU SUHONG MEDICAL INSTRUMENTS CO.,LTD

**Conformity Declaration to the European Norms 93/42/CEE, 2007/47/CEE for
Medical Devices**

We, JIANGSU SUHONG MEDICAL INSTRUMENTS CO.,LTD, registred addres: Industrial Zone,
Xueyan Town, Wujin District,Changzhou .City Postcode: 213169, MANUFACTURER the following
items:

**Weighing scales: RGZ-10; RGZ-20; RGZ-50; RGZ-120; RGZ-150; RGZ-160;
RGZ-200 RGT-10; RGT -50; RGT -140; RGZ-160; RGT -200; RCS-20; RCS-200**

SH-8039 Stretcher (ISO13485)

DECLARE

that the item inside the boxes is included in Medical Devices Class I and REP no. 42689, and it is marked
CE and it is realized in accordance with the European Norms 93/42/CEE and 2007/47/CEE

Best regards,

Sunny

江苏苏宏医疗器械有限公司
JIANGSU SUHONG MEDICAL INSTRUMENTS CO., LTD

地址:江苏省常州市雪堰镇
Add:Xueyan,Changzhou,Jiangsu Province
电话:(Tel):0519-86155923 86153227 86153228
<http://www.czuhong.com>

邮编: 213169
Post Code:213169
传真(FAX):0519-86161228
E-mail:njk@czuhong.com



REGISTRATION NO. 04718Q10000118

CERTIFICATE OF QUALITY MANAGEMENT SYSTEM FOR MEDICAL DEVICES

This is to certify that the quality management system of

JIANGSU SUHONG MEDICAL INSTRUMENTS CO., LTD.

Registered Address: Industrial Zone, Xueyan Town, Wujin District

Postcode: 213100

Manufacturing Address: Industrial Zone, Xueyan Town, Wujin District

Has been assessed and conformed to the following standard(s)

YY/T 0287-2017 idt ISO 13485:2016

The certificate is valid for the following scope:

The Design, Development, Production and Service of visual acuity chart, stretcher, reflector lamp, x-ray film viewer, manual wheelchair, weighing scale, electronic weighing scale, electronic infant scale, infant height bed, electronic infant height bed, sitting height gauge, mobility scooter, walking aid, crutch, medical kit for home visit, mobile ultraviolet rays sterilization truck for design development, production and service.

Date of issue: March 23, 2018

Date of expiry: March 22, 2022

General Manager:

BEIJING HUA GUANG CERTIFICATION
OF MEDICAL DEVICES CO., LTD.