

## EU Declaration of Conformity for Cell Density Monitoring System

We Hamilton Bonaduz AG, CH-7402 Bonaduz/Switzerland confirm that the following products

| Product name                 | Ref     | Comment |
|------------------------------|---------|---------|
| Arc View 265 Controller      | 243800  |         |
| Arc View 465 Controller      | 243801  |         |
| Arc View 465 XL Controller   | 243802  |         |
| Incyte pre-amp               | 243720  |         |
| Dencytee pre-amp             | 243760  |         |
| 4-20 mA Output Box           | 243850  |         |
| Incyte Sensor Family         | various |         |
| Dencytee Sensor Family       | various |         |
| Cell Density Integration Kit | 243809  |         |
| Cell Density ComBox          | 243810  |         |

meet the following EU directives (including all applicable amendments):



|                       |            |
|-----------------------|------------|
| RoHS 2 Directive      | 2011/65/EU |
| EMC Directive         | 2014/30/EU |
| Low Voltage Directive | 2014/35/EU |

Applied company quality management system  
EN ISO 9001

Certification Body  
TÜV Rheinland Industrie Service GmbH  
Am Grauen Stein 29  
D-51105 Köln-Poll, Germany

### Applied standards:

|                        |                                                                                                                            |
|------------------------|----------------------------------------------------------------------------------------------------------------------------|
| RoHS 2 Directive:      | EN 50581:2012                                                                                                              |
| EMC Directive:         | EN 61326-1:2013<br>EN 61000-3-2:2006 + A1:2009 + A2:2009<br>EN 61000-3-3:2008<br>EN 61000-6-2 / EN 61000-6-4 (243809 only) |
| Low Voltage Directive: | EN 60950-1                                                                                                                 |

### Additional information:

All products listed in this document comply with CFR47 FCC Part 15.

Authorize department for  
technical documents:

HAMILTON Process Analytics

**Hamilton Bonaduz AG**

  
Andreas Wieland  
CEO

Bonaduz, 12.7.2017