



OneSpray ALCOHOL-BASED FAST ACTING SPRAY DISINFECTANT FOR MEDICAL EQUIPMENT

ALCOHOL-BASED FAST ACTING DISINFECTANT WIPES FOR OneSpray Wipes **MEDICAL DEVICES** 

OneSurface ALCOHOL-FREE FAST ACTING SPRAY DISINFECTANT FOR MEDICAL

**EQUIPMENT** 

**OneSurface Wipes** ALCOHOL- FREE FAST ACTING DISINFECTANT WIPES FOR

**MEDICAL DEVICES** 

OneSept GA HIGH-LEVEL DISINFECTANT FOR MEDICAL INSTRUMENTS AND

**ENDOSCOPES** 

OneSept PA HIGH-LEVEL DISINFECTANT FOR MEDICAL INSTRUMENTS AND

**ENDOSCOPES** 

HIGH-LEVEL DISINFECTANT FOR MEDICAL INSTRUMENTS AND OneSept OPA

**ENDOSCOPES** 

OneSept PA Powder HIGH-LEVEL DISINFECTANT FOR MEDICAL INSTRUMENTS AND ENDOSCOPES

**ENDOSCOPES** 

**OneSept A ALKALINE SOLUTION FOR WASHER AND DISINFECTORS** 

**OneSept FN NEUTRALIZER SOLUTION FOR WASHER AND DISINFECTORS** 

**OneSept ENZYM ENZYMATIC SOLUTION FOR MEDICAL INSTRUMENT WASHER** 

**ENZYMATIC CLEANER FOR MEDICAL INSTRUMENTS** OneSept ENZYM

AND ENDOSCOPES

## **OneSpray**

# ALCOHOL-BASED FAST ACTING SPRAY DISINFECTANT FOR MEDICAL EQUIPMENT

- ·Ethanol, propanol and quaternary ammonium based
- ·Wide spectrum of effects

### **CHEMICAL COMPOSITION**

%30 Ethanol, %10 2-Propanol, %0,25 Didecylmethylpoly(oxethyl) Ammonium Propionate, Auxiliary Substances, Essences, Deionized Water.

### **USAGE AREAS**

It is indicated for rapid disinfection of non-invasive medical devices, incubators, aerators used in dentistry and medical devices used in ENT examinations.



KAF G20 - 1	<b>750</b> ML
KAF G20 - 2	<b>1000</b> ML
KAF G20 - 3	5 L





	Microbiological Activity		
	Test Method	Microorganism	Effect Duration
		Staphylococcus aureus	
Bactericidal	EN 13727	Pseudomonas auruginosa	
	Enterococcus hirae	<b>30</b> sec	
Fungucidal / Yeas- ticidal EN 13624	Candida albicans	Tomas .	
	EN 13624	Aspergillus niger	
Tuberculocidal /	EN 14348	Mycobacterium avium	Samuel Min
Mycobactericidal EN 14348	Mycobacterium terrae	2 min	
		Poliovirus Type 1	Willy,
Virucidal	EN 14476	Human Adenovirus Type 5	1 min
		Murine Norovirüs	Same.

## **OneSpray Wipes**

# ALCOHOL-BASED FAST ACTING DISINFECTANT WIPES FOR MEDICAL DEVICES

- ·Ethanol, propanol and quaternary ammonium based
- ·Wide spectrum of effects

#### **CHEMICAL COMPOSITION**

%30 Ethanol, %10 2-Propanol, %0,25 Didecylmethylpoly(oxethyl) Ammonium Propionate, Auxiliary Substances, Essences, Deionized Water.

### **USAGE AREAS**

**OneSpray Wipes** is used in the disinfection of non-invasive, smooth, non-porous surface medical devices (Ultrasound probes, etc.) pre-cleaned, non-invasive disinfection by immersion method and non-invasive medical devices, incubators, aerators used in dentistry and ENT examinations. It is used for rapid disinfection of devices.



KAF G35 - 1	<b>100</b> PCS
KAF G35 - 2	<b>150</b> PCS
KAF G35 - 3	<b>200</b> PCS





Microbiological Activity			
Test Method	Microorganism	Effect Duration	
	Staphylococcus aureus		
EN 13727	Pseudomonas auruginosa		
	Enterococcus hirae	<b>30</b> sec	
EN137/O/	Candida albicans	Manual .	
ticidal Yeas-	Aspergillus niger		
/ 50.37.77.0	Mycobacterium avium	San Maria	
EN 14346	Mycobacterium terrae	5 min	
	Poliovirus Type 1		
EN 14476	Human Adenovirus Type 5		
	Murine Norovirus	1 min	
	Pseudomonas aeruginosa	The state of the s	
EN 16615	Staphylococcus aureus		
	Enterococcus hirae		
	EN 13727  EN 13624  EN 14348  EN 14476	Test Method  Microorganism  Staphylococcus aureus  Pseudomonas auruginosa  Enterococcus hirae  Candida albicans  Aspergillus niger  Mycobacterium avium  Mycobacterium terrae  Poliovirus Type 1  Human Adenovirus Type 5  Murine Norovirus  Pseudomonas aeruginosa  Staphylococcus aureus	

### **OneSurface**

### ALCOHOL-FREE FAST ACTING SPRAY DISINFECTANT FOR MEDICAL EQUIPMENT

- ·Hydrogen peroxide and quaternary ammonium based
- ·Wide spectrum of effects

### **CHEMICAL COMPOSITION**

0.84% Hydrogen Peroxide, 0.56% Alkyldimethylbenzylammonium Chloride, 0.49% Didecyldimethyl Ammonium Chloride, Auxiliary Substances, Deionized Water.

### **USAGE AREAS**

It is used in the rapid disinfection of aerators, non-invasive medical devices, incubators used in dentistry, medical devices with mandatory disinfection used in ENT examinations and materials that are not resistant to alcohol.



KAF G21 - 1	<b>750</b> ML
KAF G21 - 2	<b>1000</b> ML
KAF G21 - 3	<b>5</b> L





	Microbiological Activity		
	Test Method	Microorganism	Effect Duration
		Staphylococcus aureus	
Bactericidal	EN 13727	Pseudomonas auruginosa	
		Enterococcus hirae	
Fungucidal /		Candida albicans	
Yeasticidal		Aspergillus niger	
Tuberculocidal /	EN 14348	Mycobacterium avium	1 min
Mycobactericidal		Mycobacterium terrae	The same of the sa
	/irucidal EN 14476	Poliovirus Type 1	
Virucidal		Human Adenovirus Type 5	
		Murine Norovirus	
Sperioidal	EN 17126	Bacillus subtilis	
Sporicidal	EN 13704	Bacillus subtilis	

### **OneSurface Wipes**

# ALCOHOL- FREE FAST ACTING DISINFECTANT WIPES FOR MEDICAL DEVICES

- ·Hydrogen peroxide and quaternary ammonium based
- ·Wide spectrum of effects

#### **CHEMICAL COMPOSITION**

0.84% Hydrogen Peroxide, 0.56% Alkyldimethylbenzylammonium Chloride, 0.49% Didecyldimethyl Ammonium Chloride, Auxiliary Substances, Deionized Water.

#### **USAGE AREAS**

**OneSurface Wipes** is used for rapid disinfection of non-alcoholic and non-invasive medical devices used for rapid disinfection and cleaning on medical and medical device surfaces, incubators, aerators used in dentistry and medical devices with obligatory disinfection used in ENT examinations and materials that are not resistant to alcohol.



KAF G36 - 1	<b>100</b> PCS
KAF G36 - 2	<b>150</b> PCS
KAF G36 - 3	<b>200</b> PCS





	Test Method	Microorganism	Effect Duration
		Staphylococcus aureus	
Bactericidal	EN 13727	Pseudomonas auruginosa	
		Enterococcus hirae	
Fungucidal /	EN 13624	Candida albicans	
Yeasticidal	EN 13024	Aspergillus niger	
Tuberculocidal /	EN 14348	Mycobacterium avium	
Mycobactericidal	EN 14348	Mycobacterium terrae	<b>1</b> min
Virucidal EN 14476		Poliovirus Type 1	Manual .
	EN 14476	Human Adenovirus Type 5	
		Murine Norovirus	
Coorioidal	EN 17126	Bacillus subtilis	
Sporicidal EN 13704	Bacillus subtilis		
Bactericidal/ EN 166 Yeasticidal		Pseudomonas aeruginosa	
	EN 16615	Staphylococcus aureus	
		Enterococcus hirae	

## **OneSept GA**

### HIGH-LEVEL DISINFECTANT FOR MEDICAL INSTRUMENTS AND ENDOSCOPES

- ·Anticorrosive effective
- ·Suitable for manual usage and endoscope washing machines.
- ·Wide spectrum of effects

### **CHEMICAL COMPOSITION**

2% Glutaraldehyde, Anticorrosive Substances, Auxiliaries, Deionized Water.

### **USAGE AREAS**

Used to provide high-level disinfection of all kinds of surgical instruments, surgical laparoscopy instruments, flexible and rigid endoscopes.













after activation

Solution can be used for up to 28 days









	Test Method	Microorganism	Effect Duration
		Staphylococcus	
	EN 13727	Pseudomonas auruginosa	
Bactericidal		Enterococcus hirae	
240101101441		Staphylococcus	
	EN 14651	Pseudomonas auruginosa	
		Enterococcus hirae	
	FNIA	Mycobacterium avium	
Tuberculocidal / Mycobactericidal	EN 14563	Mycobacterium terrae	S. Carlotte
	EN 14348	Mycobacterium avium	5 min
	EN 14346	Mycobacterium terrae	
EN 14562	EN 1/5/2	Candida albicans	
	EN 14562	Aspergillus niger	
Yeasticidal	EN 13624	Candida albicans	
	EN 13624	Aspergillus niger	
		Poliovirus Type 1	
Virucidal	EN 14476	Human Adenovirus Type 5	
		Murine Norovirus	
Coorioidal	EN 17126	Bacillus subtilis	15 min
Sporicidal	EN 13704	Bacillus subtilis	19 min

## **OneSept PA**

### HIGH-LEVEL DISINFECTANT FOR MEDICAL INSTRUMENTS AND ENDOSCOPES

- ·Anticorrosive effective
- ·Suitable for manual usage and endoscope washing machines.
- ·Wide spectrum of effects

### **CHEMICAL COMPOSITION**

Peracetic Acid 2250 ppm, Anticorrosive Substances, Auxiliaries, Deionized Water.

### **USAGE AREAS**

Used to provide high level disinfection of interventional surgical instruments, dental surgical instruments, surgical laparoscopy instruments, endoscopes and heat sensitive instruments.









Peracetic Ready to use Acid after activation



Solution can be used for up to 14 days









	Test Method	Microorganism	Effect Duration
		Staphylococcus	
	EN 13727	Pseudomonas auruginosa	
Bactericidal		Enterococcus hirae	
Bactonolaai		Staphylococcus	
	EN 14651	Pseudomonas auruginosa	
		Enterococcus hirae	
	EN1-7-7-	Mycobacterium avium	
Tuberculocidal /	EN 14563	Mycobacterium terrae	Section .
Mycobactericidal	EN 1/7/0	Mycobacterium avium	5 min
	EN 14348	Mycobacterium terrae	
	EN 14562	Candida albicans	
Fungucidal /		Aspergillus niger	
Yeasticidal		Candida albicans	
EN 13624	EN 13624	Aspergillus niger	
		Poliovirus Type 1	
Virucidal	EN 14476	Human Adenovirus Type 5	
		Murine Norovirus	
Sporioidal	EN 17126	Bacillus subtilis	
Sporicidal	EN 13704	Bacillus subtilis	

## **OneSept OPA**

### HIGH-LEVEL DISINFECTANT FOR MEDICAL INSTRUMENTS AND ENDOSCOPES

- ·Anticorrosive effective
- ·Suitable for manual usage and endoscope washing machines.
- ·Wide spectrum of effects

### **CHEMICAL COMPOSITION**

0.55% Ortho-phthalaldehyde, Anticorrosive Substances, Auxiliaries, Deionized Water.

### **USAGE AREAS**

Used to provide high level disinfection of heat sensitive equipment, surgical instruments, flexible and rigid endoscopes.









Ortho-phthalaldehyde



Solution can be used for up to 14 days



Ready to use









	milotobiological Activity		
	Test Method	Microorganism	Effect Duration
		Staphylococcus	
	EN 13727	Pseudomonas auruginosa	
Bactericidal		Enterococcus hirae	
240101101441		Staphylococcus	
	EN 14651	Pseudomonas auruginosa	
		Enterococcus hirae	
	EN 14563	Mycobacterium avium	
Tuberculocidal /		Mycobacterium terrae	Section of
Mycobactericidal	EN 14348	Mycobacterium avium	5 min
	EN 14546	Mycobacterium terrae	
	EN 14562	Candida albicans	
Fungucidal /		Aspergillus niger	
Yeasticidal	EN 13624	Candida albicans	
EN 13624	Aspergillus niger		
		Poliovirus Type 1	
Virucidal	EN 14476	Human Adenovirus Type 5	
		Murine Norovirus	
Sporicidal	EN 17126	Bacillus subtilis	
EN 13704	EN 13704	Bacillus subtilis	

## **OneSept PA POWDER**

### HIGH-LEVEL DISINFECTANT FOR MEDICAL INSTRUMENTS AND ENDOSCOPES

- ·Anticorrosive effective
- ·Suitable for manual usage and endoscope washing machines.
- ·It performs cleaning and disinfection at the same time.

### **CHEMICAL COMPOSITION**

Peracetic Acid (solution diluted with 2% water) 2250 ppm, Sodium Percarbonate, TAED, Anti-Corrosion, Stabilizer, Auxiliaries.

#### **USAGE AREAS**

Used to provide high-level disinfection of endoscopes, arthroscopes, surgical and dental instruments.



KAF G26 - 1	<b>2</b> KG
KAF G26 - 2	<b>2,5</b> KG
KAF G26 - 3	<b>5</b> KG







	Test Method	Microorganism	Effect Duration
Bactericidal	EN 13727	Staphylococcus	
		Pseudomonas auruginosa	
		Enterococcus hirae	
	EN 14651	Staphylococcus	
		Pseudomonas auruginosa	
		Enterococcus hirae	10 min
	EN 14563	Mycobacterium avium	
Tuberculocidal /		Mycobacterium terrae	
Mycobactericidal	EN 14348	Mycobacterium avium	
		Mycobacterium terrae	
	EN 14562	Candida albicans	
Fungucidal / Yeasticidal		Aspergillus niger	
	EN 13624	Candida albicans	
		Aspergillus niger	
Virucidal	EN 14476	Poliovirus Type 1	
		Human Adenovirus Type 5	
		Murine Norovirus	
Sporicidal	EN 17126	Bacillus subtilis	
	EN 13704	Bacillus subtilis	

### OneSept A

### ALKALINE SOLUTION FOR WASHER AND DISINFECTORS



- •It is an alkaline cleaner for cleaning and disinfection of medical devices with washers thermal disinfectors.
- •Provides in depth cleaning without any residues.
- •Suitable for use in different degrees of water hardness.

### **USAGE AREAS**

Used in washing and thermal disinfection machines for cleaning and thermal disinfection of heat resistant medical instruments and materials including laboratory equipments at 93°C and cleaning of non-heat resistant medical instruments and materials at 50-60°C.

#### **USAGE**

0.1% - 0.5% (1-5 ml/lt) is recommended for cleaning and thermal disinfection devices in the medical field. Follow the machine manufacturer's recommendations for dosing information.

#### COMPOSITION

Potassium hydroxide, complexing agents, corrosion inhibitors.



KAF G34

5 L

KAF G33

**5** I

## OneSept FN

### **NEUTRALIZER SOLUTION FOR WASHER AND DISINFECTORS**



- ·Liquid, organic acid-containing neutralization solution for washer-disinfectors
- ·Neutralization of alkaline residues from the main washing process
- •Prevents limescale formation
- ·Dissolves albumins, organic bases, carbonates and hydroxides during cleaning
- ·Suitable for use in all washers and disinfectors

### **USAGE AREAS**

Used for the removal of lime residues in the machine and for the neutralization of alkali residues resulting from the main washing process. (40-60 °C).

### **USAGE**

For cleaning and disinfection devices in the medical field, 0.05%-0.3% (0.5-3 ml/lt) is recommended. Follow the machine manufacturer's recommendations for dosing information.

#### COMPOSITION

Phosphoric acid, corrosion inhibitors and auxiliary substances.

### OneSept ENZYM

### **ENZYMATIC SOLUTION FOR MEDICAL INSTRUMENT WASHER**



Concentrate tool cleaning detergent solution containing developed triple enzyme complex for use in automatic washing machines.

### **USAGE AREAS**

Suitable for cleaning of dirt surgical and medical instruments at sterilization unit with instrument washing machines.

#### **USAGE**

0.2%-0.8% (2-8 ml/lt) is recommended for cleaning and disinfection devices in the medical field. Follow the machine manufacturer's recommendations for dosing information.

### COMPOSITION

%1-%5 (protease, amylase, lipase) glycerin, monopropylene glycol, nonionic surfactants, complexing agents anticorrosive substances.









KAF G32


5 I

KAF G40

5 L

## OneSept ENZYM

# ENZYMATIC CLEANER FOR MEDICAL INSTRUMENTS AND ENDOSCOPES



·Anticorrosive effective



### **USAGE AREAS**

- •Suitable for endoscope washing machines, ultrasonic washing machines and manual use. It provides the cleaning and pre-disinfection of all kinds of invasive surgical instruments, dental surgical instruments, surgical laparoscopy intruments, flexible and rigid endoscopes.
- •Provides a more effective disinfection process by cleaning organic wastes on the surfaces such as blood, fat, protein and starch. Compatible with glass, rubber, wood, ceramic, silicone, plastic and stainless steel.

#### **USAGE**

It is used by diluting in % 0,5 concentration. All instruments are immersed into the solution. Shows maximum activity at 20- 45 °C. It is effective in 5 minutes in manual and ultrasonic usage. The instruments, have been removed from the solution at the end of the process, are rinsed thoroughly.

### COMPOSITION

%1-%5 (protease, amylase, lipase) glycerin, monopropylene glycol, nonionic surfactants, complexing agents anticorrosive substances.

	Usa	Usage Area	Pr	Product type			Efficacy spectrum	ectrum	
	End + Sur. Ins.	Dental Units, Incubators and ENT	Ready after activation	Ready for use	Concen- trate	Bactericidal	Fungucidal	Virucidal	Sporicidal
GA	>		>			>	>	>	>
РА	>		<b>&gt;</b>			>	<b>&gt;</b>		<b>/</b>
OPA	>			>		>	>	>	>
PA POWDER	>				>	>	>	>	>
OneSpray		>		>		>	>	>	
OneSurface		>		>		>	<b>/</b>	<b>&gt;</b>	<b>&gt;</b>
OneSpray Wipes		>		>		>	<	>	
OneSurface Wipes		>		>		>	>	>	>

\*End.: Endoscopy
\*Sur. Ins.: Surgical instruments

//

