



Date: 22.12.2021

Ref nr. ocds-b3wdp1-MD-1637681843608 (21047425)

Attn: IP Office of Management of External Assistance Programs

WHEREAS, *MED Birleşim Medikal San. ve Tic. Ltd. Şti.*, who are established and reputable manufacturers of Containerized Oxygen Generating Systems and Cylinder Manifold Systems having factories at Perpa İş Merkezi B. Blok Kat 5 No:155 Okmeydanı Şişli İstanbul

Do hereby declare that, for the systems, which we offered,

- The capacity of the compressor and the working pressure required for the oxygen station provided are as follows

O2 Generator Capacity:

10 m3/h

16 m3/h

24 m3/h

30 m3/h

34 m3/h

Air Compressor Capacity:

18,5KW – 3,10 m3/min @ 8 Bar / 2,70 m3/min @ 10 Bar

30KW - 5,30 m3/min @ 8 Bar / 4,81 m3/min @ 10 Bar

37KW – 6,37 m3/min @ 8 Bar / 5,90 m3/min @ 10 Bar

45KW – 7,65 m3/min @ 8 Bar / 7,15 m3/min @ 10 Bar

55KW – 9,33 m3/min @ 8 Bar / 8,76 m3/min @ 10 Bar

Murat Kaftancı
Export Manager
Med Birleşim Ltd.

MED Birleşim
Medikal Sanayi ve Tic. Ltd. Şti.
Perpa Tic. Mrk. B Blok Kat: 5 No: 155
Şişli / İSTANBUL Tel: 00 90 212 220 21 22
Şişli T.C. 6141 305 861

On Behalf of GM