

# CERTIFICATE OF ANALYSIS

## PrioCHECK® FMDV NS 5 strip plates

Product ID

7610770

Lot No

F200101L

Expiry Date

31-Jan-2022

Prionics performed the QC testing in compliance with valid instructions and states the following results:

### Test Components

#	Name	Art. No.	Lot No	#	Name	Art. No.	Lot No
1	Test Plate	7610451	F200101L201	7	Positive Control	7610455	F200101L220
2	Conjugate	7610456	F200101L212	8	Weak Positive Control	7610454	F200101L223
3	Dilution Buffer FMDV NS	7610446	F200101L248	9	Negative Control	7610453	F200101L224
4	Additive	7610445	F200101L250	10	Chromogen (TMB) Substrate	7588925	F200101L214
5	Demineralized Water	7610303	F200101L218	11	Stop Solution	7610302	F200101L216
6	Washing Fluid	7610301	F200101L238				

Technical inserts Version numbers 1.0\_e, 1.2\_d, 1.0\_es, 1.0\_f

### Final QC (Control number: 20Q-020)

Test	Specification	Result
OD450 max.	≥ 1.000	1.829
PI weak positive control	> 50%	71%
PI positive control	> 70%	91%

### Release by Prionics AG

The above mentioned PrioCHECK® FMDV NS lot fulfils all required batch release criteria and is

RELEASED

QA manager

S. Heerkens

Date:

12-mar-2020

Signature:

