

Test Report No. 7191252182-EEC21/01-WBH
dated 26 Jan 2021



PSB Singapore

**Add value.
Inspire trust.**

Note: This report is issued subject to the Testing and Certification Regulations of the TÜV SÜD Group and the General Terms and Conditions of Business of TÜV SÜD PSB Pte Ltd. In addition, this report is governed by the terms set out within this report.

SUBJECT:

Testing of Examination Gloves submitted by Fitone Latex Products Co.,Ltd.
Guangdong on 14 Jan 2021.

TESTED FOR:

Fitone Latex Products Co.,Ltd. Guangdong
No.5 Tongyi Road, Lingbei Industrial Zone, Suixi,
524338 Zhanjiang, Guangdong,
China

TEST DATE:

20 Jan 2021

DESCRIPTION OF SAMPLES:

S/N	Product Description	Brand/ Model	Size	Colour	Lot No.	Expiry Date	Sample Received	Manufacturer
1	Single-Use Medical Latex Examination Gloves	- (see Remark 1)	M	Natural	SE201201	11/2023	160 pairs (320 pcs)	Fitone Latex Products Co.,Ltd. Guangdong

Lot size as specified by client: 150,001 to 500,000 pieces

METHOD OF TEST:

EN 455-1:2020 Medical gloves for single use
Part 1: Requirements and testing for freedom from holes



Laboratory:
TÜV SÜD PSB Pte. Ltd.
TÜV SÜD @ IBP
15 International Business Park
Singapore 609937

Phone : +65-6778 7777
E-mail: info.sg@tuvsud.com
<https://www.tuvsud.com/en-sg>
Co. Reg : 199002667R

Regional Head Office:
TÜV SÜD Asia Pacific Pte. Ltd.
TÜV SÜD @ IBP
15 International Business Park
Singapore 609937
TÜV®

RESULTS:

Sample: Single-Use Medical Latex Examination Gloves, Natural, Size M

Table: Results for EN 455-1:2020

Clause	Tests	Requirements	No. of non-compliers allowed (pieces)	Number tested (pieces)	Actual no. of non-compliers found (pieces)	Inferred results
4 5	Freedom from holes	Shall not leak	10	315	2	Passed

REMARKS:

1. The Brand/Model was not provided by client.



Yeo Poh Kwang
Associate Engineer



Wong Bee Hui
Product Manager
Medical Health Services (NAM)

APPENDIX:



Photo: Single-Use Medical Latex Examination Gloves, Natural, Size M

Test Report No. 7191252182-EEC21/01-WBH
dated 26 Jan 2021



Please note that this Report is issued under the following terms :

1. This report applies to the sample of the specific product/equipment given at the time of its testing/calibration. The results are not used to indicate or imply that they are applicable to other similar items. In addition, such results must not be used to indicate or imply that TÜV SÜD PSB approves, recommends or endorses the manufacturer, supplier or user of such product/equipment, or that TÜV SÜD PSB in any way "guarantees" the later performance of the product/equipment. Unless otherwise stated in this report, no tests were conducted to determine long term effects of using the specific product/equipment.
2. The sample/s mentioned in this report is/are submitted/supplied/manufactured by the Client. TÜV SÜD PSB therefore assumes no responsibility for the accuracy of information on the brand name, model number, origin of manufacture, consignment or any information supplied.
3. Nothing in this report shall be interpreted to mean that TÜV SÜD PSB has verified or ascertained any endorsement or marks from any other testing authority or bodies that may be found on that sample.
4. This report shall not be reproduced wholly or in parts and no reference shall be made by the Client to TÜV SÜD PSB or to the report or results furnished by TÜV SÜD PSB in any advertisements or sales promotion.
5. Unless otherwise stated, the tests were carried out in TÜV SÜD PSB Pte Ltd, 15 International Business Park Singapore 609937.
6. The tests carried out by TÜV SÜD PSB and this report are subject to TÜV SÜD PSB's General Terms and Conditions of Business and the Testing and Certification Regulations of the TÜV SÜD Group.

Effective 01 January 2021

