



Bio-Rad  
Certificate of Analysis

---

Material Description: Mueller-Hinton Agar 500g  
Material Number: 64884  
Batch Number: 64589440  
Manufacture Date: 2024-01-17  
Expiration Date: 2029-01-15

---

<b>Characteristic</b>	<b>Specification</b>		<b>Results</b>
	<b>Lower Limit</b>	<b>Upper Limit</b>	
Labeling	Specifications see above		Conforms
Appearance Dehydrated powder	Beige		Conforms
Appearance Prepared medium	Opalescent, amber		Conforms
pH	7.2	7.4	7.2



Bio-Rad  
Certificate of Analysis

Material Description: Mueller-Hinton Agar 500g  
Material Number: 64884  
Batch Number: 64589440  
Manufacture Date: 2024-01-17  
Expiration Date: 2029-01-15

<u>Characteristic</u>	<u>Specification</u>		<u>Results</u>
	<u>Lower Limit</u>	<u>Upper Limit</u>	
..... Antimicrobial Susceptibility Testings .....			
Escherichia coli ATCC_25922			Conforms
	Amoxicillin + Clavulanic acid 20/10 µg Ampicillin 10 µg Cefotaxime 5 µg Chloramphenicol 30 µg Ciprofloxacin 5 µg Gentamicin 10 µg Tetracycline 30 µg Tigecycline 15 µg Trimethoprim + Sulfamethoxazole 1,25/23,75µg		
Staphylococcus aureus ATCC_29213			Conforms
	Benzylpenicillin 1 IU Cefoxitin 30 µg Ciprofloxacin 5 µg Erythromycin 15 µg Gentamicin 10 µg Linezolid 10 µg Tetracycline 30 µg		
Pseudomonas aeruginosa ATCC_27853			Conforms
	Aztreonam 30 µg Ceftazidime 10 µg Ciprofloxacin 5 µg Gentamicin 10 µg Imipenem 10 µg Piperacillin + Tazobactam 30/6 µg Tobramycin 10 µg		
Enterococcus faecalis ATCC_29212			Conforms
	Ampicillin 2 µg Imipenem 10 µg Linezolid 10 µg Nitrofurantoin 100 µg Trimethoprim 5 µg Trimethoprim + Sulfamethoxazole 1,25/23,75µg Vancomycin 5 µg		



Bio-Rad  
Certificate of Analysis

---

Material Description: Mueller-Hinton Agar 500g  
Material Number: 64884  
Batch Number: 64589440  
Manufacture Date: 2024-01-17  
Expiration Date: 2029-01-15

---

<b>Characteristic</b>	<b>Specification</b>		<b>Results</b>
	<b>Lower Limit</b>	<b>Upper Limit</b>	
Staphylococcus aureus ATCC_43300			Conforms
	Cefoxitin 30 µg		

This lot of dehydrated Mueller Hinton agar has been tested according to ISO 16782 and meets Quality Control criteria as stated in ISO 16782.

Hélène Bruxelles

---

This certificate has been verified and electronically approved by an authorized Quality representative. This e-signature is performed within a secure system.

2024-02-12 / 15:11:47 UTC

---

Date/Time