



# Attestation of Compliance

Reference No. : LCS200817083AE

Applicant : Shenzhen Kenid Medical Devices CO., LTD

Address : Room 201, 2nd Floor, Building 3, Aojia Industrial Park, 33 Pingan Avenue, Xinnan Village, Longgang District, Shenzhen, China 518111

Trade Mark : KENID

Product : Dry imager

Model(s) : KENID plus, KND6320

Tested according to : EN 60601-1-2: 2015  
EN 61000-3-2: 2019, EN 61000-3-3: 2013/A1: 2019

The submitted products have been tested by us with the listed standards.

This Attestation of Compliance is issued according to the council Directive 2014/30/EU, Referred to as the Electromagnetic Compatibility. It confirms that the listed product complies with all essential requirements of the EMC directive and applies only to the sample and its technical documentation submitted to Shenzhen LCS Compliance Testing Laboratory Ltd. for testing.

After preparation of the necessary technical documentation as well as the EC conformity declaration the required CE marking can be affixed on the product. Other relevant Directives have to be observed.



Date of issue: September 17, 2020



Shenzhen LCS Compliance Testing Laboratory Ltd.  
Room 101, 201, Building A and Room 301, Building C, Juji Industrial Park, Yabianxueziwei, Shajing Street,  
Bao'an District, Shenzhen, Guangdong, China  
Tel: (86)0755-82591330 Fax: (86)0755-82591332  
Http://www.lcs-cert.com Email: webmaster@lcs-cert.com

扫码查询真伪  
Scan, Query authenticity