

Quality Certificate

for following calibration standards:

Standard	Art.- No.	Pc(s).	Lot No.	Date of production	Date of expiry
"A" 0,000° C	7165	-	without No.	-	See note
"B" 0,557° C	7166	2	150134-150135	14.07.25	See note
"A" 0,408° C	7186	-	-	-	See note
"B" 0,600° C	7187	-	-	-	See note
"C" 0,512° C	7188	2	153271-153272	14.07.25	See note

All solutions are manufactured according the following standards:
DIN EN ISO 5764 / ISO / FDIS 5764:2002(E) IDF 108:2002 (E)

Note / Expiry-Date

1. storage

„Unopened bottles stored between 2°C and 7°C are usable for a period of 6 months.

2. usage

„Pour samples of standard solutions from the bottles. Never use pipettes for this purpose. Do not use standard solutions from bottles less than one quarter full or more than two months old or containing visible moulds.

See standard 5764 „5.3 Sodium chloride standard solutions“

Berlin, 17. Juli 2025

Funke - Dr. N. Gerber Labortechnik GmbH



Funke-Dr. N. Gerber
Labortechnik GmbH

i.A. J. Kehnscharper
- Despatch Dept. -

FUNKE GERBER

Funke - Dr. N. Gerber Labortechnik GmbH
Ringstraße 42
12105 Berlin

Annex:
Quality Certificate
for following calibration standards:

Standard	Art-No	Tolerance	Min.- Value	Max.- Value
„A“ 0,000 °C	7165	$\leq \pm 0,3 \text{ m}^\circ\text{C}$	-0,0003 °C	+0,0003 °C
„B“ 0,557 °C	7166	$\leq \pm 0,5 \text{ m}^\circ\text{C}$	-0,5575 °C	-0,5565 °C
„A“ 0,408 °C	7186	$\leq \pm 0,5 \text{ m}^\circ\text{C}$	-0,4085 °C	-0,4075 °C
„B“ 0,600 °C	7187	$\leq \pm 0,5 \text{ m}^\circ\text{C}$	-0,6005 °C	-0,5995 °C
„C“ 0,512 °C	7188	$\leq \pm 0,5 \text{ m}^\circ\text{C}$	-0,5125 °C	-0,5115 °C

All solutions are manufactured according the following standards:

DIN EN ISO 5764
ISO/FDIS 5764:2002(E)
IDF 108:2002(E)

Berlin, 17. Juli 2025

Funke - Dr. N. Gerber Labortechnik GmbH