

DECLARATION OF CONFORMITY TO Regulation (EU) 2017/745 CONCERNING MEDICAL DEVICES

MANUFACTURER: Edan Instruments, Inc.

#15 Jinhui Road, Jinsha Community, Kengzi Sub-District, Pingshan District, 518122 Shenzhen, P.R.China

SRN: CN-MF-000009957

EUROPEAN REPRESENTATIVE: Shanghai International Holding Corp. GmbH

Eiffestrasse 80 20537 Hamburg Germany

PRODUCT/MODEL: Ambulatory Blood Pressure Monitor /

SA-10,SA-05,SA-06,SA-08 and SA-09

EMDN [NAME/CODE]: BLOOD PRESSURE HOLTER RECORDERS / Z12050404

Basic UDI-DI: 69444138SAS4S

CLASSIFICATION: Class IIa, Rule 10 According To Annex VIII of the MDR

CONFORMITY ASSESSMENT ROUTE: ANNEX IX EXCLUDING CHAPTER II.

WE, EDAN INSTRUMENTS, INC., HERE WITH DECLARE THAT THE ABOVE MENTIONED PRODUCT(S) MEET THE PROVISIONS OF REGULATION (EU) 2017/745 OF THE EUROPEAN PARLIAMENT AND OF THE COUNCIL ON MEDICAL DEVICE.

ALL SUPPORTING DOCUMENTATION IS RETAINED AT THE PREMISES OF THE MANUFACTURER. EU DECLARATION OF CONFORMITY IS ISSUED UNDER THE SOLE RESPONSIBILITY OF THE MANUFACTURER

STANDARDS APPLIED: EN 60601-1:2006+A1:2013, EN 60601-1-2:2015, EN 60601-1-6:2010+A1:2015, EN 60601-1-11:2015, EN IEC 80601-2-30: 2019, EN ISO 81060-2: 2019+A1: 2020, EN ISO 10993-1:2020, EN ISO 10993-5:2009, EN ISO 10993-10:2013, EN ISO 14155:2011, EN ISO 14971:2019, EN 62304:2006+A1:2015, EN 62366-1:2015, EN ISO 15223-1:2021, EN 1041:2008+A1:2013, EN ISO 780:2015

NOTIFIED BODY: TÜV SÜD PRODUCT SERVICE GMBH

RIDLERSTR 65, D-80339 MÜNCHEN, GERMANY

IDENTIFICATION NUMBER 012

(EC) CERTIFICATE(S): G10 091264 0025 Rev. 01 VALID UNTIL: 2026-02-17

2019-11-26

START OF CE-MARKING:

PLACE, DATE OF ISSUE: SHENZHEN, WITH THE

SIGNATURE: _______ FILE SIGNATURE:

MANAGEMENT REPRESENTATIVE