

## DECLARATION OF CONFORMITY

This Declaration of Conformity is issued under the sole responsibility of the manufacturer.

### 1. Declaration of Conformity

This Declaration of Conformity is made against the Electromagnetic Compatibility Directive 2014/30/EU, Low Voltage Directive 2014/35/EU and RoHS Directive 2011/65/EU.

### 2. Product models

FieldLogger

### 3. Product description

The FieldLogger is a data acquisition and recording instrument with 8 universal analog inputs and 8 digital I/O that features communication interfaces like Ethernet, USB and RS-485. It can act as a Modbus master and can manage alarm functions.

### 4. Conformity

The object of the Declaration described above is in conformity with the relevant Union harmonization legislation:

- EMC: EN 61326-1:2013 / IEC 61326-1:2012

Emission:	Standard	Description
CISPR11	2009 +A1: 2010	Conducted and Radiated emission

Immunity:	Standard	Description
IEC 61000-4-2 2008		Electrostatic discharge
IEC 61000-4-3 2006 +A1: 2007 +A2: 2010		Radiated electromagnetic field
IEC 61000-4-4 2012		Electrical fast transient/Burst
IEC 61000-4-5 2014 +A1: 2017		Surge
IEC 61000-4-6 2013		Conducted disturbances induced by radio-frequency fields
IEC 61000-4-8 2009		Power frequency magnetic field
IEC 61000-4-11 2004 +A1: 2017		Voltage Dips, short interruptions and voltage variations

- Safety: EN 61010-1:2010 +A1: 2016 (UL file E300526)

### 5. Manufacturer:

NOVUS PRODUTOS ELETRÔNICOS LTDA  
Rua Engenheiro Homero Carlos Simon, 737  
Canoas, RS, BRAZIL 92442-000

Canoas, 15 / July / 2021.



Sandro Rafael dos Santos  
R&D Director