

EC DECLARATION OF CONFORMITY

Hereby we,

LAUDA-GFL Gesellschaft für Labortechnik mbH
Schulze-Delitzsch-Str. 4+5
30938 Burgwedel
Federal Republic of Germany

declare that the below stated **Puridest Mono distillation apparatus** models:

PD 2 R, PD 4 R, PD 8 R and PD 12 R

with the technical data:

**230 V, 50 / 60 Hz
1.5 kW (PD 2 R)
3.0 kW (PD 4 R)**

**230 V / 3 / PE, 50 / 60 Hz or
400 V / 3 / N / PE, 50 / 60 Hz
6.0 kW (PD 8 R)
9.0 kW (PD 12 R)**

are in conformity with the following EC Directives:

I	2014/35/EU	(Low Voltage Directive)
II	2014/30/EU	(EMC Directive)
III	2011/65/EU + (EU) 2015/863	(RoHS Directive)

For conformity **with I** the following standards were applied:

**EN 61010-1:2010
EN 61010-1-010:2014**

For conformity **with II** the following standard was applied:

EN 61326-1:2013

Authorized representative for the compilation of the technical documentation:

Mr Florian Wunderling at LAUDA-GFL

LAUDA-GFL Ges. für Labortechnik mbH


Andreas Degmayr
Managing director

Burgwedel, 01 July 2020

LAUDA-GFL Gesellschaft für Labortechnik mbH
Schulze-Delitzsch-Str. 4/5 • 30938 Burgwedel • DE

T +49 (0) 5139 9958-0 • F +49 (0) 5139 9958-21
info@lauda-gfl.de • www.lauda-gfl.de

WEEE-Reg.-Nr.: DE 67770231
Amtsgericht Hannover • HRB 120071
VAT-ID.-No. DE 115042911

Geschäftsführer:
Andreas Degmayr

Commerzbank AG Hannover
IBAN DE61 2504 0066 0141 2089 00
BIC COBA DE 33 XXX

Sparkasse Hannover
IBAN DE88 2505 0180 1050 2062 65
BIC SPKH DE 33 XXX