CERTIFICATE OF ANALYSIS

Digitally signed by Bolea Petru
Date: 2026.04.03 14:15:45 EEST
Reason: MoldSign Signature
Location: Moldova
MOLDOVA EUROPEANĂ

Product : **SALMONELLA ENRICHMENT 500G**
Reference : BK194HA
Batch N° : 0082233
Expiry date : 2027.11.23

Analysis	Specifications	Results
Characters		
Appearance of dehydrated medium	Free-flowing and homogeneous powder	Conform
Color of dehydrated medium	White cream	Conform
Appearance of prepared medium	Limpid solution may contain a light precipitate	Conform
Color of prepared medium	Amber	Conform
Tests		
pH of prepared medium	6,8 to 7,2	6,9
Microbiology		
Cultural response after 18 hours of incubation at 41,5°C in Salmonella Enrichment Broth, followed by 24h of incubation at 37°C onto IRIS® Salmonella Agar		
Growth Salmonella Enteritidis CIP 82.97 (WDCM 00030) + Escherichia coli ATCC 25922 (WDCM 00013) + Pseudomonas aeruginosa ATCC 27853 (WDCM 00025)	> 10 magenta colonies	Conform
Growth Salmonella Typhimurium ATCC 14028 (WDCM 00031) + Escherichia coli ATCC 8739 (WDCM 00012) + Pseudomonas aeruginosa ATCC 27853 (WDCM 00025)	> 10 magenta colonies	Conform
Growth Escherichia coli ATCC25922 (WDCM 00013)	Colorless colonies	Conform
Growth Enterococcus faecalis ATCC 29212 (WDCM 00087)	Inhibited, score 0	Conform

Biokar Diagnostics - Rue des Quarante Mines - ZAC de Ther - Allonne - BP10245 - F60002 BEAUVAIS CEDEX - FRANCE
Tel: +33(0)3 44 14 33 33 Fax: +33(0)3 44 14 33 34

CERTIFICATE OF ANALYSIS

Product : **SALMONELLA ENRICHMENT 500G**
Reference : BK194HA
Batch N° : 0082233
Expiry date : 2027.11.23

Cultural response after 45 minutes at 1 hours of incubation at 20-25°C (NF EN ISO 11133)

Growth Escherichia coli ATCC 8739 (WDCM 00012) (expressed in % colonies/T0)	-30 to 30 %	-16
Growth Staphylococcus aureus ATCC 25923 (WDCM 00034) (expressed in % colonies/T0)	-30 to 30 %	15

Cultural response after 60 minutes of incubation at 20°C (NF EN ISO 11133)

Growth Listeria monocytogenes CIP 59.53 (WDCM 00021) (expressed in % colonies/T0)	-30 to 30 %	28
Growth Listeria monocytogenes CIP104794 (WDCM 00109)(expressed in % colonies/T0)	-30 to 30 %	2

Beauvais May 02, 2023

CONFORM **Quality Control : Celine PILLITU**

Biokar Diagnostics - Rue des Quarante Mines - ZAC de Ther - Allonne - BP10245 - F60002 BEAUVAIS CEDEX - FRANCE
Tel: +33(0)3 44 14 33 33 Fax: +33(0)3 44 14 33 34