



TM-D 4056

Since 1991
turmed[®]
HOSPITAL EQUIPMENT

**HASTA KARYOLASI (DÖRT MOTOR)
PATIENT BED (FOUR MOTOR)**



STANDART DONANIM VE ÖZELLİKLER

- Yükseklik ayarı. (motorlu)
- Sırt ayarı. (motorlu)
- Diz ayarı. (motorlu)
- Ayak ucu. (manuel)
- Sırt ve diz geriye kayma hareketi. (autocontour)
- Trendelenburg ve ters trendelenburg (motorlu).
- El kumandası.
- ABS yatak platformu. (çıkarılabilir)
- ABS başlıklar. (çıkarılabilir)
- ABS korkuluklar.
- Açı göstergeleri. (sırt ve trendelenburg)
- Serum askısı. (yükseklik ayarlı)
- Aksesuar yuvaları.
- Tekerlekler ikisi kilitli.
- Çarpma tampon tekerleri.
- Elektrostatik toz boyalı gövde. (RAL 7035)

STANDARD FEATURES

- Height adjustment. (motor)
- Backrest adjustment. (motor)
- Thigh rest adjustment. (motor)
- Footrest. (manuel)
- Auto-regression system.
- Trendelenburg and reverse trendelenburg (motor).
- Handset control.
- ABS mattress platform. (detachable)
- ABS head and foot boards. (detachable)
- ABS Side Rails.
- Angle indicator. (backrest, trendelenburg)
- IV pole. (height adjustable)
- Accessories holders.
- Castors two with break.
- Protective corner bumpers.
- Electrostatic powder coated frame. (RAL 7035)

ÖLÇÜLER

Dış boyutlar	: 2030 x 1070 mm
Yatak platformu	: 1900 x 900 mm
Karyola yükseklik ayarı	: 440 - 750 mm
Maksimum sırt açısı	: 70°
Maksimum diz açısı	: 40°
Trendelenburg / Ters trendelenburg	: + 12° / - 12°
Autocontour (sırt - diz)	: 75 - 50 mm
Korkuluk yüksekliği	: 370 mm
Başlık yüksekliği	: 425 mm
Yatak altı boşluğu	: 160 mm
Tekerlekler	: 125 mm
Güvenli çalışma yükü	: 150 kg
Taşıma (yükleme) kapasitesi	: 300 kg

TECHNICAL PARAMETERS


External Dimensions	: 2030 x 1070 mm
Mattress Platform	: 1900 x 900 mm
Height Adjustment	: 440 - 750 mm
Backrest max. angle	: 70°
Footrest max. angle	: 40°
Trendelenburg / Rev. trendelenburg	: + 12° / - 12°
Auto-regression (Backrest - knee)	: 75 - 50 mm
Height of side rails	: 370 mm
Height of head-foot boards	: 425 mm
Under-bed clearance	: 160 mm
Castors diameter	: 125 mm
Safe working load	: 150 kg
Max. Loading capacity	: 300 kg

OPSİYONEL DONANIMLAR

- Tartı sistemi.
- Karyola boyu uzatma.
- Batarya.
- X-Ray kaset yeri. (sırt)
- Sırt kısmı manuel CPR.
- Korkuluk kumandası.
- Manyetik kilitleli el kumandası.
- Ortopedi traksiyon çerçevesi.
- Tutunma kolu.
- Hemşire paneli.
- Merkezi yön kilitleli tekerlek.

OPTIONAL FEATURES

- Weight scale system.
- Bed length extension.
- Battery.
- X-Ray cassette holder. (Backrest)
- Manual CPR at backrest.
- Integrated side rail controls.
- Handset control with magnetic lock.
- Orthopedics traction frame.
- Lifting Pole.
- Nurse control.
- Central lock castors.



Firmamız, Ürün Ölçü ve Özelliklerini Değiştirme Hakkına Sahiptir. Güncel Bilgileri Web Sitemizden Takip Edebilirsiniz.
Our company has the right to change the product dimensions and specifications. You can follow up the latest information from our website.



CE DECLARATION OF CONFORMITY

We declare that the products mentioned below meets the provisions of the basic safety and health about the design and the type in the product of 93/42/EEC Medical Device Regulations.

PRODUCTS	: GYNAECOLOGICAL EXAMINATION CHAIR (TWO MOTOR)	TM-A 1020
	STRETCHER	TM-C 3003
	STRETCHER	TM-C 3017
	PATIENT BED (FOUR MOTOR)	TM-D 4056
	PATIENT BED (FOUR MOTOR)	TM-D 4059
	CNC MATTRESS	TM-D 4059

INSTRUCTIONS : 93/42/EEC Medical Device Regulations.

CLASSIFICATION : Class 1.

STANDARDS : EN ISO 14971/April 2005, TS 3888/February 1983, TS 4271/April 1984, TS 4510/April 1985, TS 4720/February 1986, TS 5541/March 1988, TS 5685/April 1988, TS 5686/April 1988, TS 5875/June 1988, TS 5876/June 1988, TS 6075/October 1988, TS 8285/April 1990, TS 8770/February 1991, TS 8771/February 1991, TS ISO 6440/April 1997 TS EN ISO 13485 Medical Devices – Quality Management Systems – For the purposes of the legislation requirements.

MANUFACTURER

NAME : TUR-MED HASTANE MALZEMELERİ SAN. VE TIC. LTD. STİ.
ADDRESS : 722/1 Sokak No: 30-32 Akincilar Buca / IZMİR / TURKEY
TEL&FAX : +90.232.254 13 16 – 254 76 33 – 254 69 22

TUR-MED HASTANE MALZEMELERİ SAN. VE TİC. LTD. ŞTİ.

722/1 SOK. NO:30-32
AKINCILAR / İZMİR / TÜRKİYE

*Has been assessed and found to Comply with the Requirements of:
Denetlenmiş ve aşağıdaki standardın gerekliliklerine uygunluğu görülmüştür:*

ISO 13485:2016

*Medical Devices -Quality Management System is applicable to:
Tıbbi Cihazlar Kalite Yönetim Sistemi*

SALES OF HOSPITAL FIXTURE, MEDICAL AND ORTHOPEDIC
PRODUCT AND MEDICAL INDUSTRIAL EQUIPMENTS

HASTANE DEMİRBAŞI, TIBBİ VE ORTOPEDİK ÜRÜN İLE
TIBBİ ENDÜSTRİYEL EKİPMANLARIN SATIŞI

Certificate Number: 2021/MDQMS/10130 Initial Certification Date: 26.03.2021
Belge Numarası: 2021/MDQMS/10130 İlk Belgelendirme Tarihi: 26.03.2021

Certification Period: 3 Years
Belgelendirme Periyodu: 3 Yıl

Certificate Validity Date: 25.03.2022
Belge Geçerlilik Tarihi: 25.03.2022



ACCREDITED
Management Systems
Certification Body
MSCB-135

