

Test Report No. 7191239631-EEC20/02-WBH
dated 14 Jul 2020



PSB Singapore

**Add value.
Inspire trust.**

Note: This report is issued subject to the Testing and Certification Regulations of the TÜV SÜD Group and the General Terms and Conditions of Business of TÜV SÜD PSB Pte Ltd. In addition, this report is governed by the terms set out within this report.

SUBJECT:

Testing of Gloves submitted by Sri Trang Gloves (Thailand) Public Company Limited (Songkhla Branch) on 01 Jul 2020.

TESTED FOR:

Sri Trang Gloves (Thailand) Public Company Limited
110 Kanchanawit Road.,
T.Patong, A.Hatyai,
Songkhla, 90230,
Thailand.

TEST DATE:

13 Jul 2020

DESCRIPTION OF SAMPLES:

S/N	Product Description	Colour	Lot No.	Size	Sample received (pieces)	Manufacturer
1	Latex Powder Free Online Chlorination Glove	White	013145956	M	250	Sri Trang Gloves (Thailand) Public Company Limited

Lot size as specified by client: 35,001 to 150,000 pieces

METHOD OF TEST:

As requested by client, the test was conducted as follows:

EN 455-1:2020 Medical gloves for single use
Part 1: Requirements and testing for freedom from hole



Laboratory:
TÜV SÜD PSB Pte. Ltd.
No.1 Science Park Drive
Singapore 118221

Phone : +65-6885 1333
Fax : +65-6776 8670
E-mail: enquiries@tuv-sud-psb.sg
www.tuv-sud-psb.sg
Co. Reg : 199002667R

Regional Head Office:
TÜV SÜD Asia Pacific Pte. Ltd.
1 Science Park Drive, #02-01
Singapore 118221
TUV[®]

Test Report No. 7191239631-EEC20/02-WBH
dated 14 Jul 2020



PSB Singapore


RESULTS:

Sample: Latex Powder Free Online Chlorination Glove, Lot No. 013145956, Size M

S/N	Tests	Requirements	No. of non-compliers allowed (pieces)	Number tested (pieces)	Actual no. of non-compliers found (pieces)	Inferred results
1	Freedom from holes	Shall not leak	7	200	3	Pass

REMARKS:

The manufacturing batch code was not provided by the client.



Yeo Poh Kwang
Associate Engineer



Wong Bee Hui
Product Manager
Medical Health Services (NAM)

APPENDIX:



Photo: Latex Powder Free Online Chlorination Glove, Lot No. 013145956, Size M

Test Report No. 7191239631-EEC20/02-WBH
dated 14 Jul 2020



Please note that this Report is issued under the following terms :

1. This report applies to the sample of the specific product/equipment given at the time of its testing/calibration. The results are not used to indicate or imply that they are applicable to other similar items. In addition, such results must not be used to indicate or imply that TÜV SÜD PSB approves, recommends or endorses the manufacturer, supplier or user of such product/equipment, or that TÜV SÜD PSB in any way "guarantees" the later performance of the product/equipment. Unless otherwise stated in this report, no tests were conducted to determine long term effects of using the specific product/equipment.
2. The sample/s mentioned in this report is/are submitted/supplied/manufactured by the Client. TÜV SÜD PSB therefore assumes no responsibility for the accuracy of information on the brand name, model number, origin of manufacture, consignment or any information supplied.
3. Nothing in this report shall be interpreted to mean that TÜV SÜD PSB has verified or ascertained any endorsement or marks from any other testing authority or bodies that may be found on that sample.
4. This report shall not be reproduced wholly or in parts and no reference shall be made by the Client to TÜV SÜD PSB or to the report or results furnished by TÜV SÜD PSB in any advertisements or sales promotion.
5. Unless otherwise stated, the tests were carried out in TÜV SÜD PSB Pte Ltd, No.1 Science Park Drive Singapore 118221.

July 2011

