

NINGBO GREETMED MEDICAL INSTRUMENTS CO., LTD.

CERTIFICATE OF ANALYSIS

Report No: 240223-1

Product name	Surgical scrub brush	Quantity	15000pcs
LOT	24022301	MFG	202404
Test based on	Enterprise standard	EXP	203103
Inspection date	2024-04-11	Report date	2024-04-11
Item	Technical requirements	Result	Status
Appearance	Neatly packaged, paper-plastic packing in good condition, clear printing complete.	Meet Specifications	PASS
Specifications	Containing chloride glucurate and nail file.	Meet Specifications	PASS
Plastic brush	Surface should be smooth and no burrs burr and cracks	Meet Specifications	PASS
	Brush is soft, no fall off	Meet Specifications	PASS
sponge	Neat edges and shape Angle	Meet Specifications	PASS
	Specifications for the 80 * 50 * 25 mm(±10%)	Meet Specifications	PASS
nail file	Smooth surface, no burrs, no fly, no impurities.	Meet Specifications	PASS
Chlorhexidine Gluconate Solution	4% 宁波柯泰医疗器械有限公司 NINGBO GREETMED MEDICAL INSTRUMENTS CO.,LTD.	4%	PASS
Thoracic Fluid Content	20ml	20ml	PASS
Conclusion	The product meets the requirements 李桂蓉		

Inspection: 杨沈静

Review: 申华马