

CERTIFICATE OF COMPLIANCE

Certificate Number 2016-2-24-E311966
Report Reference E311966-D1004-1/A0/C0-ULCB-ULCB
Issue Date 2016-2-24

Issued to: THERMO ELECTRON LED GMBH
Applicant Company: ZWEIGNIEDERLASSUNG OSTERODE
AM KALKBERG
37520 OSTERODE AM HARZ GERMANY

Listed Company: Same as Applicant

This is to certify that representative samples of Laboratory Centrifuge
Heraeus Cryofuge 8, Heraeus Cryofuge 16, Sorvall BP 8, Sorvall
BP16, Sorvall BIOS 16

Have been investigated by UL in accordance with the
Standard(s) indicated on this Certificate.

Standard(s) for Safety: UL 61010-1 Ed.3, Published May 11, 2012, Revision dated July
15, 2015, CAN/CSA C22.2 No. 61010-1-12 Published May 2012,
Update No.1 July 2015, IEC 61010-1:2010 (Third Edition)

Additional Information: See the UL Online Certifications Directory at
www.ul.com/database for additional information.

Only those products bearing the UL Certification Mark should be considered as being covered by UL's
Certification and Follow-Up Service.

Look for the UL Certification Mark on the product.

This is to certify that representative samples of the product as specified on this certificate were tested
according to the current UL requirements.

B. Mahlenz *John H. Hines*

Bruce Mahlenz, Assistant Chief Engineer, Global Inspection and Field Services, UL LLC
Joseph Hosey, General Manager, Director of Sales – Canada, UNDERWRITERS LABORATORIES OF CANADA INC.

Helena Y. Wolf

Helena Y. Wolf, Director, Global Market Access Operations, UL LLC

Any information and documentation involving UL Mark services are provided on behalf of UL LLC (UL) or any authorized licensee of UL. For questions, please contact a local UL
Customer Service Representative www.ul.com/contactus

