

EC Certificate

Full Quality Assurance System Directive 93/42/EEC on Medical Devices, Annex II excluding (4)

Registration No.:

HD 1169866-1

Manufacturer:

Ziehm Imaging GmbH Lina-Ammon-Str. 10 90471 Nürnberg

Germany

Products:

mobile Radiographic/Fluoroscopic Units (mobile C-arms)

Replaces Certificate, Registration No.: HD 60144926 0001

The Notified Body hereby declares that the requirements of Annex II, excluding section 4 of the directive 93/42/EEC have been met for the listed products. The above named manufacturer has established and applies a quality assurance system, which is subject to periodic surveillance, defined by Annex II, section 5 of the aforementioned directive. For placing on the market of class III devices covered by this certificate an EC design-examination certificate according to Annex II section 4 is required.

Report No.:

1083421-20

Effective date:

2021-03-11

Expiry date:

2024-05-26

Issue date:

2021-03-11

Reland Gruber

Tillystraße 2 · 90431 Nürnberg · Germany

TÜVRheinland

TÜV Rheinland LGA Products GmbH is a Notified Body according to Directive 93/42/EEC concerning medical devices with the identification number 0197.

Page 1 of 1