



Product Service

# EC Certificate

## EC Design-Examination Certificate

Directive 93/42/EEC on Medical Devices (MDD), Annex II (4)  
(Devices in Class III)

No. G7 17 08 73322 014

**Manufacturer:** **Beijing Target Medical Technologies Inc.**

No.60, Shunren Rd.  
Shunyi District  
101300 Beijing  
PEOPLE'S REPUBLIC OF CHINA



**EC-Representative:** **Lepu Medical (Europe) Cooperatief U.A.**

Abe Lenstra Boulevard 36  
8448 JB Heerenveen  
THE NETHERLANDS


**Product:** **Catheter for Cardiology and Angioplasty  
Disposable central venous catheter kit**

The Certification Body of TÜV SÜD Product Service GmbH declares that a design examination has been carried out on the respective devices in accordance with MDD Annex II (4). The design of the devices conforms to the requirements of this Directive. For marketing of these devices an additional Annex II certificate is mandatory. See also notes overleaf.

**Report no.:** BJ1771106

**Valid from:** 2018-04-09

**Valid until:** 2020-09-26



**Date,** 2018-04-09

Stefan Preiß

TÜV SÜD Product Service GmbH is Notified Body with identification no. 0123

Page 1 of 2



Product Service

**EC Certificate****EC Design-Examination Certificate**Directive 93/42/EEC on Medical Devices (MDD), Annex II (4)  
(Devices in Class III)

No. G7 17 08 73322 014

**Model(s):** Disposable central venous catheter kit**Parameters:**

Size Range:

Single Lumen: 14G-20G(OD:1.05-2.1mm)

Double Lumen: 4F-11.5F(OD:1.35-3.85mm)

Triple Lumen: 5.5F-7F(OD:1.85-2.35mm)

Length of Tubing: 50-600mm

**Facility(ies):**Beijing Target Medical Technologies Inc.  
No.60, Shunren Rd., Shunyi District, 101300 Beijing, PEOPLE'S  
REPUBLIC OF CHINA