







No.: SUIN2009007656MR

Date: Oct 12, 2020

Page: 1 of 4



CUSTOMER NAME: NANJING MINDRAY BIO-MEDICAL ELECTRONICS CO., LTD

ADDRESS: 666# MIDDLE ZHENGFANG ROAD, JIANGNING, 211111 NANJING,

JIANGSU

Sample Name : Medical supply units enclosure

Above information and sample(s) was/were submitted and confirmed by the client. SGS, however, assumes no responsibility to verify the accuracy, adequacy and completeness of the sample information provided by client.

SGS Ref. No. : NJIN2009004103ML

Date of Receipt : Sep 30, 2020
Testing Start Date : Sep 30, 2020
Testing End Date : Oct 12, 2020

Test result(s) : For further details, please refer to the following page(s)

(Unless otherwise stated the results shown in this test report refer only to

the sample(s) tested)

Signed for SGS-CSTC Standards Technical Services Co., Ltd. Suzhou Branch

Penny Zhang

Authorized signatory



Unless otherwise agreed in writing, this document is issued by the Company subject to its General Conditions of Service printe overleaf, available on request or accessible at http://www.scs.com/en/Terms-and-Conditions.aggs and, for electronic format documents at http://www.scs.com/en/Terms-a-Document.aggs. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document subject in the formation contained hereon reflects the Company's findings at he time of its enversion only and within the limits of the service of the servi



No.: SUIN2009007656MR

Date: Oct 12, 2020

Page: 2 of 4

Summary of Results:

No.	Test Item	Test Method	Result	Conclusion
1	Adhesion Test of Coatings	ISO 2409:2013	See result	/

Note: Pass : Meet the requirements;

Fail: Does not meet the requirements;

/: Not Apply to the judgment.





No.: SUIN2009007656MR

Date: Oct 12, 2020

Page: 3 of 4

Test Item: Adhesion Test of Coatings Sample Description: White coated plate

Test Method: ISO 2409:2013


Test Condition: Space: 1mm Tape: 3M®600#

Test Result:

Test Item	Adhesion classification
Adhesion Test of Coatings	0

Note: According to ISO 2409:2013, in the adhesion classification--classification 0 is the best and classification 5 is the worst.

Test Photo:






Unless otherwise agreed in writing, this document is issued by the Company subject to its General Conditions of Service printed overleaf, available on request or accessible at http://www.ags.com/en/Terms-and-Conditions for electronic Format documents at <a href="http://www.ags.com/en/Terms-and-Conditions/fer



No.: SUIN2009007656MR

Date: Oct 12, 2020

Page: 4 of 4



- The test report was published by SGS-CSTC SuZhou Branch.
- Part/all of the test methods are not in CMA accredited scope, the test report shall only be used for client scientific research, teaching, internal quality control, product research and development, etc. and just for client internal reference.

****** End of report******

