



# MOVEE 1.0 — Patient Transport Stretcher

- Backrest adjustment with gas spring
- Trendelenburg and reverse trendelenburg adjustment assisted by hydrolic with manual control
- Foldable side rails
- Cross lockable castors
- Electorstatic painted metal frame
- Plastic crash bumpers
- Removable, washable at 95°, fire retardant mattress cover.

Length: 200 cm

Width: 80 cm

Castor Diameter: 15 cm

Safe Working Load: 200 kg

MEYSA TIBBİ CİHAZLAR SAN. VE TİC. A.Ş.  
O.S.B. 16. Cad. No: 73 Melikgazi/KAYSERİ/TÜRKİYE  
Tel: +90 352 503 16 60 Fax: +90 352 506 16 61  
www.meissa.com.tr info@meissa.com.tr



# MOVEE 2.0

HİDROLİK KOTROLLÜ ACİL MÜDAHALE SEDYESİ  
PATIENT EMERGENCY TREATMENT STRETCHER

- Amortisör yardımı ile ayarlanabilen sırt kısmı
- Hızlı ve güvenli ayarlanabilen yükseklik trendelenburg ve ters trendelenburg pozisyonları
- Kaldırılabilir kenar korukları sayesinde kolay ve güvenli kullanım.
- Sağlam hijyenik ve çarpmalara karşı korumalı çalkı şase.
- 20 cm ebatında merkezi kilit tekerler.
- Her iki taraftan kullanıma imkân veren pedal sistemi
- X-Ray geçişgen kompakt laminant sırt.
- Hasta bağlama kemeri.
- Sedyenin alt kısmında (ABS plastik üzerinde) oksijen tüpü yuvası.
- 95°'de ykılabilen, elev alınmaz, çift taraflı esnek, şilte kılıfı.
- Yüksekliği ayarlanabilir paslanmaz çalkı serum askısı.
- Elastik çarpma tamponları.
- Elektrostatik Boyalı metal aksam.

## TEKNİK ÖZELLİKLER

- Uzunluk / Overall Length : 206 cm
- En / Overall Width : 80 cm
- Yükseklik(Şilte) / Height Range : 56- 84 cm
- Yatma Yüzeyi En / Sleep Surface Width : 62 cm
- Yatma Yüzeyi En / Sleep Surface Width : 192 cm
- Trendelenburg / Trendelenburg : 14°
- Sırt Kısmı(Maximum) / Backrest Angle (Max.) : 90°
- Teker Çapı / Castor Diameter : 20 cm
- Yatac Alt Boşluğu / Under Bed Clearance : 10 cm
- Güvenli Çalışma Kapasitesi / Safe Working Load : 230 kg
- (Aksesuarlar dahil) / (With Accessories)

## EBAT / DIMENSIONS

- Backrest adjustment assisted by gas spring.
- High low trendelenburg reverse trendelenburg adjustment assisted by hydraulic.
- Foldable side guards, easy and safe handling.
- Solid and hygienic construction, no danger of accidental crushing.
- 20 cm castors with central loading.
- Dual sided adjustment pedals.
- X-Ray translucent backrest.
- Patient belt for fastening.
- Accessories holder.
- Washable at 95° fire retardant mattress cover.
- Height adjustable stainless steel IV pole.
- Plastic crash bumpers.
- Electrostatic painted metal frame.

## TECHNICAL DETAILS



GİZLEMLİ RİME BARI  
FLEXIBLE PUSH BAR



KONTROL PEDALLARI  
POSITION PEDALS




ÇARPMA TAMPONU  
PATIENT HANDSET




MERKEZİ FREN SİSTEMİ  
CENTRAL LOCK




Sırt Hareketi  
Backrest



Ayağı Hareketi  
Legrest



Yükseklik  
High/Low



Düz Trendelenburg  
Trendelenburg




Ters Trendelenburg  
Reverse Trendelenburg

# MOVEE 2.0

# MOVEE 2.0

## MOVEE 2.0

HİDROLİK KOTROLLÜ ACİL MÜDAHALE SEDYESİ  
PATIENT EMERGENCY TREATMENT STRETCHER



Low Bed



Sirt Hareketi / Backrest



Düz Trendelenburg / Trendelenburg



Yüksek / High Low





## AT UYGUNLUK BEYANI / EC DECLARATION OF CONFORMITY

BELGE No/CERT.No: CE-03

MEYSA TIBBİ CİHAZLAR SANAYİ VE TİCARET ANONİM ŞİRKETİ

Aşağıda imal yılı seri numarası verilen mamul/mamullerin ilk dökümanlarda tanımlandığı gibi 93/42/EEC Tıbbi Cihazlar Direktifi yönetmeliği karşılamak şekilde ürettiğimizi beyan ederiz.

WE CLEARLY DECLARE THAT OUR PRODUCT(S), GIVEN BELOW WITH SERIAL NUMBER AND PRODUCTION YEAR, MANUFACTURED IN COMPLIANCE WITH 93/42/EEC MEDICAL DEVICES DIRECTIVE, AS DEFINED IN FIRST DOCUMENTS



Type : Pioneer

Model : 2 Motors Electrical Patient Bed, 3 Motors Electrical Patient Bed, 4 Motors Electrical Patient Bed, 4 Motors Intensive Care Patient Bed, 4 Motors Intensive Care Patient Bed, 4 Motors Intensive Care Patient Bed, 4 Motors Column Model Intensive Care Patient Bed, Grande 2,3,4,CL, Junior 2,3,4, Elite 2,3,4

Year of prod. : 2021

Standard No. : TS EN 60601-1:2009 , TS EN 60601-2-52:2010/AC:2011, EN 1865-1:2010+A1:2015, EN 60204-1 , TS EN ISO 13849-1 ,

Address : ORGANIZE SAN. 16. CD. NO: 73

WAS TURKEY

Certification Manager  
Ahmet YILDIRIM

06/03/2021

[www.wasturkey.com](http://www.wasturkey.com) [info@wasturkey.com](mailto:info@wasturkey.com)

Ebulula Mardin Cad. Yıldırım Oğuz Göker Sokak  
Park Maya Sitesi Carlton 17 Blok Daire:9 34335 Akatlar/İstanbul





**MEYSA TIBBİ CİHAZLAR SAN. ve TİC. A.Ş.**

O. S. B. 16.Cad. No :73 Melikgazi / KAYSERİ / TÜRKİYE

+90 352 503 16 60

www.meissa.com.tr

+90 352 503 16 61

info@meissa.com.tr



## DECLARATION OF CONFORMITY

IN accordance with 93/42/EC Medical Devices Directive of the Council of European Union, whose purpose is providing conformity of laws, directives and administrative of documents of member countries in respect to Medical Device;

**Name and Address of the Company:** MEYSA TIBBİ CİHAZLAR SAN VE TİC A.S.

**Tel:** +90 352 5013 16 6 **Fax:** +90 352 503 16 61

**Place and Address of Manufacture:** OSB. 16.Cad No:73 Melikgazi Kayseri / TURKEY

**Applied Directives:** 93/42/EC- Medical Devices Directive

**Classification and Annex Applied:** Product is subject to Medical Devices Directive Class 1. Applied.

**Name of Product and Types:**

**Patient and ICU Beds:** PIONEER CLW, CL, 4, 4.3PLUS, 4.2CARE, 4.1TREND, 3, 2, ELITE CL, 4.3 PLUS,4, 3, 2 GRANDE CL, 4 ,3, 2, DUO

**Manual Patient Beds:** PRIMITIVE 3,2

**Stretchers:** MOVEE2.1, MOVEE 2.0, DYNAMIC 2.0, MOVEE VERSUS

**Gynecological Ex. Table:** Mathercare 2

**Pediatric Bed:** JUNIOR 4, 3, 2, JUNIOR, JUNIOR ELITE

**Baby Cot:** NICE 1, NICE 2

**Examination Table:** Diagno 01,02,03

**Patient Bed Side Cabinets:** SERA 1,2,3,4,5,6,7,8

**Patient Overbed Table:** MIRACH 1,2,3

**Declaration;**

Our company manufactures the products stated above in accordance with the requirements of the current EN 980-1996/A1 (Graphs and Symbols Used on labels) EN 1401 (Information provided with the Product by Manufacturer) ISO 13485:2003 Medical Devices Quality Management System.

**Used Standards;**

The mentioned products are complying with the requirements of the following **standards**; EN 60601-1:2006/A1:2013 EN 60601-1-2:2007 EN 60601-1-6:2010, IEC60601-2-38, EN 60601-2-52: 1-2010

The products described above were subjected to initial type experiments by Manufacturer and factory manufacture control was carried out by regular tests.

Date of Issue 01.06.2018

General Manager







# CERTIFICATE

## MEYSA TIBBİ CİHAZLAR SANAYİ VE TİCARET ANONİM ŞİRKETİ

KAYSERİ OSB MAH. 16 CAD. NO:73  
MELİKGAZI / KAYSERİ / TÜRKİYE

*Has been assessed and found to Comply with the Requirements of:  
Denetlenmiş ve aşağıdaki standardın gerekliliklerine uygunluğu görülmüştür:*

### ISO 13485:2016

*Medical Devices-Quality Management System is applicable to:  
Tıbbi Cihazlar Kalite Yönetim Sistemi*

**DESIGN AND PRODUCTION OF HOSPITAL FURNITURES, MEDICAL BED  
AND PATIEND BED**

**HASTANE MOBİLYALARI, MEDİKAL YATAK VE HASTA YATAĞI  
TASARIMI VE ÜRETİMİ**

Certificate Number: 2021/MDQMS/10256  
Belge Numarası: 2021/MDQMS/10256

Initial Certification Date: 26.05.2021  
İlk Belgelendirme Tarihi: 26.05.2021

Certification Period: 3 Years  
Belgelendirme Periyodu: 3 Yıl

Certificate Validity Date: 25.05.2022  
Belge Geçerlilik Tarihi: 25.05.2022




IQR Sertifikasyon Onayı

**IQR ULUSLARARASI BELGELENDİRME HİZMETLERİ LTD.ŞTİ.**

Beşevler Mah. Kocayunus Sk. No:3 Arslan Han Plaza K:2 Nilüfer / BURSA  
Tel.: +90.224.266 00 16 Faks: +90.224.249 41 13 www.iqrcert.com e-posta: info@iqrcert.com





## QUALITY SYSTEM CERTIFICATE

The Certification Body

B.E.C BELGELENDİRME VE MUAYENE HİZMETLERİ LTD.ŞTİ.

Certified that the

**MEYSA TIBBİ CİHAZLAR  
SANAYİ VE TİCARET A.Ş.**

Organize sanayi bölgesi 16. Cd. No:73  
MELİKGAZİ/KAYSERİ-TÜRKİYE

**has established and applies quality systems for**

**"DESIGN AND PRODUCTION OF HOSPITAL FURNITURES;  
MEDICAL AND PATIENT BEDS"**

**has implemented and maintains a quality management system.**

**an audit, documented in 23.409.17/A report, has verified that this quality management system fulfills the requirements of the following standart.**

**TS EN ISO 9001:2015**

**Valid Date : 15.09.2023**  
**Certificate number : 38.23.409.17/A**  
**EA Code -NACE code : 23 NACE 32.50**  
**Certificate Issuance Date /Rev. No: 15.09.2020 /0**

MANAGER OF SYSTEM  
CERTIFICATION



TÜRKAK BDS NO  
YS-54B8-86CD



**1.VISA 18.08.2020**



**2.VISA 17.08.2021**

**3.VISA 17.08.2022**

This document is valid until the date determined above under the conditions that rules of certification are obeyed and it is fixed that periodical supervisions are undergone and appropriateness of the system is sustained. Confirm the validity of the document at [www.bec.com.tr](http://www.bec.com.tr) or on fax number + 90 262 641 16 68

B.E.C LTD. -555/2 SK. NO:8 GEBZE / KOCAELİ / TÜRKİYE - +90.262.642.04.02- [www.becturkey.com](http://www.becturkey.com)