

**PLUM Sp. z o. o. is certified for production, final product inspection and testing according to module D of Directive 2014/32/EU of the European Parliament and of the Council: certificate of quality system no. 0119-SJ-A002-09.**

## REPORT OF CONFORMITY TEST

Date of issue: 02.04.2024 (dd.mm.yyyy)

Report No: 751/34/LA/C/2024

Page 1/2

<b>OBJECT OF CONFORMITY TEST</b>	<b>Volume conversion device</b> Type: <b>MacBAT 5</b> Serial number: <b>1007542008</b> Program version / Bootloader version: <b>S013.78_V142637 / B15</b> CRC: <b>CrcMain=934C8A73; CrcBoot=34CD6BC3</b> Gas pressure measurement range (P1): <b>(0.8 ÷ 10) bar abs</b> Gas temperature measurement range: <b>(-30 ÷ 70) °C</b> Pressure sensor serial number (P1): <b>1007295922</b> Temperature sensor serial number: <b>1007441550</b>
<b>MANUFACTURER</b>	PLUM Sp. z o.o. ul. Wspólna 19, Ignatki, 16-001 Kleosin, Poland
<b>EU-TYPE EXAMINATION CERTIFICATE</b>	DE-19-MI002-PTB004
<b>METHOD AND TEST CONDITIONS</b>	Test procedure: ILAJ 5.4/10, according to point A.1.4.2 of EN 12405-1. Base conditions: $p_b = 1.01325$ bar, $T_b = 20$ °C Combustion reference temperature: $T_1 = 20$ °C Test algorithm range: $(-30 \div 70)$ °C Algorithm: AGA8-92DC  Test gas (%mol): C1=85.9, C2=8.5, C3=2.3, n-C4=0.35, i-C4=0.35, n-C5=0.05, i-C5=0.05, neo-C5=0, C6+=0, N2=1, CO2=1.5, C6H14=0, C7H16=0, C8H18=0, C9H20=0, C10H22=0, H2=0, H2O=0, H2S=0, CO=0, He=0, Ar=0, O2=0
<b>CONFORMITY WITH REQUIREMENTS</b>	The object of conformity test meets the requirements specified in Annex IV of Directive 2014/32/EU of the European Parliament and of the Council. The results of conformity test have been presented on page 2 of this report including uncertainty of measurement.



**DEPUTY MANAGER**  
of the Measurement Laboratory  
*Piotr Chociej*  
**Piotr Chociej MSc. Eng.**

**RESULTS OF CONFORMITY TEST** Test results are the following.

**Conversion factor C**

Reference value			Indication of tested device			Error	Uncertainty of measurement
Temperature	Pressure	Conversion factor	Temperature	Pressure	Conversion factor		
$t_{cv}$	$p_{cv}$	$C_{cv}$	$t$	$p$	$C$	$e_c^{1)}$	$U_c$
°C	bar abs	-	°C	bar abs	-	%	%
-30.01	0.8000	0.953031	-30.01	0.8001	0.953156	0.01	0.15
-30.01	3.1000	3.733017	-30.01	3.1003	3.733420	0.01	0.13
-30.01	5.4000	6.574768	-30.01	5.4003	6.575239	0.01	0.09
-30.01	7.7000	9.481333	-30.01	7.7004	9.481881	0.01	0.09
-30.01	10.0000	12.456152	-30.01	10.0004	12.456740	0.00	0.09
20.00	10.0000	10.097095	19.99	10.0003	10.097703	0.01	0.09
20.00	7.7000	7.729113	19.98	7.7002	7.729603	0.01	0.08
20.00	3.1000	3.075549	19.99	3.1003	3.075963	0.01	0.12
20.00	0.8000	0.789128	19.99	0.8001	0.789291	0.02	0.15
69.98	0.8000	0.673587	69.97	0.8001	0.673672	0.01	0.15
69.98	3.1000	2.618618	69.97	3.1003	2.618922	0.01	0.12
69.98	5.4000	4.576206	69.97	5.4000	4.576321	0.00	0.08
69.98	7.7000	6.546339	69.97	7.7002	6.546746	0.01	0.08
69.98	10.0000	8.529194	69.97	10.0003	8.529640	0.01	0.08

<sup>1)</sup> relative error

**Volume at base conditions  $V_b$**

Reference value			Indication of tested device	Error	Uncertainty of measurement
Temperature	Pressure	Volume at base conditions			
$t_{cv}$	$p_{cv}$	$V_{cv}$	$V_b$	$e_v^{2)}$	$U_v$
°C	bar abs	m <sup>3</sup>	m <sup>3</sup>	%	%
20.00	5.4000	538.87	538.90	0.01	0.09

<sup>2)</sup> relative error

Maximum error  $|e_c|$  = 0.02% < 0.5%  
 Error  $|e_v|$  = 0.01% < 0.5%  
 Maximum error  $|e_p|$  = 0.02% < 0.2%  
 Maximum error  $|e_t|$  = 0.00% < 0.1%

**ENVIRONMENTAL CONDITIONS** Ambient temperature: (23.2 ÷ 24.1) °C  
 Relative humidity: (39 ÷ 49) %

**DATE OF TEST** 02.04.2024 (dd.mm.yyyy)

**TRACEABILITY** This report provides traceability of measurement to national measurement standards, which realize the units of measurement according to the International System of Units (SI). Measurement standards applied for calibration have been listed below.

Name of measurement standards	Type	Serial number
Electronic Thermometer	MacREJ 5	1004466418
Standard Pulse Counter	MacIMP E	92701i
Pressure Controller	CM2-B-13G	10329693

**UNCERTAINTY OF MEASUREMENT** Uncertainty of measurement has been evaluated in compliance with EA-4/02. The expanded uncertainty assigned corresponds to a coverage probability of 95% and the coverage factor  $k = 2$ .