	ONESPRAY WIPES			
	Alcohol-Based Fast Acting Disinfectant Wipes for Medical Devices			
	Document Number	Release date	Revision No	Revision date
	TD.04.123	10.06.2019	03	24.05.2021




ONESPRAY

**ALCOHOL-BASED FAST ACTING DISINFECTANT
WIPES FOR MEDICAL DEVICES**



KAF GRUP SAĞLIK HİZMETLERİ İNŞ. SAN. VE TİC. LTD. ŞTİ.

**Atakent Mah. 221 Sk. No:3A Rota Office A Blok Kat:14 D:82-83
K.ÇEKMECE / İSTANBUL**

	ONESPRAY WIPES Alcohol-Based Fast Acting Disinfectant Wipes for Medical Devices			
	Document Number	Release date	Revision No	Revision date
	TD.04.123	10.06.2019	03	24.05.2021

General Features

OneSpray Wipes used for disinfection of pre-cleaned, non-invasive, smooth, smooth and non-porous medical devices (ultrasound probes etc.) It is ready to use. It does not contain aldehyde and phenol. It is based on a combination of ethanol, propanol and quaternary ammonium and does not cause color fading or dyeing on the material used. It is fragrant.

Our OneSpray Wipes products are offered for sale in 3 different forms as 100 pieces, 150 pieces and 200 pieces.

Chemical Composition

The content of the solution impregnated with the wipe;

Chemical Composition	ATC Codes
Ethanol	D08AX53
2-Propanol	D08AX05
Didecylmethylpoly(oxethyl) Ammonium Propionate	D08AJ06
Alcohols C10, ethoxylated	-
Essence	-
Deionized Water	-


Usage Area and Purpose

The product is used for rapid disinfection of non-invasive medical devices, incubators, aerators used in dentistry and medical devices with mandatory disinfection used in ENT examinations.

Effect Mechanism:

The mechanism of action of our OneSPRAY Wipes products as disinfectants is bactericidal, virucidal and fungicidal. They disrupt the lipid structure of the cell membrane of the microorganisms in the medical device and also denature the cell proteins. These substances act by disrupting the three-dimensional structure of proteins and causing random ringing and helicalization of the polypeptide chain. In the table below, the microorganisms on which our products are effective are indicated according to the tests performed.

Products	Shelf Life	Tub Life Time	Usable surface materials	Microorganisms and Effect times
Onespray Wipes	2 years	30 days	<ul style="list-style-type: none"> - Glass - Ceramic - Silicone - Plastic 	<i>Pseudomonas aeruginosa</i> - 1 minute (EN 16615) <i>Staphylococcus aureus</i> - 1 minute (EN 16615) <i>Enterococcus hirae</i> - 1 minute (EN 16615)

	ONESPRAY WIPES Alcohol-Based Fast Acting Disinfectant Wipes for Medical Devices			
	Document Number	Release date	Revision No	Revision date
	TD.04.123	10.06.2019	03	24.05.2021

			- PVC - Acrylic - polymer - Polycarbonate - Polystyrene - HDPE - Stainless steel - Porcelain	<i>Candida albicans</i> - 30 seconds (EN 13624) Polyvirus Type 1- 1 minute (EN 14476) Human Adenovirus Type 5- 1 minute (EN 14476) Murine norovirus - 1 minute (EN 14476) <i>Staphylococcus aureus</i> - 30 seconds (EN 13727) <i>Pseudomonas aeruginosa</i> - 30 seconds (EN 13727) <i>Enterococcus hirae</i> - 30 seconds (EN 13727) <i>Mycobacterium avium</i> - 5 minutes (EN 14348) <i>Mycobacterium terrae</i> -5 minutes (EN 14348) <i>Candida albicans</i> -30 seconds (EN 13624) <i>Aspergillus niger</i> -30 sec (EN 13624)
--	--	--	---	---

** Corrosion effect when our products are subjected to corrosion test according to TS 6773 standard was not shown.

Usage

It is used for disinfection of pre-cleaned medical devices (ultrasound probes etc.)

- With enough wipes, the entire surface of the device is wiped.
- 30 seconds - 5 minutes wait.

Before application, make sure that the medical devices are compatible with alcohol. Unplug the electrical devices before the application.

Usage Concentration

It is ready to use.


INDICATIONS-CONTRAINDICATIONS,SIDE EFFECTS, WARNINGS AND PRECAUTIONS

Indications

It is indicated for disinfection of pre-cleaned, non-invasive, smooth, smooth and non-porous medical devices (ultrasound probes etc.)

Contraindications

- It should not be used directly (without pre-cleaning) on surfaces and instruments containing organic compounds (blood, tissue, urine, etc.).

	ONESPRAY WIPES Alcohol-Based Fast Acting Disinfectant Wipes for Medical Devices			
	Document Number	Release date	Revision No	Revision date
	TD.04.123	10.06.2019	03	24.05.2021

- It should not be used together with other chemicals as it may react with chemicals.
- Persons sensitive to ethyl alcohol, isopropyl alcohol, didecylmethylpoly(oxethyl) ammonium propionate should not come into direct contact with the product.

Side effects

Our products have no known side effects.

WARNINGS and PRECAUTIONS

- Be sure to read the instructions for use before using the product.
- Avoid contact with eyes.
- Use protective clothing (gloves and eye/face protection) equipment.
- IF SWALLOWED: Rinse your mouth, do not try to vomit.
- In case of prolonged contact with the skin, wash with plenty of soap and water. In case of skin irritation get medical attention.
- Do not reuse the packaging of the finished product .
- **IF IN EYES:** Rinse carefully with water for a few minutes. mention of eye irritation If so, consult your doctor.
- Avoid spreading to the environment.
- Store in its original container.
- Dispose of waste and used packaging in accordance with official regulations.

Storage and Shipping Conditions:

The expiry date is 2 years from the date of manufacture. Keep the package tightly closed, in a well-ventilated place between 0 °C and 25 °C . Keep out of reach of children.


PATIENT POPULATION











Our products are not applied to patients, they can only be applied to the surfaces of non-invasive medical devices.


➤ Sterilization






Our products are used non-sterile.

➤ Symbols Used on the Label and Their Meanings

	ONESPRAY WIPES Alcohol-Based Fast Acting Disinfectant Wipes for Medical Devices			
	Document Number	Release date	Revision No	Revision date
	TD.04.123	10.06.2019	03	24.05.2021

	Notified Body Number		Expiration date
	Production date		Non-sterile
	Read the Instructions for Use.		Protect from sunlight.
	Store in dry place.		Storage temperature range
	Manufacturer Information		Do not use damaged package.

	ONESPRAY WIPES Alcohol-Based Fast Acting Disinfectant Wipes for Medical Devices			
	Document Number	Release date	Revision No	Revision date
	TD.04.123	10.06.2019	03	24.05.2021

	Barcode number		Reference number
	Lot number		Caution (Warnings)
	Flammable		

KAF GRUP
SAĞLIK HİZMETLERİ



KAF GRUP SAĞLIK HİZMETLERİ İNŞ. SAN. VE TİC. LTD. ŞTİ.

**Merkez Adres: Atakent Mah. 221 Sk. No:3A Rota Office A Blok Kat:14 D:82-83
K.ÇEKMECE / İSTANBUL**

Üretim Adresi: Bardakçı Mah. Teknokent Sk. No:3 Tuşba/VAN