

	<p>PRODUCT DESCRIPTION</p>	<p>PD08-03</p>
---	-----------------------------------	-----------------------

6.7. Double swivel patient connection

7. Applied Standarts:

EN ISO 13485, 93/43/EEC Medical Device Directive, ISO 12342, EN 5356, ISO 11135-1, ISO 10993-7, ISO 10993-1, EN ISO 11607-1, EN ISO 11607-2.

8. Tests performed on the product :

8.1. Dimensional controls (Gauge, caliper and weight controls)

8.2. Leakage

8.3. Resistance to flow

8.4. Routine assembling process controls

8.5. Pull test

8.6. Biocompatibility

8.7. Accelerated aging test