

Shenzhen BEL Technology Co.,Ltd.

Add: 3rd Floor, Xingfu Building, Tongfuyu Industrial Zone, Shiyan  
Town, Bao'An District, Shenzhen, Guangdong, China.



# Certificate of Compliance

**Certificate Number: BEL20200000101899**

**Applicant** : Shenzhen Wanlianxin Electronic Co., Ltd.  
**Address** : Xingye Huating, 48 District, Haihua community, Xin'an street, Bao'an District, Shenzhen 202  
**Manufacturer** : Shenzhen Wanlianxin Electronic Co., Ltd.  
**Address** : Xingye Huating, 48 District, Haihua community, Xin'an street, Bao'an District, Shenzhen 202  
**Product** : FINGERTIP PULSE OXIMETER  
**M/N** : WLX501  
WLX502, WLX503, WLX504, WLX505  
WLX506, WLX507, WLX508, WLX509  
**Test Standard** : EN 60601-1:2006+A12:2014  
EN 60601-1-2:2015

The EUT described above has been tested by us with the listed standards and found in compliance with the council Medical products general directive 2007/47/EC. It is possible to use CE marking to demonstrate the compliance with this EMC Directive.

It is only valid in connection with the test report number: BEL20200000101899



Manager

May. 24, 2020

This certificate of conformity is based on a single evaluation of the submitted sample(s) of the above mentioned product. It does not imply an assessment of the whole product and relevant Directives have to be observed.

Tel:+86-755-27808385, Email:Lucky@belemc.cn,  
Web:Http://www.belemc.com www.belemc.cn