

SHANDONG CHENGWU MEDICAL PRODUCTS FACTORY

Address:Southern end of Quancheng Road, Chengwu County, Heze City,Shandong Province 274200 P.R. China

Tel: +86-15102183521 Web:www.chengwumed.com www.vvzp.com

Certificate of Analysis

Product Name	Needle Holder	Specification Model	CY-A		Lot No.		20230702
Quantity	100000	Test Quantity	32pcs		Inspection Date		2023.07.02
Test items	Standard requirements			Test results			Result
Appearance	There should be no impurities, stains, lack of material, clear and accurate scale; smooth, without sharp edges and burrs.			✓ Meet the requirements □ Does not meet the requirements			✓ Passed □Unqualified □ other
Color	The color is uniform, without color difference or heterochromia.			Ø	Meet	the	✓ Passed
				requirements			□Unqualified
				□ Does not meet the			□ other
				requirements			
Size	Size The height of the needle holder shall not be less than 5.2cm, and the						✓ Passed
				Height: 5.1-5.2cm			□Unqualified
	maximum diameter shall not be greater			Maximum diameter:			□ other
	than 2.1cm. (±0.1cm)			2.0-2.1cm			
Deformation			Ø	Meet	the	☑ Passed	
	without deformation			requirements			□Unqualified
				□ Does	not meet	the	□ other
			requirements				
Package	The printing o	n the packagir	ng box	Ø	Meet	the	☑ Passed
	should meet the requirements, whether the printed handwriting is clear, the shipping mark is correct, and the			requirements			□Unqualified
				□ Does not meet the			□ other
	product quantity is correct.			requirements			
Conclusion	☑ Passed □Unqualified			\$\$\$\$ 秦君县庆 # 250			
				金米 人 學等			
Note	1 This						

Inspector: Sun Fangfang Date: 2023.07.02 Reviewer: Li Shanshan Date: 2023.07.02