



DECLARATION OF CONFORMITY

The undersigned ELMI SIA company declares that the following instrument(s) conforms the below mentioned directives and standards:

Manufacturer:	ELMI SIA Bukultu street 7b, Riga LV-1005, Latvia Phones: (+371) 67558 743, (+371) 67551 934
Equipment name:	CM-50, CM-50M, CM-50MP, CM-7S Plus, CM-8S, CM-7S, CM-70M.07, CM-70M.09, CM-6M, CM-6MT, CM-70
Equipment type:	Laboratory centrifuge
Serial number	8 digits styled XX Y ZZZZ T, where XX is year of production, Y — quarter of production, ZZZZ – batch number, T – device modification
Directives:	EMC Directive 2014/30/EU Low Voltage Directive 2014/35/EU RoHS3 2015/863/EU WEEE 2012/19/EU
Applied standards:	<u>EN 61326-1: 2013</u> Electrical equipment for measurement, control and laboratory use. EMC requirements. General requirements. <u>EN 61010-1: 2011</u> Safety requirements for electrical equipment for measurement, control and laboratory use. General requirements. <u>EN 61010-2-020: 2016</u> Particular requirements for laboratory centrifuges.

Vitalijs Mironovs
Technical Director

08.04.2022

Date

Maksims Sviridovs
Manufacturing Director

08.04.2022

Date

**ISO 13485
Certified**