

Wir

We

B. Braun Avitum AG
Schwarzenberger Weg 73-79
34212 Melsungen
Germany

erklären in eigener Verantwortung,
dass das/die Produkt/e
Kit für Dialyse und Haemo(dia)filtration
(Artikelnummern siehe Anlage I)
mit den Anforderungen der folgenden Richtlinie
übereinstimmt/übereinstimmen:

Richtlinie 93/42/EWG des Rates vom
14. Juni 1993 über Medizinprodukte

Konformitätsbewertungsverfahren:
nach Anhang II mit Ausnahme der Nummer (4)
der oben genannten Richtlinie
Klassifizierung
gemäß Anhang IX der oben genannten Richtlinie:
Klasse IIb, Regel 3
EG-Zertifikat Nr.
G1 066097 0096 Rev. 02

Konformitätsbewertungsverfahren:
nach Anhang V und Anhang VII
der oben genannten Richtlinie
Klassifizierung
gemäß Anhang IX der oben genannten Richtlinie:
Klasse I Sterile, Regel 1
EG-Zertifikat Nr.
G2S 17 08 66097 082
Benannte Stelle:
TÜV SÜD Product Service GmbH
Ridlerstraße 65, 80339 München, Deutschland
Kennnummer 0123

Datum der ersten CE-Kennzeichnung:
2015-01

Doc #: 02/15-RA-fo
Doc Rev #: 11.0
Rev date: 2020-03-04

Gültigkeit dieser Erklärung:
von 2020-03-09
bis 2023-01-09

Radeberg, 2020/03/09


Anton Deisser
Head of CoE Fluids, Concentrates and Disposables

hereby declare in our own responsibility
that the product/s
Kit for Dialysis and Haemo(dia)filtration
(article numbers see attachment I)
is/are in compliance with the following directive:

Council Directive 93/42/EEC of 14 June 1993
concerning medical devices

Conformity assessment procedure:
according to annex II excluding (4)
of the Directive named above
Classification
according to annex IX of the Directive named above:
Class IIb, Rule 3
EC Certificate No.
G1 066097 0096 Rev. 02


Conformity assessment procedure:
according to Annex V and Annex VII
of the Directive named above
Classification
according to annex IX of the Directive named above:
Class I Sterile, Rule 1
EC Certificate No.
G2S 17 08 66097 082
Notified body:
TÜV SÜD Product Service GmbH
Ridlerstraße 65, 80339 Munich, Germany
Identification number 0123

Date of first CE-marking:
2015-01

Doc #: 02/15-RA-fo
Doc Rev #: 11.0
Rev date: 2020-03-04

Validity of this declaration:
from 2020-03-09
until 2023-01-09

Mirandola, 2020.03.09


Dr. Giuliana Gavioli
Head of Division RA

Anlage I / Attachment I

Art. No.	Description	Class	Rule
7211136	OMNIset Including 0.8 sqm Hemofilter	IIb	3
7211137	OMNIset Including 1.2 sqm Hemofilter	IIb	3
7211151	OMNIset Including 1.6 sqm Hemofilter	IIb	3
7211265	OMNIset Plus Including 1.6 sqm Hemofilter	IIb	3
7211427	OMNIset® Plus	IIb	3
7211432	OMNIset® ECCO ₂ R	IIb	3
7211287	OMNIset PRO Including 1.3 sqm Hemofilter	IIb	3
7211288	OMNIset PRO Including 1.6 sqm Hemofilter	IIb	3
7211367	OMNIset® L 1.6 m ²	IIb	3
7211368	OMNIset® 0.8 m ²	IIb	3
7211369	OMNIset® 1.2 m ²	IIb	3
7211370	OMNIset® 1.6 m ²	IIb	3
7211371	OMNIset® Pro 0.8 m ²	IIb	3
7211372	OMNIset® Pro 1.2 m ²	IIb	3
7211373	OMNIset® Pro 1.6 m ²	IIb	3
7211422	OMNIset® 1.3 m ²	IIb	3
7211423	OMNIset® 1.6 m ²	IIb	3
7211428	OMNIset® 1.3 m ²	IIb	3
7211429	OMNIset® 1.6 m ²	IIb	3
7211425	OMNIset® Pro 1.3 m ²	IIb	3
7211426	OMNIset® Pro 1.6 m ²	IIb	3
7211376	OMNIset® Pro 1.3m ²	IIb	3
7211377	OMNIset® Pro 1.6 m ²	IIb	3
7211065	OMNIBag 7000 mL Effluent bag	I sterile	1

Radeberg, 2020/03/09

Anton Deisser
Head of CoE Fluids, Concentrates and Disposables

Mirandola, 2020/03/05

Dr. Giuliana Gavioli
Head of Division RA



Benannt durch/Designated by
Zentralstelle der Länder
für Gesundheitsschutz
bei Arzneimitteln und
Medizinprodukten
www.zlg.de
ZLG-BS-244.10.08



Product Service

EC Certificate

Full Quality Assurance System

Directive 93/42/EEC on Medical Devices (MDD), Annex II excluding (4)
(Devices in Class IIa, IIb or III)

No. G1 066097 0096 Rev. 02

Manufacturer:

B. Braun Avitum AG

Schwarzenberger Weg 73-79
34212 Melsungen
GERMANY

Product Category(ies): Active and non-active medical devices for extracorporeal blood treatments:
- Active medical devices for extracorporeal blood treatment: hemodialysis, acute dialysis, plasmapheresis.
Reverse osmosis Systems, central concentrate supply Systems, ring piping and hot disinfection Systems for dialysis;
- Kit for Dialysis and Haemo(dia)filtration (extracorporeal Circuit and dialyser); Lines for Dialysis and Haemo(dia)filtration;
- Kit for Plasma Treatment (Extracorporeal Circuit and Plasma filter); Lines for Plasma Treatment;
-Dialyzers, hemofilters, hemodiafilters, dialysis fluid filters, hemofilters for continuous renal replacement therapy;
- Plasma Filter Haemoselect;
- Blood filtration devices; H.E.L.P. SYSTEM kit;
- S.A.F.E. Apheresis Set;
- A.V. Fistula needles
- Catheters and Catheter Sets for Dialysis;
- Disinfectants for Dialysis Machines;
- Sterile and non-sterile hemodialysis concentrates (class IIb), solid dosage form for hemodialysis (class IIb) and irrigation Solutions (class IIa)

The Certification Body of TÜV SÜD Product Service GmbH declares that the aforementioned manufacturer has implemented a quality assurance system for design, manufacture and final inspection of the respective devices / device categories in accordance with MDD Annex II. This quality assurance system conforms to the requirements of this Directive and is subject to periodical surveillance. For marketing of class III devices an additional Annex II (4) certificate is mandatory. See also notes overleaf.

Report No.: 713168200

Valid from: 2020-02-28

Valid until: 2024-05-26

Date, 2020-02-28

Christoph Dicks
Head of Certification/Notified Body