

Certificate of Analysis

nerbe plus GmbH & Co. KG hereby certifies that the product group:

Article number	09-400-5063	09-401-5003	09-420-5063	09-421-5003
Lot number	11124	11125	11126	11126
Expiry date	01/2028	01/2028	01/2028	01/2028

are manufactured under strictly defined and monitored conditions as per product specifications. This here by certifies that the supplied product complies with following standards of production.

Quality System Compliance: The product conforms to written material specifications and was manufactured under the ISO 9001/2015 quality system.

The delivery of each batch of Calibrated Disposable Inoculating Loops are tested per Evans Blue Dye Method described by the American Society for Microbiology. The process control assures that the products Loops 1µl and Loops 10µl maintain an inner dimensional accuracy of +/- 10%.

nerbe plus is certified according to EN ISO 13485.

22.04.2024
Quality Management
nerbe plus GmbH & Co. KG

This certificate was issued digitally and is valid without a signature.