

ACON BIOTECH (HANGZHOU) CO.,LTD
CERTIFICATE OF ANALYSIS

PRODUCT NAME: HCV Antibody EIA Test Kit
PURCHASE NO.: 37740068A
BATCH NO.: 2004722
EXPIRATION DATE: 2022/02/11
QUANTITY: 50kits

SPECIFICATION		STANDARD	CONCLUSION
Physics Check		No visible deposit or floccules in solution.	Complies
Blank Well		OD < 0.100 (450nm) OD < 0.050 (450/630 nm)	Complies
Sensitivity	H (n=2)	OD ≥ 2.500	Complies
	M (n=2)	OD ≥ 1.500	Complies
	L1 (n=2)	OD ≥ 0.150	Complies
	L2 (n=2)	OD ≥ 0.150	Complies
	L3 (n=2)	OD ≥ 0.150	Complies
Negative Control		OD < 0.100	Complies
Positive Control		OD > 1.000	Complies
Coincidence of specificity standard		≥ 99/100	Complies
Precision		CV < 15%	Complies

Quality Department: Hedy Zhang

Date: Sep 23, 2020

