

STATEMENT

We, ACON Laboratories, Inc. having a registered office at *5850 Oberlin Drive #340, San Diego, CA 92121* assign SRL Sanmedico having a registered office at *A. Corobceanu street 7A, apt. 9, Chisinău, MD-2012, Moldova*, as authorized representative in correspondence with the conditions of directive 98/79/EC.

We declare that the company mentioned above is authorized to register, notify, renew or modify the registration of medical devices on the territory of the Republic of Moldova.

This authorization will be valid for one year after the date of this statement.

Date: June 1, 2021

Signature:



Qiyi Xie, Md, MPH
Sr. Officer, Regulatory & Clinical Affairs
ACON Laboratories, Inc.
Ph: 858-875-8011
Email: qxie@aconlabs.com



Benannt durch/Designated by
Zentralstelle der Länder
für Gesundheitsschutz
bei Arzneimitteln und
Medizinprodukten
www.zlg.de
ZLG-BS-245.10.07



Product Service

EC Certificate

Full Quality Assurance System
Directive 98/79/EC on In Vitro Diagnostic Medical Devices (IVDD), Annex IV excluding (4, 6)
(List A and B and devices for self-testing)

No. V1 104507 0003 Rev. 01

Manufacturer: **ACON Laboratories, Inc.**
5850 Oberlin Drive, #340
San Diego CA 92121
USA

Product Category(ies): **In Vitro diagnostics for the detection of human infections and tumor markers, blood glucose measuring self-testing systems, self-testing devices for clinical chemistry, hematology and pregnancy and ovulation**

The Certification Body of TÜV SÜD Product Service GmbH declares that the aforementioned manufacturer has implemented a quality assurance system for design, manufacture and final inspection of the respective devices / device families in accordance with IVDD Annex IV. This quality assurance system conforms to the requirements of this Directive and is subject to periodical surveillance. For marketing of List A devices an additional Annex IV (4) certificate is mandatory. See also notes overleaf.

Report no.: SH1974310

Valid from: 2019-10-24
Valid until: 2022-09-12

Date, 2019-10-24

Stefan Preiß
Head of Certification/Notified Body

TÜV SÜD
 ZERTIFIKAT ◆ CERTIFICATE ◆ 認證書 ◆ CERTIFICADO ◆ CERTIFICAT



Benannt durch/Designated by
 Zentralstelle der Länder
 für Gesundheitsschutz
 bei Arzneimitteln und
 Medizinprodukten
 www.zlg.de
 ZLG-BS-245.10.07



Product Service

EC Certificate

Full Quality Assurance System
 Directive 98/79/EC on In Vitro Diagnostic Medical Devices (IVDD), Annex IV excluding (4, 6)
 (List A and B and devices for self-testing)

No. V1 104507 0003 Rev. 01

Model(s): For Detail Models see attachment

Facility(ies):

- ACON Laboratories, Inc.
5850 Oberlin Drive, #340, San Diego CA 92121, USA
- ACON Laboratories, Inc.
10125 Mesa Rim Road, San Diego CA 92121, USA
- AZURE Institute, Inc.
10125 Mesa Rim Road, San Diego CA 92121, USA

TÜV SÜD
 ZERTIFIKAT ◆ CERTIFICATE ◆ 認證書 ◆ CERTIFICADO ◆ CERTIFIKAT ◆ CERTIFICATE ◆ CERTIFICADO ◆ CERTIFIKAT



Benannt durch/Designated by
Zentralstelle der Länder
für Gesundheitsschutz
bei Arzneimitteln und
Medizinprodukten
www.zlg.de
ZLG-BS-245.10.07



Product Service

EC Certificate

Full Quality Assurance System
Directive 98/79/EC on In Vitro Diagnostic Medical Devices (IVDD), Annex IV excluding (4, 6)
(List A and B and devices for self-testing)

No. V1 104507 0003 Rev. 01

For the product(s)/product category (ies):

On Call Plus Blood Glucose Monitoring System,
On Call Plus Blood Glucose Test Strips,
On Call EZ II Blood Glucose Monitoring System,
On Call Redi Blood Glucose Monitoring System,
On Call Redi II Blood Glucose Test Strips,
On Call Advanced Blood Glucose Monitoring System,
On Call Advanced Blood Glucose Test Strips,
On Call Platinum Blood Glucose Monitoring System,
On Call Platinum Blood Glucose Test Strips,
On Call Chosen Blood Glucose Monitoring System,
On Call Chosen Blood Glucose Test Strips,
On Call Vivid Blood Glucose Monitoring System (OGM-101),
On Call Vivid Blood Glucose Test Strips (OGS-101),
On Call Vivid Pal Blood Glucose Monitoring System (OGM-102),
On Call Sharp Blood Glucose Monitoring System (OGM-121),
On Call Sharp Blood Glucose Test Strips (OGS-121)
On Call Plus II Blood Glucose Monitoring System (OGM-171),
On Call Plus II Blood Glucose Test Strips (OGS-171),
On Call Extra Blood Glucose Monitoring System (OGM-191),
On Call Extra Blood Glucose Test Strips (OGS-191),
On Call GK Dual Blood Glucose & Ketone Monitoring System (OGM-161),
On Call Blood Ketone Test Strips (OGS-161),
D-ONE Blood Glucose Monitoring System,
D-ONE Blood Glucose Test Strips,
Urinalysis Reagent Strips (Urine),
UTI Urinary Tract Infection Test Strips,
Toxoplasma IgG EIA Test Kit,
Toxoplasma IgM EIA Test Kit,
Rubella IgG EIA Test Kit,
Rubella IgM EIA Test Kit,
CMV IgG EIA Test Kit,

Page 3 of 4

TÜV SÜD Product Service GmbH is Notified Body with identification no. 0123

TÜV SÜD Product Service GmbH • Certification Body • Ridlerstraße 65 • 80339 Munich • Germany

TÜV®



Benannt durch/Designated by
 Zentralstelle der Länder
 für Gesundheitsschutz
 bei Arzneimitteln und
 Medizinprodukten
 www.zlg.de
 ZLG-BS-244.10.08



Product Service

EC Certificate

Full Quality Assurance System
 Directive 93/42/EEC on Medical Devices (MDD), Annex II excluding (4)
 (Devices in Class IIa, IIb or III)

No. G1 104507 0002 Rev. 01

Facility(ies):

ACON Laboratories, Inc.
 5850 Oberlin Drive, #340, San Diego CA 92121, USA

ACON Laboratories, Inc.
 10125 Mesa Rim Road, San Diego CA 92121, USA

AZURE Institute, Inc.
 10125 Mesa Rim Road, San Diego CA 92121, USA

TÜV SÜD TÜV SÜD TÜV SÜD TÜV SÜD TÜV SÜD TÜV SÜD TÜV SÜD TÜV SÜD TÜV SÜD TÜV SÜD TÜV SÜD TÜV SÜD TÜV SÜD TÜV SÜD
 ZERTIFIKAT ◆ CERTIFICATE ◆ 認證書 ◆ CERTIFICADO ◆ CERTIFIKAT ◆ CERTIFICATE

Letter of Declaration

To whom it may concern:

We **Acon Laboratories, Inc.**, who is the legal manufacturer of Blood Glucose Monitoring System (Including Glucose Meter, Glucose test strip, Control Solution, Lancet and lancing device etc, to test the glucose level of human blood), have registered office at 10125 Mesa Rim Road, San Diego, CA 92121 USA, here to declare that:

- **On Call® Plus Strips** correspond with **On Call® Plus** Blood Glucose Monitoring System.
- We currently have in stock the tender required quantity of Meters, Strips and Lancets (1000/50000/50000).

This clarification letter will only be used for product registration, tender submission, sales and marketing of **On Call® Plus** Blood Glucose Monitoring System in **Moldova** it should not be used for any other business or non-business purposes.

Sincerely yours,


Eddie Shi
International Sales & Marketing Sales Manager
Diabetes Care

Acon Laboratories, Inc.





Specification

Feature	Specification
Technology	Biosensor/Electrochemical, Glucose oxidase (GOD)
Result Calibration	Plasma-equivalent
Test Time	10 seconds
Sample Size	0.5 µL
Sample Type	Fresh capillary whole blood
Hematocrit Range	25 - 60%
Glucose Test Range	20 - 600 mg/dL (1.1 - 33.3 mmol/L)
Memory Storage	300 results with date and time
Test Averaging	7, 14, 30-day averages
Data Transfer	USB
Control Solution	3 levels
Audio Feature	Optional beep for sample detection, error messages
Automatic Shutoff	2 minutes after last action
Battery	One (1) CR 2032 3.0V coin cell battery
Battery Life	1,000 measurements
Operating Conditions	41 - 113 °F (5 - 45°C) and 10 - 90% relative humidity
Strip Storage Temperature	2-35°C
Expiration Date	24 months (6 months after first opening)


Catalog

Product Name	Catalog No.	Contents			
On-Call [®] Plus Blood Glucose Monitoring System	G113-111 v †	1 Meter 1 Manual 10 Lancets	10 Test Strips 1 Carrying Case 1 Code Chip	1 Control Solution 1 1 Quick Reference Guide 1 Clear Cap (for testing on forearm and palm)	1 Lancing Device 1 Warranty Card
On-Call [®] Plus Blood Glucose Meter	G113-211 v †	1 Meter 1 Manual	1 Control Solution 1 1 Warranty Card	1 Carrying Case 1 Quick Reference Guide	
	G113-214 v	1 Meter 1 Manual 10 Lancets	1 Lancing Device 1 Carrying Case 1 Warranty Card	1 Control Solution 1 1 Quick Reference Guide 1 Clear Cap (for testing on forearm and palm)	
On-Call [®] Plus Blood Glucose Test Strips	G133-111 v †	50 Test Strips (25/vial)		1 Code Chip	1 Package Insert
		50 Test Strips (50/vial)		1 Code Chip	1 Package Insert
	G133-112 v	100 Test Strips (25/vial)		1 Code Chip	1 Package Insert
	G133-114 v	10 Test Strips (10/vial)		1 Code Chip	1 Package Insert
	G133-115 v	25 Test Strips (Individually Foil Wrapped)		1 Code Chip	1 Package Insert
	G133-117 v	50 Test Strips (Individually Foil Wrapped)		1 Code Chip	1 Package Insert
G133-118 v	25 Test Strips (25/vial)		1 Code Chip	1 Package Insert	
On-Call [®] Plus Blood Glucose Test Strips and Lancets	G133-211 v	50 Test Strips (25/vial)	50 Lancets (25/bag)	1 Code Chip	1 Package Insert
On-Call [®] Plus Blood Glucose Control Solution	G123-311 v †	1 Control Solution 0	1 Control Solution 1	1 Control Solution 2	1 Package Insert
On-Call [®] Lancets	G124-10A v †	100 Lancets (25/bag)			
On-Call [®] Lancing Device	G124-11AV	1 Lancing Device		1 Package Insert	
On-Call [®] Diabetes Management Software Kit	G124-13A †	1 USB Data Transfer Cable		1 Installation Disk	

v CE Marked for sale in the European Community 0123 † US 510(k) Cleared and CLIA Waived



On-Call[®] Plus
Blood Glucose Monitoring System




Delivers Value and Quality

- 0.5 µL Blood Sample
- Accurate & Reliable Results
- 25 - 60% HCT Range
- US 510(k) & CE




Accurate and Reliable


Extensive clinical studies proved the accuracy of *On Call Plus* Blood Glucose Monitoring System with fresh capillary blood samples, which can comply with EN ISO 15197: 2015.



Consensus Error Grid Analysis Clinical Trial - Fingertip Capillary Blood, by Technican ACON On Call Plus Blood Glucose Monitoring System vs. YSI		
System Accuracy Results for Glucose Concentration ≥ 100 mg/dL		
Within ±5%	Within ±10%	Within ±15%
290 / 462 (62.8%)	432 / 462 (93.5%)	462 / 462 (100.0%)
System Accuracy Results for Glucose Concentration <100 mg/dL		
Within ±5 mg/dL	Within ±10 mg/dL	Within ±15 mg/dL
145 / 198 (73.2%)	193 / 198 (97.5%)	198 / 198 (100.0%)
System Accuracy Results for both Glucose Concentration ≥ 100 mg/dL and < 100 mg/dL		
Within ±15% or ±15 mg/dL		
658 / 660 (99.7%)		



Consensus Error Grid Analysis Clinical Trial - Forearm Capillary Blood, by Technican ACON On Call Plus Blood Glucose Monitoring System vs. YSI		
System Accuracy Results for Glucose Concentration ≥ 100 mg/dL		
Within ± 5%	Within ± 10%	Within ± 15%
202 / 444 (45.5%)	375 / 444 (84.5%)	440 / 444 (99.1%)
System Accuracy Results for Glucose Concentration <100 mg/dL		
Within ±5 mg/dL	Within ±10 mg/dL	Within ±15 mg/dL
110 / 168 (65.5%)	154 / 168 (91.7%)	168 / 168 (100.0%)
System Accuracy Results for both Glucose Concentration ≥ 100 mg/dL and < 100 mg/dL		
Within ±15% or ±15 mg/dL		
608 / 612 (99.3%)		



Consensus Error Grid Analysis Clinical Trial - Palm Capillary Blood, by Technican ACON On Call Plus Blood Glucose Monitoring System vs. YSI		
System Accuracy Results for Glucose Concentration ≥ 100 mg/dL		
Within ±5%	Within ±10%	Within ±15%
219 / 444 (49.3%)	395 / 444 (89.0%)	441 / 444 (99.3%)
System Accuracy Results for Glucose Concentration < 100 mg/dL		
Within ±5 mg/dL	Within ±10 mg/dL	Within ±15 mg/dL
130 / 168 (77.4%)	166 / 168 (98.8%)	168 / 168 (100.0%)
System Accuracy Results for both Glucose Concentration ≥ 100 mg/dL and < 100 mg/dL		
Within ±15% or ±15 mg/dL		
609 / 612 (99.5%)		

Key Features

- 0.5 µL blood sample
- HCT 25 - 60% HCT range
- 2 - 35°C strip storage temperature
- Optional individually packaged test strips available
- Alternative testing sites including fingertip, forearm and palm
- Automatic detection of insufficient sample
- 300 test memory with date and time
- 7, 14, 30 - day averages calculation
- Easy PC data transfer and smart App data analysis

Authority Certificate



CE certificate



USFDA CFG certificate



Health Canada certificate

On-Call[®] Plus

Blood **Glucose** Monitoring System

User's Manual



Distributed by:

CLIAwaived.com[™]

11578 Sorrento Valley Road, Suite 25/26
San Diego, CA 92121
888-882-7739

On-Call[®] Plus

Blood Glucose Monitoring System

Self monitoring of blood glucose (SMBG) is an integral part of diabetes care, but the high cost of testing can make this impossible. At ACON, our goal is to provide high quality glucose monitoring at a price that allows you to test as often as necessary. Together, we can better manage your diabetes and help you live a longer and healthier life.

Welcome, and thank you for choosing the On Call[®] Plus Blood Glucose Monitoring System. The On Call[®] Plus Blood Glucose Monitoring System will give you accurate blood glucose results in just a few simple steps.

To ensure accurate results from your On Call[®] Plus Blood Glucose Monitoring System, please follow these guidelines:

- Read instructions before use.
- Use the code chip that accompanies each box of test strips.
- Use only On Call[®] Plus Blood Glucose Test Strips with the On Call[®] Plus Blood Glucose Meter.
- For *in vitro* diagnostic use only. Your blood glucose monitoring system is to be used only outside the body for testing purposes.
- For self testing and professional use.
- Test only whole blood samples with the On Call[®] Plus Blood Glucose Test Strips and Meter.
- For self-testers, consult your physician or diabetes healthcare professional before making any adjustments to your medication, diet or activity routines.
- Keep out of reach of children.
- For help with any additional questions or issues, please contact Customer Support at 1-800-838-9502.

By following the instructions outlined in this User's Manual, you will be able to use your On Call[®] Plus Blood Glucose Monitoring System to monitor your blood glucose and better manage your diabetes.

Table of Contents

En


Getting Started	1
Component Descriptions	2
<i>On Call® Plus</i> Blood Glucose Meter	3
Meter Display	4
<i>On Call® Plus</i> Blood Glucose Test Strips	5
<i>On Call® Plus</i> Glucose Control Solution	8
Install the Battery	10
Meter Setup Before Testing	11
Step 1 - Coding the Meter	11
Step 2 - Adjusting the Meter Settings	12
Performing a Quality Control Test	14
Testing Your Blood	17
Step 1 - Getting a Drop of Blood	17
Step 2 - Testing Blood Glucose	22
"HI" and "LO" Messages	24
Using the Meter Memory	26
Viewing Stored Records	26
Clearing the Memory	27
Maintenance	28
Replacing the Battery	28
Caring for Your <i>On Call® Plus</i> Blood Glucose Monitoring System	28
Suggested Testing Times and Target Goals	29
Comparing Meter and Laboratory Results	30
Troubleshooting Guide	31
Specifications	33
Warranty	34
Index	35

Getting Started

Before testing, read the instructions carefully and learn about all the components of your *On Call® Plus* Blood Glucose Monitoring System. Depending on the *On Call® Plus* product you purchase, some of the components may need to be purchased separately. Please check the list of contents on the outer box for details on which components are included with your purchase.



Blood Glucose Meter



Test Strips



Code Chip



Lancing Device



Clear Cap



Sterile Lancet



Control Solution




Carrying Case

Component Descriptions

1. Blood Glucose Meter: Reads the test strips and displays the blood glucose concentration.
2. Test Strips: Strips with a chemical reagent system used with the meter to measure glucose concentration in blood.
3. Code Chip: Automatically calibrates the meter with the code number when inserted into the meter.
4. Lancing Device: Used with sterile lancets to prick the fingertip, palm (at the base of the thumb) or forearm for blood sample collection. The packaged lancing device has multiple depth settings, allowing users to adjust the depth of the puncture and minimize discomfort.
5. Clear Cap: Used with the lancing device and sterile lancet to draw blood sample from the forearm and palm.
6. Sterile Lancets: Used with the lancing device to draw a blood sample. Sterile lancets are inserted into the lancing device with each blood draw and discarded after use.
7. Control Solution: Verifies the proper operation of the blood glucose monitoring system by checking the test strips and meter against a precalibrated control solution. Control Solution 1 is all you need most of the time. Control Solution 2 is also available if you want to do a level 2 test. The two levels of control solution, Control 1 and Control 2, are available in the *On Call® Plus* Glucose Control Solution package which is sold separately.
8. Carrying Case: Provides portability for blood glucose testing wherever you go.
9. User's Manual: Provides detailed instructions on using the blood glucose monitoring system.
10. Quick Reference Guide: Provides a brief overview of the blood glucose monitoring system and testing procedures. This small guide can be kept in your carrying case.
11. Quick Start Guide: A short set of instructions to get you started testing with your new system.
12. Logbook: Allows users to record their blood glucose data and get a better picture of their broader trends.
13. Warranty Card: Should be completed and returned to the warranty center to qualify for the 5-year meter warranty.

On Call® Plus Blood Glucose Meter

The meter reads the test strips and displays the blood glucose concentration. Use this diagram to become familiar with all the parts of your meter.



Liquid Crystal Display (LCD): Shows your test results, and helps you through the testing process.

M Button: Recalls previous test results from the meter memory and performs other menu selection functions.

S Button: Selects meter settings, performs other menu selection functions.

Strip Port: Test strips are inserted into this area to perform a test.

Battery Carrier: The battery carrier is located on the back of the meter.

Code Chip Slot: Insert the code chip here.

Code Chip: For coding the meter. A new code chip comes with every box of test strips.

Data Port: Not Currently Available for Use.


Meter Display

Pound Sign (#)

Appears with the control solution test result or when you mark an invalid result to prevent it from being included in the averages.

Control Solution Symbol

Indicates a control test result. A pound sign (#) will also be displayed when control solution symbol appears.



Meter Use and Precautions

- The meter is pre-set to display blood glucose concentration in either millimoles per liter (mmol/L) or milligrams per deciliter (mg/dL) depending on which unit of measure is standard in your country. This unit of measure cannot be adjusted. The meter will be set to mg/dL by default when sold in the United States.
 - Do not get water or other liquids inside the meter.
 - Keep the strip port area clean.
 - Keep your meter dry and avoid exposing it to extremes in temperature or humidity. Do not leave it in your car.
 - Do not drop the meter or get it wet. If you do drop the meter or get it wet, check the meter by running a quality control test. Refer to **Quality Control Test** on page 14 for instructions.
 - Do not take the meter apart. Taking the meter apart will void the warranty.
 - Refer to the **Caring for Your Meter** section on page 28 for details on cleaning the meter.
 - Keep the meter and all associated parts out of reach of children.
- Note: Follow proper precautions and all local regulations when disposing of the meter and used batteries.

On Call® Plus Blood Glucose Test Strips

The *On Call® Plus* Blood Glucose Test Strips are thin strips with a chemical reagent which work with the *On Call® Plus* Blood Glucose Meter to measure the glucose concentration in whole blood. After the strip is inserted into the meter, blood is applied to the sample tip of the test strip, then automatically absorbed into the reaction cell where the reaction takes place. A transient electrical current is formed during the reaction and the blood glucose concentration is calculated based on the electrical current detected by the meter, then the result is shown on the meter display. The meter is calibrated to display plasma equivalent results.

Sample Tip

Apply blood or control solution here.




Check Window

Check to confirm that sufficient sample has been applied.


Contact Bars

Insert this end of the test strip into the meter until it stops.

IMPORTANT: Apply sample only to the sample tip of the test strip. Do not apply blood or control solution to the top of the test strip as this may result in an inaccurate reading.




Correct




Incorrect

Hold the blood drop to the sample tip of the test strip until the check window is full and until the meter begins to count down. If the check window does not fill, do not add more blood to the test strip. You may get an E-5 message or an inaccurate test result. Discard the strip and retest. Even if the meter begins to countdown but the check window does not fill, discard the strip and begin the test again with a fresh test strip.




Correct



Incorrect

Code Number



Each package of test strips is printed with a code number (CODE), lot number (LOT), unopened expiration date (📅) and control range (CTRL 1 and CTRL 2).

Storage and Handling

Please review the following storage and handling instructions:

- Store test strips in a cool, dry place at room temperature, 59-86°F (15-30°C). Store them away from heat and direct sunlight.
- Do not freeze or refrigerate.
- Do not store or use test strips in a humid place such as a bathroom.
- Do not store the meter, the test strips or control solution near bleach or cleaners that contain bleach.
- Replace the cap on the test strip vial immediately after removing a test strip.
- The test strip should be used immediately after removing it from container.
- Do not use your test strips past the unopened expiration date printed on the label. Using test strips past the unopened expiration date may produce incorrect test results.

Note: The expiration date is printed in Year-Month format. 2008-01 means January, 2008.

Special Instructions for Test Strip in the Vial

- Test strips should be stored tightly capped in their protective vial to keep them in good working condition.
- Do not store test strips outside their protective vial. Test strips must be stored in the original vial with the cap tightly closed.
- Do not transfer test strips to a new vial or any other container.
- Replace the cap on the test strip vial immediately after removing a test strip.
- A new vial of test strips may be used for 3 months after first being opened. The opened expiration date is 3 months after the date the vial was first opened. Write the opened expiration date on the vial label after opening. Discard the vial 3 months after you first open it, usage after this period may result in inaccurate readings.

Special Instructions for Test Strip in Foil Pouch

- Tear the pouch carefully starting from the tear gap. Avoid damaging or bending the test strip.
- Use test strip immediately after removing it from the pouch.

Test Strip Precautions

- For *in vitro* diagnostic use. Test strips are to be used only outside the body for testing purposes.
- Do not use test strips that are torn, bent, or damaged in any way. Do not reuse test strips.
- Before running a blood glucose test, make sure that the code number on the meter display matches the number shown on the test strip vial or on the pouch.
- Keep the test strip vial or the foil pouch away from children and animals.
- Consult your physician or healthcare professional before making any changes in your treatment plan based on your blood glucose test results.
- Not intended for the diagnosis of or screening for diabetes mellitus.
- Not for use on neonates.
- When testing alternative sites, test when in steady state only (such as before eating, before taking medication, before exercising, or 2 hours after eating).

See the test strip insert for more details.

On Call® Plus Glucose Control Solution

The *On Call® Plus* Glucose Control Solution contains a known concentration of glucose. It is used to confirm that your *On Call® Plus* Blood Glucose Meter and Test Strips are working together properly and that you are performing the test correctly. It is important to run a quality control test regularly to make sure you are getting correct results.

You should run a quality control test:

- Before you first use your meter, to familiarize yourself with its operation.
- Before using a new box of test strips.
- When you suspect that the meter or test strips are not working properly.
- When you suspect that your test results are inaccurate, or if they are inconsistent with how you feel.
- When you suspect your meter is damaged.
- After cleaning your meter.
- At least once a week.

Refer to **Quality Control Test** on page 14 for instructions on running a quality control test.

Please review the following storage and handling instructions:



Storage and Handling

- Store the control solution at room temperature, 59-86°F (15-30°C).
 - Do not refrigerate or freeze.
 - If the control solution is cold, do not use until it has warmed to room temperature.
 - Use before the unopened expiration date that is shown on the bottle.
- Note: The expiration date is printed in Year-Month format. 2008-01 means January, 2008.
- Each bottle of control solution can be used for 3 months after you first open it. The control solution will expire 3 months after the bottle is opened for the first time. Record this opened expiration date on the bottle label.

Control Solution Precautions

- For *in vitro* diagnostic use. The control solution is for testing only outside the body. Do not swallow or inject.
- Shake well before using.
- Control solution tests are specified to be accurate only when tested between 59 and 104°F (15-40°C).
- The control ranges shown on the test strip vial (or on the foil pouch) are not recommended ranges for your blood glucose level. Your personal blood glucose target ranges should be determined by your diabetes healthcare professional.
- Do not touch the test strip with the tip of the control solution bottle.
- Use only the same brand of control solution that was provided with your kit.


Please contact Customer Support at 1-800-838-9502 for more information on obtaining the control solution kit.

See the control solution insert for more details.


Install the Battery

Battery may not be preinstalled in the meter. One CR 2032 3.0V coin cell battery is required. Please find the battery in your carrying case and install it according to following steps:

1. Pull the battery carrier on the left side of the meter. The battery carrier should be easily opened with your finger.



2. Place a new CR 2032 3.0V coin cell battery. Make sure it is aligned with the (+) side facing up in the battery carrier.



3. Close the battery carrier and make sure that it snaps shut.

Meter Setup Before Testing


Before testing, the following steps should be followed:

Step 1 - Coding the Meter

Simply insert the code chip to code the meter. Every time you change to a new box of test strips, you need to insert the code chip packed with the new box of test strips. You can see the code number appears on the meter. Make sure this number matches the code number printed on the test strip vial label (or on the foil pouch) and the number printed on the code chip.

You can easily find a code chip in your starter kit box. This code chip is used with the test strip packed in your carrying case when you first open the carrying case. If there is already one code chip inserted, remove it and insert the new code chip.

1. Take the code chip from the test strip box. Compare the code number on the code chip with the code number printed on the test strip vial label (or on the foil pouch). If the two numbers are not identical, you may get inaccurate results. If the code number on the code chip does not match the number on the vial or foil pouch of strips with which it was packaged, please contact Customer Support at 1-800-838-9502.
2. With your meter turned off, insert the new code chip into the code chip slot of the meter. It should easily snap into place. The code chip should remain in the meter, do not take it out until you change to another new box of test strips.




Note: If a test strip is inserted and no strip code is stored in memory, the display will flash "-- CODE".

Step 2 - Adjusting the Meter Settings


Adjust the meter settings to set the clock, ensuring that results stored in the memory are shown with the correct date and time. You can also turn the meter audio feature on or off. You need to adjust the meter settings before you first use your meter.

You will need to set the clock settings after replacing the battery.


1. Press the S button to enter the meter setup mode. The meter will automatically enter the setup mode when turned on for the first time by any method.
2. First, set the clock for either 12 or 24 hour mode. Press the M button to switch between the two settings, then press the S button to save your choice and start setting the year, month and date.



3. The year will appear at the top of the display. Press the M button until the correct year is displayed. Once you have selected the correct year, press the S button to save your choice and start setting the month. Press the M button until the correct month is displayed, then press the S button to save your choice and start setting the date. Press the M button until the correct date is displayed, then press the S button to save your choice and start setting the time.




4. The hour will appear at the top of the display. Adjust the hour with the M button until the correct hour is displayed. Press the S button to save your choice and set the minutes. Press the M button to change to the correct minute. Press the S button to save your choice and move to set the audio feature.



5. Audio Feature

The meter comes with the meter audio feature enabled. The meter will give one short beep when it is turned on, after sufficient sample has been applied to the test strip and when the result is ready. The meter will sound three short beeps to sound a warning when an error has occurred. Please check the error number on the display to confirm what kind of error has occurred.

Press the M button to switch between turning the meter beep "On" and "Off". Press the S button to confirm your selection. Pressing S at this point will end the setup mode and power off the meter.




Performing a Quality Control Test

The quality control test confirms that the test strips and meter are working together properly, and that you are performing the test correctly. It is important to perform this test:


- Before you first use your meter.
- Before using a new box of test strips.
- When you suspect that the meter or test strips are not working properly.
- When you suspect that your test results are inaccurate, or if they are inconsistent with how you feel.
- When you suspect your meter is damaged.
- At least once a week.

1. Insert a test strip into the strip port, contact bars end first and facing up, to turn on the meter and display all the display segments. If the audio option is on, the meter will beep, signaling the meter is turned on.




2. Check the display to confirm that all the display segments turn on (see display illustration above).

3. Following this display check, the system will enter the test mode. The display will show the date and time and the strip icon with the blood sample icon blinking. The code number will be displayed in the center of the screen. Make sure that the code number that appears on the display matches the code number (CODE) on the test strip vial (or on the foil pouch). If not, make sure to locate and insert the code chip that came with the box of strips. If the codes still do not match, do not perform a test. You will need a new package of test strips to perform a test. The blinking test strip and blood drop icon indicates that the test strip is inserted correctly and a drop of control solution can be added.



Note: If the test strip has been inserted incorrectly, the meter will not turn on.


4. Press the M button to mark the test as a control solution test. Once the M button is pressed, the control solution symbol will appear on the display.



5. Shake the control solution bottle well, then squeeze it gently and discard the first drop. If the tip clogs, tap the tip gently on a clean, hard surface, shake again, and then use. Squeeze out a second small drop on a clean nonabsorbent surface. Touch the sample tip of the test strip to the control solution drop. If the audio option is turned on, the meter will beep to indicate a sufficient sample has been applied.

Notes:

- Do not apply control solution to the test strip directly from the bottle.
 - If the control solution sample does not fill the check window, do not add a second drop. Discard the test strip and start over with a new test strip.
6. Once sufficient sample has been applied, the meter display will count down from 9 to 1 and then display the result. The control solution test results should be within the control range (CTRL1) printed on the test strip vial (or on the foil pouch). This means that your blood glucose monitoring system is working properly and that you are performing the procedure correctly.



Test results are displayed either in mmol/L or mg/dL depending on the unit of measure most common in your country. The meter will be set to mg/dL by default when sold in the United States.

7. Remove and discard the test strip.

The display should also show a pound sign (#) indicating the test is a control solution test. This shows that the number will not be counted in the 7, 14 and 30-day averages. The pound sign (#) will also be displayed when reviewing the results stored in memory.

If the result falls outside the indicated control range:

- Confirm you are matching the correct range. Control Solution 1 results should be matched to the CTRL 1 range printed on the test strip vial (or on the foil pouch).
- Check the expiration date of the test strip and control solution. Make sure that the test strip vial and control solution bottle have not been opened for more than 3 months. Discard any test strips or control solution that has expired.
- Confirm the temperature in which you are testing is between 59 and 104°F (15-40°C).
- Make sure that the test strip vial and control solution bottle have been tightly capped.
- Make sure code number on the strip vial label or on the foil pouch matches the code number appears on the meter display.
- Confirm that you are using the same brand of control solution that was provided with your kit.
- Make sure that you followed the test procedure correctly.

After checking all of the conditions listed above, repeat the quality control test with a new test strip. If your results still fall outside of the control range shown on the test strip vial (or on the foil pouch), your meter may be defective. Please contact Customer Support at 1-800-838-9502 for help.

Two levels of control solution are available labeled Control Solution 1 and Control Solution 2. Control Solution 1 is sufficient for most all self-testing needs. If you think your meter or strips may not be working correctly, you may also want to do a level 2 test. The ranges for both (CTRL 1 and CTRL 2) are displayed on the test strip vial (or on the foil pouch). Simply repeat step 4 through 6, using Control Solution 2.

For confirmation of results, Control Solution 1 tests should fall within the CTRL 1 range, and Control Solution 2 tests should fall within the CTRL 2 range. If the control solution test results do not fall within the respective ranges, DO NOT use the system to test blood, as the system may not be working properly. If you cannot fix the problem, please contact Customer Support at 1-800-838-9502 for help.

Please contact Customer Support at 1-800-838-9502 for more information on obtaining the *On Call® Plus* control solution kit, which contains Control Solution 1 and Control Solution 2.


Testing Your Blood

The following steps show how to use the meter, test strips, lancing device and sterile lancets together to measure your blood glucose concentration.

Step 1 - Getting a Drop of Blood

The *On Call® Plus* Blood Glucose Monitoring System requires a very small sample of blood which may be obtained from the fingertip, palm (at the base of the thumb) or forearm. See page 20 for information on obtaining a blood sample from the palm or forearm. Before testing, choose a clean, dry work surface. Familiarize yourself with the procedure and make sure you have all the items needed to obtain a drop of blood.



IMPORTANT: Prior to testing, wipe the test site with an alcohol swab or soapy water. Use warm water to increase blood flow if necessary. Then dry your hands and the test site thoroughly. Make sure there is no cream or lotion on the test site.




Fingertip Testing

For fingertip sampling, adjust the depth penetration to reduce the discomfort. You do not need the clear cap for fingertip sampling.


1. Unscrew the lancing device cover from the body of the lancing device. Insert a sterile lancet into the lancet holder and push it until the lancet comes to a complete stop in the lancet holder.




2. Hold the lancet firmly in the lancet holder and twist the safety tab of the lancet until it loosens, then pull the safety tab off the lancet. Save the safety tab for lancet disposal.



3. Carefully screw the cover back onto the lancing device. Avoid contact with the exposed needle. Make sure the cover is fully sealed on the lancing device.



4. Adjust the puncture depth by rotating the lancing device cover. There are a total of 5 puncture depth settings. To reduce the discomfort, use the lowest setting that still






produces an adequate drop of blood.

Adjustment:


- 1 and 2 for delicate skin
- 3 for normal skin
- 4 and 5 for calloused or thick skin

Note: Greater pressure of the lancing device against the finger will also increase the puncture depth.



5. Pull the cocking barrel back to set the lancing device. You may hear a click. The device is now loaded and ready for obtaining a drop of blood.



6. Prior to testing, wipe your hand with an alcohol swab or wash your hands with soap. Use warm water to increase blood flow in your fingers if necessary. Then dry your hands thoroughly. Massage the hand from the wrist up to the fingertip a few times to encourage blood flow.



7. Hold the lancing device against the side of the finger to be lanced with the cover resting on the finger. Push the release button to prick your fingertip. You should hear a click as the lancing device activates. Gently massage from the base of the finger to the tip of the finger to obtain the required blood volume. Avoid smearing the drop of blood.



For the greatest reduction in pain, lance on the sides of the fingertips. Rotation of sites is recommended. Repeated punctures in the same spot can make your fingers sore and callused.

Forearm or Palm (at the base of the thumb) Testing


The forearm and palm areas have fewer nerve endings than the fingertip so you may find that obtaining blood from these sites is less painful than from the fingertip. The technique for forearm and palm sampling is different. You need the clear cap to draw blood from these sites. The clear cap is not adjustable for puncture depth.

IMPORTANT: There are important differences between forearm, palm and fingertip samples that you should know. Important information about forearm and palm glucose testing:


- You should talk to your healthcare professional before doing forearm or palm glucose testing.
- When blood levels are changing rapidly such as after a meal, insulin dose or exercise, blood from the fingertips may show these changes more rapidly than blood from other areas.
- Fingertips should be used if testing is within 2 hours of a meal, insulin dose or exercise and any time you feel glucose levels are changing rapidly.
- You should test with the fingertips anytime there is a concern for hypoglycemia or you suffer from hypoglycemia unawareness.

Please refer to **Fingertip Testing** to insert the lancet and load the lancing device.


1. Screw the clear cap onto the lancing device.




2. Choose a puncture site on the forearm or palm. Select a soft, fleshy area of the forearm that is clean and dry, away from bone, and free of visible veins and hair.



To bring fresh blood to the surface of the puncture site, massage the puncture site vigorously for a few seconds until you feel it getting warm.





- Place the lancing device against the puncture site. Press and hold the clear cap against the puncture site for a few seconds. Press the release button of the lancing device, but **do not immediately lift the lancing device** from the puncture site. Continue to hold the lancing device against the puncture site until you can confirm a sufficient blood sample has formed.



Disposal of the Lancet

- Unscrew the lancing device cover. Place the safety tab of the lancet on a hard surface and carefully insert the lancet needle into the safety tab.
- Press the release button to make sure that the lancet is in the extended position. Pull the lancet straight out of the lancet holder and discard it in an appropriate container. Place the lancing device cover back on the lancing device.



Lancet Precautions

- Do not use the lancet if the safety tab is missing or loose when you take the lancet out of the bag.
- Do not use the lancet if the needle is bent.
- Use caution whenever the lancet needle is exposed.
- Never share lancets or the lancing device with other people to prevent possible infections.
- In order to reduce the risk of infection from prior use of the instrument, always use a new, sterile lancet. Do not reuse lancets.
- Avoid getting the lancing device or lancets dirty with hand lotion, oils, dirt or debris.
- For AST testing, if current lancet is not obtaining enough blood due to skin or other conditions, please contact Customer Support at 1-800-838-9502 for information on different lancet options.


Specimen collection and preparation by healthcare professionals

Please refer to test strip insert if applicable.


Step 2 - Testing Blood Glucose

Note: Insertion of a new test strip at any time, except while in the data transfer mode (detailed on page 28) will cause the meter to automatically enter the test mode.

- Insert a test strip into the strip port, contact bars end first and facing up, to turn on the meter and display all the display segments. If the audio option is on, the meter will beep, signaling the meter is turned on.




Make sure that the code number that appears on the display matches the code number (CODE) on the test strip vial (or on the foil pouch). If not, make sure to locate and insert the code chip that came with the box of strips. If the codes still do not match, do not perform a test. You will need a new package of test strips to perform a test.




- The blinking test strip and blood drop icon will indicate that the test strip is inserted correctly and a drop of blood can be added.

- Touch the blood sample to the sample tip at the end of the test strip. If the audio option is turned on, the meter will also beep to indicate the sample is sufficient and the measurement has started.



DO NOT:

- Apply sample to the front or back of the test strip.
- Smear the blood drop onto the test strip.
- Press your finger against the test strip.
- Apply a second drop of blood.




4. The meter will count down from 9 to 1 and then display the measurement results. The meter will also beep to indicate that measurement is complete.

To mark invalid results and to prevent them from being included in the 7, 14 and 30 day averages, press the M and S buttons together. A pound sign (#) will appear on the display to show that the result will not be included when calculating the 7, 14 and 30-day averages. If a result is marked by accident, press the M and S buttons again to unmark the result. After marking the invalid result, run the test again with a new test strip.

If an error message appears on the display, refer to the **Troubleshooting Guide** on page 31. If a "HI" or "LO" error appears on the display, refer to "HI" and "LO" messages below.


5. After inspection, record valid results in your logbook with the date and time, and compare them to the target goals set by your healthcare professional. Refer to **Suggested Testing Times and Target Goals** on page 29 and your logbook for more details on your target blood glucose concentration goals.
6. Remove and discard the test strip.



"HI" and "LO" Messages

The meter can accurately measure blood glucose concentrations between 1.1 to 33.3 mmol/L (20 to 600 mg/dL). "HI" and "LO" messages indicate results outside of this range.

If "HI" appears on the display, the measured concentration value is above 33.3 mmol/L (600 mg/dL). The test should be retaken to ensure that no mistake was made in the procedure. If you are certain the meter is functioning properly and no mistakes were made in the procedure, and your blood glucose is still consistently measured as "HI", it indicates severe hyperglycemia (high blood glucose). You should contact your healthcare professional immediately.



If "LO" appears on the display, the measured concentration value is below 1.1 mmol/L (20 mg/dL). The test should be retaken to ensure that no mistake was made in the procedure. If you are certain the meter is functioning properly and no mistakes were made in the procedure, and your blood glucose is still consistently measured as "LO", it may indicate severe hypoglycemia (low blood glucose). You should treat yourself for hypoglycemia immediately as recommended by your healthcare professional.



Precautions and Limitations

- The meter, test strips and other components have been designed, tested and proven to work together effectively to provide accurate blood glucose measurements. Do not use components from other brands.
- Use only with whole blood. Do not use with serum or plasma samples.
- Do not use for testing newborns.
- Do not use the meter in any manner not specified by the manufacturer. Otherwise, the protection provided by the meter may be impaired.
- Very high (above 55%) and very low (below 30%) hematocrit can cause false results. Talk to your healthcare professional to find out your hematocrit level.

- Abnormally high levels of Vitamin C (ascorbic acid), Acetaminophen, Uric Acid, L-Dopa, Tolazamide or other reducing substances will produce falsely high blood glucose measurements.
- Fatty substances (Triglycerides up to 3,000 mg/dL or Cholesterol up to 500 mg/dL) have no major effect on blood glucose test results.
- The *On Call® Plus* Blood Glucose Monitoring System has been tested and shown to work properly up to 8,516 ft (2,595 meters).
- Severely ill persons should not run the glucose test with the *On Call® Plus* Blood Glucose Monitoring System.
- Blood samples from patients in shock, or with severe dehydration or from patients in a hyperosmolar state (with or without ketosis) have not been tested and are not recommended for testing with *On Call® Plus* Blood Glucose Monitoring System.
- Dispose of blood samples and materials carefully. Treat all blood samples as if they are infectious materials. Follow proper precautions when disposing of materials.


Using the Meter Memory

The meter automatically stores up to 300 test records. Each record includes the test result, time and date. If there are already 300 records in memory, the oldest record will be erased to make room for a new one. The meter will also calculate the average values of records from the last 7, 14 and 30 days.


Viewing Stored Records

To view stored records:

1. Press the M button to turn the meter on and enter memory mode. The most recent value and the word "MEM" will appear on the display.



If you are using the meter for the very first time, the meter display will show three dashed lines (---), the word "MEM" and the unit of measure. This shows that no data have been stored in memory.



2. The date and time will be displayed together with the results stored in memory. A pound sign (#) indicates records that will be omitted from the 7, 14 and 30 day averages.
3. Press the M button to go through the stored records.
4. Press the S button to view the data averages. The words "DAY AVG" will appear on the screen.
Note: If you do not wish to view your average glucose measurements, you can press the S button again to turn off the display.

- While in memory mode, press the M button to switch between the 7, 14 and 30 day averages. The meter will calculate the average that you selected. The number of records used in the DAY AVG will also appear in the display.



- If there are fewer than 7, 14 or 30 days in memory, all the unmarked readings currently stored in memory will be averaged instead.


If you are using the meter for the very first time, no value will appear on the display. This means that no records have been stored in memory.

- Press the S button to turn off the display.
Note: Results from quality control tests will not be included in the averages. When viewing results in memory, these values are marked with a pound sign (#) to show that they will not be included in the 7, 14 and 30 day averages.

Clearing the Memory

Extreme caution should be used when clearing the memory. This is not a reversible operation. To clear the memory:

- With the meter powered off, press and hold the M button for three seconds. This will turn on the meter and enter the delete mode.
- To clear the memory, press and hold both the M and S buttons for two seconds.
- The display will show "MEM" and "---", the meter will clear its memory and after a moment turn itself off.
- If you entered the delete mode but want to exit without deleting the recorded data, press the S button. This will turn the meter off without deleting any data.



Maintenance


Proper maintenance is recommended for best results.

Replacing the Battery

When the battery icon (🔋) appears, it means the battery is running low and you should replace the battery as soon as possible. An "E-6" error message will appear if the battery is too low to perform any more blood glucose tests. The meter will not function until the battery is replaced.

Instructions:

- Make sure the meter is off before removing the battery.
- Pull the battery carrier on the left side of the meter. The battery carrier should be easily opened with your finger.
- Remove and discard the old battery. Replace it with a new CR 2032 3.0V coin cell battery. Make sure it is aligned with the (+) side facing up in the battery carrier.
- Close the battery carrier and make sure that it snaps shut.
- Recheck and reset the clock setting as necessary after battery replacement to ensure time is set correctly. To set the meter clock, see **Meter Setup Before Testing** on page 11.



Caring for Your On Call® Plus Blood Glucose Monitoring System

Blood Glucose Meter

Your On Call® Plus Blood Glucose Meter does not require special maintenance or cleaning. A cloth dampened with water and a mild detergent solution can be used to wipe the outside of the meter. Take care to avoid getting liquids, dirt, blood or control solution into the meter through the strip or data ports. It is recommended that you store the meter in the carrying case after each use. The On Call® Plus Blood Glucose Meter is a precision electronic instrument. Please handle it with care.

Lancing Device

Use mild soap and warm water to clean with a soft cloth as required. Carefully dry the device thoroughly. Do not immerse the lancing device. Please refer to the lancing device insert for more details.

Suggested Testing Times and Target Goals

Tracking your blood glucose concentration through frequent testing is an important part of proper diabetes care. Your diabetes health care professional will help you to decide the normal target range for your glucose levels. They will also help you determine when and how often to test your blood glucose. Some suggested times are:

- When you wake up (fasting level)
- Before breakfast
- 1-2 hours after breakfast
- Before lunch
- 1-2 hours after lunch
- Before or after exercise
- Before dinner
- 1-2 hours after dinner
- Before bedtime
- After a snack
- At 2 or 3 AM, if taking insulin

You may need to test more often whenever¹:

- You add or adjust your medication for diabetes.
- You think your blood glucose levels may be too low or too high.
- You are ill, or feeling uncomfortable over long periods of time.

Expected blood glucose levels for people without diabetes:²

Time	Range, mg/dL	Range, mmol/L
Fasting and Before Meals	70 - 100	3.9 - 5.6
2 Hours after Meals	Less than 140	Less than 7.8

Talk to your diabetes healthcare professional to set your own daily target ranges.

Time of Day	Your Target Range
Waking up (Fasting level)	
Before meals	
2 hours after meals	
Bedtime	
2 AM to 3 AM	
Other	

(Note: 1 mmol/L = 18 mg/dL)

Use the logbook to record your blood glucose measurements and related information. Bring the logbook with you when visiting your physician so that you can determine how well your blood glucose is being controlled. This can help you and your health care professional make the best decisions about your glucose control plan.

1. Jennifer Mayfield and Stephen Havas, "Self-Control: A Physician's Guide to Blood Glucose Monitoring in the Management of Diabetes An American Family Physician Monograph"
2. ADA Clinical Practice Recommendations, 2010.

Comparing Meter and Laboratory Results

Your *On Call® Plus* Blood Glucose Monitoring System and laboratory results both report the glucose concentration in the serum or plasma component of your blood. However, the results may differ somewhat due to normal variation. This is expected, but the difference under normal operating conditions should be no greater than 20%.

Users should make accurate comparisons periodically between meters and laboratory results. Follow the guidelines below.

Before you go to the lab:

- Bring your meter, test strip and control solution with you to the lab.
- Make sure your meter is clean.
- Perform a quality control test to make sure the meter is working properly.
- Comparisons will be more accurate if you do not eat for at least four hours (preferably eight hours) before testing.



At the lab:

- Wash your hands before obtaining a blood sample.
- Obtain blood samples for a laboratory test and for your meter within 10 minutes of each other. This will ensure an accurate comparison of results.
- Never use your meter with blood that has been placed in test tubes containing fluoride or other anticoagulants. This will cause falsely low results.

Troubleshooting Guide

The meter has built-in messages to alert you of problems. When error messages appear, note the error number, turn off the meter and then follow these instructions.

Display	Causes	Solution
Meter fails to turn on	Battery may be damaged or not be charged	Replace battery.
	Meter is too cold	If meter has been exposed to or stored in cold conditions, wait 30 minutes to allow meter to reach room temperature then repeat test.
E-0	Power On self check error	Remove battery for 30 seconds and then put battery back and turn meter on again. If problem persists, contact customer support.
E-1	Internal calibration check error	If a cell phone, radio frequency source or a high power electrical source is nearby, place more distance between the meter and any of these sources then retest. If the problem persists, contact customer support.
E-2	Test strip was removed during the test	Repeat the test and ensure test strip remains in place.
E-3	Sample was applied to the test strip too soon	Repeat test and apply sample after blood drop/test strip icon appears.
E-4	Test strip is contaminated or used	Repeat test with a new test strip.
E-5	Insufficient sample	Repeat test and apply enough sample to fill the test strip check window.
Hi L	Temperature has exceeded the operating temperature of the system	Move to a cooler environment and repeat the test.

Display	Causes	Solution
L O L	Temperature is below the operating temperature of the system	Move to a warmer environment and repeat the test.
	Battery is discharged but has enough power to run 10 more tests	Test results will still be accurate, but replace the battery as soon as possible.
E-6	Battery has discharged and meter does not allow more tests until replacement with a new battery	Replace the battery and repeat the test.
 CODE	No code chip in the meter	Insert the code chip that accompanied the box of test strips.
E-7	Damaged code chip or the code chip was removed during a test	If the code chip is damaged, use a new code chip with the correct code number and run the test. If the chip is removed during a test, confirm the code chip matches the test strip code and repeat the test.
E-8	Meter electronics failure	If the problem persists, contact customer support.
E-9	Incorrect code chip inserted in the meter	Indicates an incorrect code chip was inserted in the meter. Please make sure you use the <i>On Call® Plus</i> brand of test strips with the <i>On Call® Plus</i> Blood Glucose Meter. If the problem persists, contact customer support.
E 10	Communications failure	There is an error in transferring data to the PC. See the package insert included with the Data Management Kit for troubleshooting.

For help with any additional questions or issues, please contact Customer Support at 1-800-838-9502.

Specifications

Feature	Specification
Measurement Range	20 to 600 mg/dL
Result Calibration	Plasma-equivalent
Sample	Fresh capillary whole blood
Minimum Sample Size	1 µL
Test Time	10 seconds
Power Source	One (1) CR 2032 3.0V coin cell battery
Battery Life	12 months or approximately 1,000 tests
Glucose Units of Measure	The meter is pre-set to either millimoles per liter (mmol/L) or milligrams per deciliter (mg/dL) depending on the standard of your country. The meter will be set to mg/dL by default when sold in the United States.
Memory	Up to 300 records with time and date
Meter Size	85mm x 54mm x 20.5mm
Display Size	35mm x 32.5mm
Weight	Approximately 49.5 g (with battery installed)
Operating Temperature	41 - 113°F(5-45°C)
Operating Relative Humidity	20-90% (non-condensing)
Hematocrit Range	30-55%

Warranty

Please complete the warranty card that came with this product and mail it to the following address:

On Call® Plus Warranty Center
10125 Mesa Rim Road,
San Diego, CA 92121-2915, USA

If the meter fails to work for any reason other than obvious abuse within the first five (5) years from purchase, we will replace it with a new meter free of charge. For your records, also write the purchase date of your product here.

Date of purchase: _____

Note: This warranty applies only to the meter in the original purchase, and does not apply to the battery supplied with the meter.

Averaging Results	26	Procedure	
Battery, Replacing the	28	Precautions and Limitations.....	24
Carrying Case	1	Testing Your Blood.....	17
Clearing the Memory	27	Quality Control Test	14
Data Port	3	Control Solution.....	8
Date Format	7, 9	How to Run.....	14
Guidelines	i	Results	
Hematocrit	24, 33	Blood Glucose.....	23
Install the Battery	10	Control Solution.....	15
Lancing Device	1, 17, 28	Meter vs. Lab Results.....	30
Lancets	1, 17	Target Goals.....	29
Logbook	2, 29	Unit of Measure.....	15, 29
Maintenance and Disposal ..	5, 28	Suggested Testing Times	29
Measure, Unit of	4, 15	Test Strip	5
Meter	3	Code.....	6
Audio feature.....	13	Expiration.....	7
Cleaning.....	28	Precautions.....	8
Code chip.....	2	Testing Your Blood	17
Display.....	4	Get a Drop of Blood.....	17
Error Messages.....	31	Test Blood Glucose.....	22
"HI" and "LO" Messages.....	24	Troubleshooting	31
M Button.....	3	Viewing Stored Records	26
Memory.....	26	Warranty	2, 34
Meter code.....	11		
Meter Setup.....	11		
Meter Use and Precautions.....	5		
S Button.....	3		
Set the Clock.....	12		
Specifications.....	33		

CUPRINS

Informații importante cu privire la sistemul de monitorizare a glicemiei On Call Plus	2
Glucometrul On Call Plus și Banda de Testare On Call Plus	
Glucometrul On Call Plus modul de funcționare	4
Afisajul Glucometrului On Call Plus	6
Banda de testare On Call Plus	7
Soluția de control On Call Plus	8
Instalarea bateriei	8
Configurarea glucometrului	8
Etapa 1 Introducerea № de cod	9
Etapa 2 Setarea parametrilor necesare	9
Semnal Sonor	11
Realizarea testului cu Soluție de control	11
Teste de sânge	15
Etapa 1 Obținerea picăturii de sânge	16
Etapa 2 Realizarea unui test de glicemie	20
Mesajele "HI" și "LO"	21
Utilizarea sistemului On Call Plus	22
Memoria Glucometrului	23
Vizualizarea rezultatelor salvate	23
Curățarea memoriei	24
Întreținerea etnică	25
Remediarea defectiunilor	26
Specificatii tehnice	28

Cum trebuie sa utilizati sistemul On Call Plus

Sistemul este realizat cu intenția de a fi utilizat în exteriorul corpului (cu alte cuvinte, numai diagnosticare în vitro)

Sistemul este utilizat pentru măsurarea glicemiei (nivelul zaharului în sânge) din probe prelevate din braț, antebraț , mână , coapsă , gambă , sau degete)

Glucometrul On Call Plus trebuie fi utilizat numai împreună cu soluția de control On Call Plus, cu benzile de testare On Call Plus. Utilizarea altor mărci de benzi de testare, soluții de control, împreună cu acest glucometru aduc la rezultate eronate.

Utilizați Glucometrul On Call Plus pentru:

- Testarea glicemiei.
- Testarea numai cu probe de sânge curente.

Nu utilizați glucometrul On Call Plus pentru:

- Diagnosticarea diabetului.
- Testarea nou nascutilor.
- Testarea sângelui arterial.

Avertisment: Vă rugăm să citiți toate instrucțiunile oferite în acest Manual de Utilizare și exersați procedurile de testare înainte de utilizarea sistemului On Call Plus .
Monitorizarea glicemiei trebuie fi făcută sub îngrijirea medicului.



Glucometru



Banda de testare



Cip



Dispozitivul de înțepare



Capacul




Lanțetă



Soluție de control



Glucometrul On Call Plus și Modul în care funcționează



Displayul LCD Afișează rezultatele testului dvs. și alte informații importante

Butonul S este folosit pentru reamintirea rezultatelor testelor anterioare și efectuează alte funcții de selecție din meniu.

Butonul M este folosit pentru pornirea glucometrului și efectuează alte funcții de selecție din meniu.

Portul Benzii de Testare Inserați aici capătul Părții Superioare a noii Benzii de Testare On Call Plus

Înlocuirea bateriei Instalați baterie îndreptat spre dvs.

Numărul bateriei CR 2032 3.0V.

USB Port Transmite informațiile la calculator prin cablu suplimentar, pentru a vizualiza, analiza și imprima rezultatele salvate. Cablu trebuie fi comandat separat.


Afișajul Glucometrului On Call Plus

Simbolul


Afișează faptul că ați primit rezultat cu soluție de control al testului, sau dacă doriți să îndepărtați rezultatul greșit din valorile medii

Simbolul Rezultatului Testului cu Solutia de control

Afișează faptul că ați marcat un rezultat de testare cu soluția de control al testului. Simbolul soluției de control apară împreună cu simbolul #




Banda de Testare On Call Plus




Aplicați soluția de control sau picătura de sânge pe vârful corespunzător ale Benzii de Testare.

Nu aplicați sânge pe ambele suprafețe ale benzii de testare.

Atingeți cu picătura de sânge locul de aplicare pe banda de testare până când Fereastra de control se va umple complet.




CORECT




GREȘIT

În cazul în care fereastră de control nu s-a completat, nu aplicați o cantitate suplimentară de sânge pe banda de testare. Acest lucru poate aduce la rezultate incorecte. Îndepărtați banda de testare și efectuați testarea din nou.



CORECT



GREȘIT

Benzele de testare pot fi folosite doar o data. Descarcați benzile de testare folosite. Continuați să umpleți banda de testare în timp de până la 60 de sec.

AVERTISMENT: Capacul sau flaconul cuprinde agenți de uscare care pot fi dăunatori dacă sunt inhalați sau înghițiți și pot distruge pielea sau produce iritații ale ochilor.

Informatii Importante privind Banda de Testare

- Sistemul On Call Plus trebuie pastrat în loc uscat, la temperatura de 15-30° C
- Feriți de acțiunea directă a razelor de soare și călduri
- Depozitați Benzile de testare doar în flaconul lor original. Nu le transferați într-un alt flacon sau recipient
- Nu depozitați niciodată Benzile de Testare individuale în afara flaconului
- După ce ați luat o nouă bandă de testare din flaconul său, închideți imediat capacul flaconului
- Nu folosiți benzile de testare după data de expirare a acestora. Aceasta poate determina rezultate eronate
- Nu îndoiiți, tăiați sau modificați forma Benzii de Testare

Soluția de control On Call Plus

Soluția de control On Call Plus este un lichid care conține o cantitate determinată de glucoză care este folosită pentru a va:

- Asigura că Glucometru dumneavoastră și benzile de testare funcționează corespunzător
- Dacă nu sunteți siguri, că glucometru și benzile de testare indică rezultatul corect
- Înainte de a utiliza banda de testare din flaconul nou
- După curățarea glucometrului dvs.

Soluția de control On Call Plus este valabilă timp de 3 luni de la deschiderea recipientului. Calculați perioada de 3 luni în avans față de data la care ați deschis noul recipient cu soluția de control. Aceasta reprezintă data la care trebuie să aruncați recipientul. Notați această data pe partea laterala a recipientului cu soluția de control.

Informații importante privind Soluția de control

- Folosiți numai pentru diagnosticarea *in vitro*. Nu folosiți pentru băut sau pentru injectare.
- Înainte de a folosi agitați flaconul
- Nu adaugați apa sau orice alt fel de lichid în soluția de control
- Se specifică faptul că testele pentru SOLUȚIA DE CONTROL prezintă rezultate corecte doar când sunt folosite la temperaturi între 15° - 40° C

- Rezultatele testelor pentru Soluția de control nu reflectă nivelul glicemiei dvs. Rezultatele testului pentru soluția de control trebuie să se încadreze în valorile indicate pe eticheta recipientului Benzei de testare.
- Folosiți doar soluția de control On Call Plus împreună cu glucometrul On Call Plus

Instalarea bateriei


Aveți nevoie de o baterie cu numărul CR 2023 3.0V.

1. Scoateți suportul pentru baterie din partea stânga a glucometrului.
2. Instalați în suport o baterie nouă cu semnul pozitiv (+) îndreptat spre dvs.
3. Închideți suportul pentru baterie.

Configurarea glucometrului înainte de începerea testării

ETAPEA 1- Introducerea numărului de cod

Pentru codificarea glucometrului instalați cipul pe care îl veți găsi în flaconul nou. Fiți pregătit de a ajusta codul de pe ecranul Glucometrului astfel încât să se potrivească cu numărul de cod de pe flaconul Benzii de testare. Asigurați-vă că numărul de cod de pe cip se potrivește cu numărul de cod de pe partea laterala a flaconului al Benzii de testare.




Notă: Dacă ați introdus banda de testare în Glucometrul fără cip, pe ecran va apărea inscripția ”CODE”

Etapa 2 Setarea parametrilor necesare


Setați ora și data pentru a vizualiza rezultatele, stocate în memoria glucometrului, cu timpul și data corectă .

1. Glucometrul dvs. după pornirea pentru prima dată va intra automat în regimul de configurarea. Tastați butonul "S"



2. În primul rând setați ceasul în 12- sau 24- format de timp. Tastați butonul "M" pentru a alege formatul dorit, apoi tastați butonul "S" pentru a salva rezultatul și a continua prin setarea anului, lunii și datei curente.





3. Pe partea de sus a ecranului va apărea anul. Tastați butonul "M" până când veți vedea valoarea corectă a anului. După alegerea anului tastați butonul "S" pentru a salva rezultatul, apoi începeți setarea datei. Tastați butonul "M" până când veți vedea valoarea corectă a datei. După alegerea datei tastați butonul "S" pentru a salva rezultatul, apoi începeți setarea orei.



4. Pe partea de sus a ecranului va apărea ora. . Tastați butonul "M" până când veți vedea valoarea corectă a orei. După alegerea orei tastați butonul "S" pentru a salva rezultatul, apoi începeți setarea minutelor.




SAU



SAU

Semnal sonor

Glucometru are semnal sonor. După pornirea glucometrului, după ce ați aplicat pe banda de testare o cantitate suficientă de probă, la primirea rezultatul final - veți auzi un semnal sonor scurt. După o eroare Glucometrul va emite trei semnale sonore scurte. Verificați codul de eroare de pe ecran pentru a determina cauza.




Tastați butonul "M" pentru a activa sau dezactiva semnalul sonor. Tastați butonul "S" pentru a salva alegerea dvs. După tastarea butonului "S" glucometrul se va opri.

Efectuarea unui Test cu Soluția de Control

Test de control al calității confirmă că Benzile de Testare și Glucometrul lucrează împreună într-un mod corespunzător. Efectuați un Test cu Soluția de control când:

- Utilizați Glucometrul pentru prima dată
- Înainte de a lucra cu benzile de testare dintr-un nou flacon
- Dacă nu sunteți siguri că Glucometrul sau Benzile de Testare funcționează în modul corespunzător
- Cel puțin o dată pe săptămână

1.Înserați partea superioară a Benzii de testare în Glucometrul dvs. Glucometrul va porni




2. Verificați dacă toate segmentele sunt afișate pe ecran

3. După aceasta Glucometrul dvs. va intra în modul de testare. Așteptați ca Glucometrul să vă afișeze simbolurile pentru data, ora, picătura de sânge și Banda de Testare. În mijlocul ecranului va apărea numărul de cod. Numărul de cod de pe ecran trebuie să se potrivească cu numărul de cod de pe flaconul Benzii de Testare. Clipirea simbolurilor de Banda de Testare și picătura de sânge va indica faptul că banda de testare este amplasată corect și este posibil să aplicați soluția de control. Glucometrul nu va porni în cazul în care banda de testare nu va fi introdusă corect.



4. Tastați butonul "M" pentru a marca acest test ca un test de control. După tastarea butonului "M" pe ecran va apărea simbolul Soluției de Control.




5. Agitați recipientul cu soluția de control. Picurați o picătură de soluție pe o suprafață neabsorbantă. Atingeți încet vârful de banda de testare în picătura de soluție de

control. În cazul în care funcția mesajului sonor este activă, Glucometrul va produce un mesaj sonor când banda de testare va fi plină.

Notă:

- Nu aplicați Soluția de Control din recipientul direct pe Banda de Testare
- Dacă Soluția de Control nu va umple complet gaura de testare, nu adăugați a doua picătură. Îndepărtați banda de Testare și începeți un nou test cu o nouă bandă de testare.

6. Atunci când o cantitate suficientă de soluție va fi aplicată, Glucometrul dvs. va începe numărarea de la 9 până la 1, după ce vă va arăta rezultatul. Comparați rezultatele testului cu soluția de control, cu valorile indicate pe eticheta flaconului pentru banda de testare. Rezultatele dvs. pentru soluția de control trebuie să se încadreze în valorile indicate pe partea laterală a flaconului pentru banda de testare.



Rezultatele de testare sunt afișate în mmol/L sau mg/dl

7. Dezinstalați și aruncați banda de testare.

După testare, pe ecranul dvs. trebuie să apară simbolul #, el vă va arăta că testarea a fost efectuată cu Soluția de Control.

Rezultatele testului care nu se încadrează în valorile CTRL1 pot fi determinate de:


- O soluție de control sau banda de testare expirată.
- Testul pentru Soluția de control a fost realizat în afară regimului nominal de temperatură, cuprins între valori de la 15° la 40° C
- Codul de pe flaconul pentru Banda de Testare nu corespunde cu codul inițializat pe ecranul Glucometrului
- Asigurați-vă că comparați rezultatele testului cu valorile CTRL1
- Verificați dacă ați efectuat corect testarea

Dacă ați făcut testarea în mod corect, dar rezultatele întotdeauna depășesc spre numărul seriei de control CTRL1, contactați distribuitorul dvs.

Teste de sânge

Etapa 1


Pentru a obține rezultatul cu folosirea sistemii On Call Plus aveți nevoie de o mică picătură de sânge, pe care le puteți obține din vârfurile degetelor, mâna sau braț. Înaintea testării alegeți o suprafață curată și uscată. Pentru a asigura rezultatele precise, spălați-vă pe mâini cu apa și săpun. Asigurați-vă că nu există urme de cremă sau loțiune pe suprafață testată.




Cum să obțineți picătura de sânge din vârful degetului

Setați adâncimea înțepării pentru a reduce disconfortul.


1. Îndepărtați capacul de protecție de pe dispozitivul de înțepare. Înserați o nouă lanțetă stirilă On Call Plus în dispozitivul de înțepare.




2. Mențineți cu fermitate Dispozitivul de înțepare, apoi rotați tip de protecție de la lanțetă și eliminați-l.



3. Răsuciți cu atenție capacul de protecție. Aveți grijă să nu atingeți acele expuse de pe dispozitivul de înțepare. Asigurați-vă că capacul de protecție acoperă complet dispozitivul de înțepare.



4. Dispozitivul de înțepare On Call Plus vă oferă 5 nivele diferite de setări. Pentru a reduce disconfortul, utilizați cea mai mică setare pentru a obține o picătură de sânge.



Nivele de setări:


1 și 2 pentru pielea sensibilă

3 pentru pielea normală


4 și 5 pentru pielea aspră

Nota: O mare presiune pe deget mărește adâncimea înțepării.


5. Trageți mânerul de îmbinare până când veți auzi un clic. Acum sunteți pregătiți să realizați testul de glucoză.



6. Înaintea testării spălați-vă mâinile cu săpun. Uscați-vă mâinile. Pentru a aduce sângele la suprafața locului de testare, frecați locul de testare pentru câteva secunde până când simțiți ca se încălzește.




7. Țineți capacul transparent pe suprafața de testare. Apăsați butonul de declanșare. Ridicați dispozitivul de înțepare. Aveți grijă să nu întindeți proba de sânge de pe suprafața de testare.




Cum să obțineți picătura de sânge din mână sau braț

Important: Dacă doriți să luați o proba de sânge din mână sau din braț trebuie să fiți conștienți, că rezultatele pot fi diferite. Pentru a face testarea în termen de 2 ore după masă sau după introducerea insulinei, folosiți vârful degetului pentru a obține un rezultat precis.

1. Efectuați pașii din secțiunea 1 și 2 ” **Cum să obțineți picătura de sânge din vârful degetului**”
2. Răsuciți cu atenție capacul transparent în locul capacului de protecție.




3. Trageți mânerul de îmbinare până când veți auzi un clic. Acum sunteți pregătiți să realizați testul de glucoză.
4. Întindeți-vă mâna sau brațul. Evitați zonele cu vene, alunite proeminente , zonele cu oase sau cu tendoane.




Pentru a aduce sângele la suprafața locului de testare, frecați locul de testare pentru câteva secunde până când simțiți ca se încălzește.

5. Țineți capacul transparent pe suprafața de testare. Apăsați butonul de declanșare. Nu ridicați. Continuați să țineți dispozitivul pe suprafața de testare pentru a obține cantitatea suficientă de sânge.




Îndepărtarea lanțetei

1. Scoateți capacul dispozitivului de înțepare cu lanțetă. Introduceți acul lanțetei în tip de protecție.
2. Tastați butonul de declanșare și asigurați-vă că lanțetă se află în starea descărcată. Scoateți lanțetă din dispozitivul de înțepare. Acoperiți-vă dispozitivul de înțepare cu capac de protecție.




Efectuarea unui test de glicemie

1. Înserați o nouă bandă de testare în Glucometru până când aceasta se oprește. Dacă funcția semnalului sonor este activată, atunci veți auzi un semnal sonor scurt. Acest semnal va informa că dispozitivul e pornit. Asigurați-vă că numărul de cod de pe cip coincide cu numărul de cod de pe partea laterală al flaconului Benzii de testare.



2. Clipirea simbolurilor benzii de testare și picăturii de sânge vă va renunța că Banda de Testare este inserată corect și puteți să aplicați picătura de sânge. În cazul în care banda de testare este introdusă incorect, dispozitivul nu va porni
3. Poziționați Banda de Testare ușor înclinată pe mostra de sânge. Dacă funcția semnalului sonor este activată, atunci veți auzi un semnal sonor care vă va informa că testarea a început.



Nota:

- Nu apăsați Banda de Testare pe zona de testare
 - Nu aplicați nici o picătură de sânge pe partea laterală a Benzii de Testare
 - Nu radeți sânge pe banda de testare
 - Nu aplicați sânge pe Banda de Testare atunci, când banda de Testare se află în afara Glucometrului.
 - Nu aplicați a doua picătura de sânge pe Banda de Testare
4. Dispozitivul va începe numărarea de la 9 până la 1, apoi vă va arăta rezultatul. Pentru a marca rezultatele greșite și pentru a le exclude din medie de 7, 13, și 30 de zile, tastați simultan butoanele **M** și **S**. Pe ecranul va apărea simbolul #

care va arată că acest rezultat nu va fi luat în considerare din medie de 7, 13 și 30 de zile. Dacă rezultatul este marcat în mod accidental, tastați din nou pe butoane **M** și **S** pentru a elimina semnul. Repetați testarea cu o nouă banda de testare.

Dacă pe ecranul dvs. va apărea un mesaj de eroare, adresați-vă pe pagina 25. Ghidul de îndepărtarea defectelor. Dacă pe ecranul va apărea eroarea de felul "HI" sau "LO", uitați-vă în secțiune "HI" și "LO" de mai jos.



5. După primirea rezultatelor, notați-vă în agenda dvs. rezultatele corecte, cu implicarea datei și orei, apoi comparați-le cu indicatorii valizi, stabiliți de către medicul dumneavoastră.
6. Extrageți și aruncați Banda de Testare .


Mesajele "HI" ("Ridicat") și "LO" ("Scăzut")

Glucometrul dvs. măsoară concentrația de glucoză de între 1.1 până la 33.3 mmol/L (de la 20 până la 600mg/dl). Mesaje "HI" și "LO" indică rezultatele în afară acestui interval.

Dacă pe ecranul dvs. va apărea mesajul "HI", valoarea concentrației măsurate este mai mare decât 33.3 mmol/L (600 mg/dL). Repetați testarea pentru a avea siguranța că ați făcut testarea în mod corespunzător. Dacă sunteți siguri că dispozitivul dvs. funcționează corect, dar după testarea repetată pe ecranul dvs. apare mesajul "HI", această indică un nivel ridicat de glucoza în sânge. Neapărat adresați-vă medicului dvs.



Dacă pe ecranul dvs. va apărea mesajul "LO", valoarea concentrației măsurate este mai mică de 1.1 mmol/L (20mg/dl). Repetați testarea pentru a avea siguranța că ați făcut testarea în mod corespunzător. Dacă sunteți siguri că dispozitivul dvs. funcționează corect, dar după testarea repetată pe ecranul dvs. apară mesajul "LO", această indică un nivel scăzut de glucoza în sânge. Neapărat adresați-vă medicului dvs.



Cum trebuie sa utilizati sistemul On Call Plus

Sistemul este realizat cu intenția de a fi utilizat în exteriorul corpului (cu alte cuvinte, numai diagnosticare în vitro)

Sistemul este utilizat pentru măsurarea glicemiei (nivelul zaharului în sânge) din probe prelevate din braț , antebraț , mână , coapsă , gambă , sau degete)

Glucometrul On Call Plus trebuie fi utilizat numai împreună cu soluția de control On Call Plus, cu benzile de testare On Call Plus. Utilizarea altor mărci de benzi de testare, soluții de control, împreună cu acest glucometru aduc la rezultate eronate.

Utilizați Glucometrul On Call Plus pentru:

- Testarea glicemiei.
- Testarea numai cu probe de sânge curente.

Nu utilizați glucometrul On Call Plus pentru:

- Diagnosticarea diabetului.
- Testarea nou nascutilor.
- Testarea sângelui arterial.

Memoria Glucometrului

Glucometrul dvs. păstrează în memoria sa 300 de înscrieri. Fiecare înscriere include în sine rezultatul testării, data și ora. Dacă în memoria sunt acumulate mai mult decât 300 de teste, cea mai veche înregistrare va fi ștearsă automat pentru a face loc pentru nouă.

Glucometrul dvs. va calcula rezultatele medii pentru ultimele 7, 14 și 30 zile.


Vizualizarea rezultatelor salvate

Pentru a vizualiza rezultatele salvate:


1. Tastați butonul **M** pentru a porni dispozitivul și pentru a intra în modul de memorie. Pe ecranul dvs. va apărea ultima înregistrare sau cuvântul "MEM".



Dacă folosiți dispozitivul pentru prima dată, pe ecranul dvs. va apărea trei linii punctate(---), cuvântul "MEM" și unitatea de măsură. Aceasta înseamnă că nu aveți în memorie rezultatele stocate.



2. Data și ora va apărea pe ecranul împreună cu rezultatele care sunt stocate în memorie. Simbolul # pe ecranul dvs. va indica faptul că înregistrările marcate nu sunt luate în considerare la mediile de 7, 14 și 30 zile.
3. Pentru a vizualiza rezultatele stocate tastați butonul M.
4. Pentru a vizualiza rezultatele medii tastați butonul S. Pe ecran vor apărea cuvintele "DAY AVE". Dacă nu doriți să vizualizați rezultatele medii nivelului glucozei, tastați din nou pe butonul S, pentru a opri ecranul.
5. În timp ce vă aflați în regimul memoriei, tastați butonul M pentru a muta prin mediile de 7, 14 și 30 zile. Dispozitivul dvs. va calcula media selectată. Numărul de înregistrări, folosite în mediile DAY AVE, va fi indicat pe ecranul dispozitivului.




6. Dacă numărul rezultatelor stocate în memorie glucometrului este mai puțin decât 7, 14 sau 30 de zile, în locul lor va fi mediate toate rezultatele nemarcate în prezent stocate în memoria.
Dacă folosiți glucometrul pentru prima oară, pe ecranul dvs. nu va apărea nici o indicație. Aceasta va însemna că nu aveți nici un rezultat înscris.
7. Tastați butonul S pentru a opri ecranul.

Curățarea memoriei

Fiți foarte atenți când începeți curățarea memoriei. Aceasta operațiune este ireversibilă. Pentru a curăța memoria procedați în felul următor:


1. Tastați butonul M în timp ce dispozitivul este deconectat și țineți-l timp de 3 secunde. Dispozitivul va intra în regimul curățării.



2. Pentru a curăța memoria tastați simultan butoanele M și S și țineți-le timp de 2 secunde.
3. Pe ecranul va apărea "MEM" și "---", memoria glucometrului va fi eliminată și după un timp el se va dezactiva.
4. Dacă ați intrat în regimul curățării memoriei dar nu doriți să eliminați rezultatele, tastați butonul S. Glucometrul va dezactiva fără îndepărtarea rezultatelor.

Întreținerea tehnică

Bateriile

Glucometrul este livrat cu o baterie CR 2032 3, 0V. În cazul, în care bateria dvs. nu asigură suficientă energie, simbolul bateriei () apare pe ecranul până când schimbați bateria. Dacă pe ecranul va apărea simbolul "E-6", dispozitivul nu va lucra până când nu veți schimba bateria.

Instrucțiuni:

1. Înainte de a schimba bateria opriți glucometrul.
2. Apăsați clapeta pentru baterie pentru a o deschide.
3. Îndepărtați bateria veche. Instalați baterie nouă cu semnul pozitiv (+) îndreptat spre dvs.



4. Înclinați clapeta bateriei înapoi.

5. După ce ați schimbat bateria reinstalați ora. Pentru instalarea orei vedeți punctul **Setarea dispozitivului înainte de testare** pe pagina 10.

Îngrijirea glucometrului dvs.



Pentru a evita pătrunderea murdăriei, prafului, sângelui, a soluției de control, apei sau a oricărui alt lichid pe banda de testare a glucometrului și în porturile de date, curățați suprafața exterioară a Glucometrului folosind o cârpă îmbibată cu una dintre următoarele soluții:

- Detergent/săpun lichid și apă
- 70% alcool isopropil

Nu scufundați niciodată Glucometrul în apă sau în orice altă soluție lichidă.

Remediarea defecțiunilor

ECRANUL	CAUZA	ELIMINAREA
Dispozitivul nu pornește	Bateria este descărcată sau deteriorată	Schimbați bateria
	Dispozitivul este foarte rece	Dacă dispozitivul a fost menținut la temperaturile scăzute, așteptați 30 de minute până când o să ajungă la temperatura camerei.
E-0	Dispozitivul pornit, eroare de autocontrol	Îndepărtați bateria pe timp de 30 de secunde, apoi introduceți-o înapoi și porniți dispozitivul
E-1	Eroarea de calibrarea internă	Dacă alături de dispozitivul dvs. se află un telefon mobil, sursa de unde radio, poziționați pe dispozitivul într-un alt loc, apoi repetați testarea
E-2	În timpul testării	Repetăți testarea și sigurați-vă că banda

	banda de testare era eliminată	de testare se află în locul său
E-3	Poba de testare era aplicată prea devreme	Repetăți testarea și aplicați proba de testare după ce pe ecranul dvs. va apărea simbolurile "picătura de sânge/banda de testare"
E-4	Bande de testare este contaminată sau utilizată	Repetăți testarea cu o nouă banda de testare
E-5	Insuficiența probei de testare	Repetăți testarea și aplicați pe banda de testare o cantitate suficientă de probă
HI .t	Temperatura ambianței a depășit temperatura obișnuită de lucru a dispozitivului	Deplăsați-vă într-un loc răcoros
LO .t	Temperatura ambianței e mai jos decât temperatura obișnuită prin care dispozitivului lucrează	Deplasați-vă într-un loc mai călduros
	Bateria se descarcă	Rezultatele testelor totuși va fi corecte, dar trebuie să înlocuiți bateriile
E-6	Bateriile sunt consumate	Înlocuiți bateria și repetați testarea
	Cipul nu este instalat	Instalați cipul
E-7	Cipul este deteriorat sau îndepărtat în timpul testării	Dacă cipul este deteriorat, utilizați unul nou cu codul corect. Dacă cipul a fost îndepărtat în timpul testării, instalați cipul și repetați testarea.

E-8	Eroare dispozitivului	Dacă problema persistă, contactați distribuitorul dvs.
E-9	Cipul instalat nu aparține Acon Labs On Call Plus	Asigurați-vă, că folosiți banda de testare On Call Plus. Dacă problema persistă, contactați distribuitorul dvs.
E-10	Eroare de comunicare	Pentru a soluționa problema contactați distribuitorul dvs.

Specificațiile tehnice

Caracteristica	Descriere
Limita rezultatelor	1.1-33.3 mmol/L (de la 20 până la 600mg/dL)
Calibrare	Echivalent plasma
Proba	Sînge integral capilar
Volumul minim de probă	1µl
Timp de testare	10 secunde
Sursa de alimentare	O baterie CR 2032 3.0V
Durată bateriei	12 luni sau 1000 de testări
Unitatea de măsură	mmol/L sau mg/dl
Memorie	300 de teste cu data și ora
Oprirea automată	După expirarea 2 minute după ultima acțiune a utilizatorului
Dimensiunile dispozitivului	88mm x 54mm x 20.5mm
Dimensiunea Ecranului	35mm x 32.5mm
Greutate	49.5 gr (cu bateria inclusă)
Temperatura de operare	5-45° C (41-113° F)
Hematocrit	30-55 %