



CERTIFICATE OF QUALITY ANALYSIS

Description: povidone-iodine prep pad, quantity: 120,000pcs

packing: 100pcs/box,100box/ctn, 4ctns, Total: 4ctns,

Lot No.: 201811 MFG Date: 201811, Expired Date: 202311

specification: Bag size: $50 \pm 2\text{mm} \times 50 \pm 2\text{mm}$

Non-woven materials : $65 \times 30\text{mm}$, $70 \pm 2\text{g}/\text{M}^2$

povidone-iodine: 10%, weight: $0.45 \pm 0.05\text{g}$

FINAL INSPECTION RESULT: Samples qty: 150pcs

No	Inspection items	Standard requirements	Defects quantity	result	Testing Instruments No.
1	Packing	According to Contract.	0	OK!	By visual Inspection
2.	Printing quality	Nice print, very clear	0	OK!	By visual Inspection
3	Size of bag, non-woven.	$50 \pm 2\text{mm} \times 50 \pm 2\text{mm}$ $2\text{g}/\text{M}^2$	0	OK!	Digital scale Vernier caliper
4.	Weight of 10% povidone-iodine	$0.45 \pm 0.05\text{g}$	0	OK!	Digital scale
5	Cutting of sides	No obviously burr	0	OK	By visual Inspection

Final Result: acceptable

Inspection: ShenQingSong

Approved by: Ding Yan Ping Dec 14th, 2018

Place, Date of Issue: Ningbo, Dec 14, 2018

Signature:

Name: Li Guirong

Position: General Manager





NINGBO GREETMED MEDICAL INSTRUMENTS CO., LTD.
