

# **Government of Pakistan**

# Pakistan Council of Scientific & Industrial Research

PCSIR- CAST METALS & FOUNDRY TECHNOLOGY CENTRE Punjab Small Industries Estate, Daska



CMFT/20/28451

Dated:27-07-2020

## **TEST REPORT**

In respect of

Centre / Division / Section

Sample Name

Your reference Letter No. / Date

Client's Sample Nos.

M/s Surgicon Sialkot Pakistan

Industrial Research & Development Section

Operating Scissor

Nil/08-07-2020

01

### Task

The CFC Center Sialkot was asked to provide the Chemical Analysis as per provided Sample.

### 1. Chemical Analysis Results

Element	Percentage	Element	Percentage
Carbon (C)	0.288	Silicon (Si)	0.344
Phosphorus (P)	0.027	Sulpher (S)	0.022
Manganese (Mn)	0.496	Nickel (Ni)	0.274
Chromium	12.675	Iron (Fe)	Balance

The above mentioned chemical Composition appears to be of AISI 420 J2

This analytical / test report is based solely on the particular samples (s) supplied by the party and should not be reproduced in part. Sampling has not been performed by the PCSIR-CM&FT and it does not accept the responsibility that the samples (s) supplied is/are truly representative samples (s) of any batch or stock or entire production. While the CM&FT agrees to take every reasonable precaution to ensure validity of its test results, it assumes no liability for the accuracy thereof beyond the amount of the fee charged for the analysis/test.

The party shall assume full responsibility for the ethical use of the results in the analysis/test reports and the laboratory shall be held free from any and all claims which may result from the use of such data by the party or other. After completion of the report the sample will be preserved for one month until negotiated otherwise

**Industrial Research & Development Section.** 

Counter signature

Sr. Instructor

PCSIR- Cast Mirals & Foundry Technology (CM) Centre Daska.

Engr. Ghulam Shabbir Sr.Engineer/Principal CM&FT Daska

Tel: 052-6625086 Fax: 052-6625089

E-mail: cmft@wol.net.pk & cmftdaska@yahoo.com