



Innovative Catheter Technology

Unique braid/coil construction for the lowest profile of any 0.014"/0.012" guidewire compatible catheter

Headway® Duo Microcatheter

Microcatheter with Hydrophilic Coating

1 per box; includes shaping mandrel and introducer sheath

Product Name	Product Code	ID Body/Tip (inches)	Tip Shape	Usable Length (cm)	Flexible Distal Length (cm)	OD Prox./Dist. (French)	Dead Space	Tip Markers
Headway Duo, 156cm	MC162156S	0.0165 / 0.0165	STR	156	9	2.1 / 1.6	0.34 ml	2
Headway Duo, 167cm	MC162167S	0.0165 / 0.013	STR	167	9	2.1 / 1.3	0.35 ml	1

Features

Hybrid Braid/Coil Design

Tight pitch coil provides lumen integrity and bending flexibility
Proximal braid provides firmness and a torqueable catheter body
Resists kinking and ovalization

Small OD, Large ID

156cm Headway® Duo microcatheter compatible with coils and $\leq 0.014"$ OD guidewires
0.0165" body ID microcatheter with low profile distal tip
167cm Headway® Duo microcatheter compatible with $\leq 0.014"/\leq 0.012"$ (proximal/distal OD) guidewires
0.0165" body ID microcatheter with 0.013" ID distal tapered tip

Superior Trackability

True 1:1 push/pull while tracking through tortuous anatomy

Liquid Embolic Compatibility

Compatible with PHIL™ Liquid Embolic, Onyx™ Liquid Embolic, nBCA and DMSO
High burst strength (700 psi)
Long length allows distal lesion access

Soft Distal Tip

Smooth, atraumatic tracking
Excellent tip "painting" ability



MicroVention, Inc.
Worldwide Headquarters
1311 Valencia Avenue
Tustin, CA 92780, USA
MicroVention UK Limited
MicroVention Europe, S.A.R.L.
MicroVention Deutschland GmbH
Web

PH +1.714.247.8000

PH +44 (0) 191 258 6777
PH +33 (1) 39 21 77 46
PH +49 211 210 798-0
microvention.com



MICROVENTION and Headway are registered trademarks of MicroVention, Inc. PHIL is a trademark of MicroVention, Inc. Onyx is a trademark of Covidien. • Scientific and clinical data related to this document are on file at MicroVention, Inc. Refer to Instructions for Use for additional information. ©2014 MicroVention, Inc. MM268(I) Rev.C 8/14