

ACON BIOTECH (HANGZHOU) CO.,LTD

CERTIFICATE OF ANALYSIS

PRODUCT NAME: On Call Plus Blood Glucose Test Strips(50)

PURCHASE NO.: 37740068B

BATCH NO.: 1691214

EXPIRATION DATE: 2022-07-31

QUANTITY: 15,000pcs

ANALYSIS		CRITERION	TEST RESULT
SYSTEM ACCURACY	Blood glucose conc. ≥ 75 mg/dL(4.2mmol/L)	The %Bias within $\pm 20\%$.	Complies
	Blood glucose conc. < 75 mg/dL(4.2mmol/L)	The Bias within ± 15 mg/dL(0.83mmol/L)	Complies
SYSTEM PRECISION	Blood glucose conc. ≥ 75 mg/dL(4.2mmol/L)	CV $< 7.5\%$	Complies
	Blood glucose conc. < 75 mg/dL(4.2mmol/L)	SD < 7.7 mg/dL(0.42mmol/L)	Complies

(Ninety-five percent (95 %) of the individual glucose results meet bias test result requirement)

Quality Department: Hedy zhang

Date: Sep. 23, 2020