



INCOTECH
Echipamente pentru detectia gazelor toxice si explozive

CERTIFICAT DE CALIBRARE

Beneficiar:	RHL PRO CONSULTING	Oras:	Bucuresti
C.U.I.	RO17375870	Nr.Reg.Com:	J40/5227/2005
Adresa beneficiar:	str. 9 Mai nr.4 bl.22A ap.14, sector 6		
Tel/fax:	021 221 37 05		
Nr. raport:	137		
Firma producatoare:	Lighthouse Worldwide Solutions		
Model aparat:	Handheld 3016		
Serie:	080502006		

Executat departament service INCOTECH S.R.L. conf. protocol calibrare MCD20050214V04 Lighthouse

		Conditii calibrare	
Data calibrare:	04 Februarie 2022	Temperatura:	23°C
Calibrarea expira la data:	04 Februarie 2023	Umiditate:	44%
Setari semnal (voltaj)			
Dimensiune particule: 0,3µm	Lot#231958	Val: Oct. 2023	Semnal canal 1 : 47 mV
Dimensiune particule: 0,5µm	Lot#229560	Val: Aug. 2023	Semnal canal 2 : 280 mV
Dimensiune particule: 0,7µm	Lot#232671	Val: Oct. 2023	Semnal canal 3 : 608 mV
Dimensiune particule: 1µm	Lot#231965	Val: Oct. 2023	Semnal canal 3 : 819 mV
Dimensiune particule: 2,0µm	Lot#231222	Val: Sep. 2023	Semnal canal 5 : 1264 mV
Dimensiune particule: 5µm	Lot#238352	Val: Mar. 2023	Semnal canal 5 : 3190 mV
Masuratori			
Debit:		2,82 l/min	Limita: ±5% (nom.2,83 l/min)
False count rate:	Nr. part. Obs in 5 min	0 counts	Nu se specifica in ISO21501-4
Laser:		48 mV	Limita: ±15% (ref. 47 mV)

Echipamente utilizate pentru efectuarea calibrarii

Model	Serie	Certificat calibrare
Analizor spectru MCA8000D	1116	E-596/2021/25.06.2021
Osciloscop Tektronix TDS2012C	C019462	15.102.24.05.21/24.05.2021
Calibrator debit TSI 4043	4043 1216 012	15.101.20.07.21/20.07.2021
Multimetru digital Voltcraft	1120036754	05.101.24.05.21/24.05.2021
NP Solair Transfer Standard	160699001	LH78547/09.11.2021

Nume: Pirva Olga

Semnatura:

Service Incotech

INCOTECH
S.R.L.

Service Incotech