



## Test Verification of Conformity

On the basis of the referenced test report(s), sample(s) of the below product have been found to comply with the harmonized standards and Directives listed on this verification at the time the tests were carried out. Other standards and Directives may be relevant to the product.

Once all product relevant  mark directives are verified in compliance, the manufacturer may indicate compliance by signing a Declaration of Conformity themselves and applying the mark to product identical to the test sample(s) if the product complies with all relevant CE mark Directives requirements.

|   |   |
|---|---|
| <b>Applicant Name &amp; Address</b>                           | : Zhongke Meiling Cryogenics Limited Company<br>No 1862 Zishi Road Economic Technological Development<br>Zone Hefei 230601 China  |
| <b>Manufacturing site Name &amp; Address</b>                  | : Same as applicant   |
| <b>Product(s) Tested</b>                                      | : Pharmacy refrigerator   |
| <b>Ratings and principal characteristics</b>                  | : 220-230V~, 50Hz, Class I, Climate Class: 4, Blowing Gas: CP<br>YC-315L: R600a/20g, 1.35A<br>YC-365L: R600a/26g, 1.4A<br>YC-395L: R290/45g, 1.8A                                 |
| <b>Model(s)</b>   | : YC-315L, YC-365L, YC-395L   |
| <b>Brand name</b>   | :    |
| <b>Relevant Standard(s) / Specification(s) / Directive(s)</b> | : IEC 60601-1 Edition 3.1 (2012)<br>EN 60601-1:2006 + A1:2013 + A12:2014<br>IEC 60601-1-2: 2014<br>EN 60601-1-2:2015  |
| <b>Verification Issuing Office Name &amp; Address</b>         | : UL-CCIC Company Limited<br>Guangzhou Branch Electronic Building, Parage Electronic<br>Industrial Park, No. 8 Nanyun Er Road, Guangzhou Science<br>Park, Guangzhou 510670, China |
| <b>Report Number(s)</b>                                       | : E507240-D1005-1/A0/C0-Informative<br>4789205811 2.1   |

**NOTE 1: This verification is part of the full test report(s) and should be read in conjunction with it.**

This Verification is for the exclusive use of UL-CCIC Client and is provided pursuant to the agreement between UL-CCIC and its Client. IUL-CCIC responsibility and liability are limited to the terms and conditions of the agreement. UL-CCIC assumes no liability to any party, other than to the Client in accordance with the agreement, for any loss, expense or damage occasioned by the use of this Verification. Only the Client is authorized to copy or distribute this Verification. Any use of the UL-CCIC name or one of its marks for the sale or advertisement of the tested material, product or service must first be approved in writing by UL-CCIC. The observations and test results referenced from this Verification are relevant only to the sample tested. This Verification by itself does not imply that the material, product, or service is or has ever been under an IUL-CCIC certification program.



**Signature**

**Name:** Paul Zhang  
**Position:** Reviewer  
**Date:** 16 July., 2020