

NÜVE SANAYİ MALZEMELERİ İMALAT VE TİCARET ANONİM ŞİRKETİ

Saracalar Mahallesi Saracalar Kümeevleri No:4/2 Akyurt Ankara, Turkey

**Design, Manufacture, Sales And After Sales Services Of Laboratory
And Sterilization Equipments**

with a scope of

ISO 9001:2015


Has established a quality management system in accordance
with international standard.

"Following elements of the standard are excluded"
"None"

Certificate No	: M 8215
Initial Certification Date	: 28 April 2010
Certification Date	: 11 March 2022
Expiration Date	: 10 March 2025

Kiwa Belgelendirme Hizmetleri A.Ş.
İTOSB 9. Cadde No. 15 Tepeören Tuzla
İstanbul / Turkey


Tel: + 90 216 593 25 75
Faks: + 90 216 593 25 74
info@kiwa.com.tr
www.kiwa.com.tr



General Manager

Certificate is valid till expiration date,
subject to successful completion of
periodical surveillance audits.
Please contact above numbers for
detailed information.





CERTIFICATE

NÜVE SANAYİ MALZEMELERİ İMALAT VE TİCARET ANONİM ŞİRKETİ

Saracalar Mahallesi Saracalar Kümeevleri No:4/2 Akyurt Ankara, Turkey

**Design, Manufacture, Sales And After Sales Services Of Laboratory
And Sterilization Equipments**

with a scope of

EN ISO 13485:2016

Has established a management system in accordance
with international Medical Devices Quality Management System Standard

"Following elements of the standard are excluded"


"7.5.5" "7.5.7" "7.5.9.2"

Certificate No	: M 8216
Initial Certification Date	: 28 April 2010
Certification Date	: 11 March 2022
Expiration Date	: 10 March 2025

Kiwa Belgelendirme Hizmetleri A.Ş.
İTOSB 9. Cadde No. 15 Tepeören Tuzla
İstanbul / Turkey

Tel: + 90 216 593 25 75
Faks: + 90 216 593 25 74
info@kiwa.com.tr
www.kiwa.com.tr

General Manager



TÜRKAK BDS NO
YS-FE6F-F002

Certificate is valid till expiration date,
subject to successful completion of
periodical surveillance audits.
Please contact above numbers for
detailed information.



NF200

Bench Top Centrifuge



- Compact design for separation of blood samples, urine particle sedimentation and other routine applications in emergency and small laboratories
- N-Prime™ programmable microprocessor control system for accurate and reliable operations
- Easy programming with one button, just turn and push
- Two large LED displays for speed/RCF and time
- Possibility of programming RCF
- Pulse key for short centrifugation requirements
- Reduced temperature increase in the chamber by means of continuous air cooling through air channel in the lid
- Quiet brushless induction drive
- Safety ensured by steel body, lid and chamber, lid locking system and motor over temperature protection
- Delivered with 12x15 ml. angle rotor

N-prime

TECHNICAL SPECIFICATIONS

Maximum Speed	6.000 rpm
Maximum RCF	4.065xg.
Tube Capacity	12x15 ml.
Control System	N-Prime™ Programmable Microprocessor Control
Speed Set Range	1.000-5.000 rpm
Speed Set Step	10 rpm
Timer Set Range	1-99 Minutes and Hold Position
Timer Set Step	1 minute
Motor	Induction Motor
Body, Lid and Chamber	Epoxy-Polyester Powder Coated Steel
Power Supply	230 V / 50-60 Hz (115 V / 60 Hz)
Power Consumption	150 W
External Dimensions (WxDxH) mm.	280x360x260
Packing Dimensions (WxDxH) mm.	350x420x480
Net / Packed Weight, kg.	12 / 16

ROTOR

Rotor Type	Capacity	Max. Tube Dia. (mm)	Radius (mm)	Max. Speed	Max. RCF
Angle	12x15 ml.	17	101	6.000 rpm	4.065xg.

ADAPTORS

Code	Tube Type	Max. Tube Dia. (mm)
G 03 057	1.5/2 ml tubes	11
G 03 014	7 ml vacuumed/ non vacuumed tubes	13
G 03 015	5 ml vacuumed/ non vacuumed tubes	13



NÜVE SANAYİ MALZEMELERİ İMALAT VE TİCARET A.Ş.

Saracalar Mah. Saracalar Kümeevleri No: 4/2 Akyurt 06750 ANKARA / TURKEY t. +90 312 399 28 30 f. +90 312 399 21 97
nuve.com.tr sales@nuve.com.tr

ISO 9001: 2008
ISO 13485: 2003

