

# EU Type Examination Certificate

This is to certify that:

WEINI TECHNOLOGY DEVELOPMENT CO., LTD  
No2 Xingye Road, Xicheng Industrial Zone  
Renhe Town, Baiyun District  
Guangzhou  
Guangdong  
510470  
China

Holds Certificate Number:

CE 709415

In respect of:

**Filtering half masks to EN 149:2001+A1:2009.  
Various models - see continuation sheet for details.**


on the basis that BSI carried out the relevant Type Examination procedures under the requirements with the Regulation (EU) 2016/425 of the European Parliament and Council relating to Personal Protective Equipment Regulation (PPE) Annex V (Module B) and meets the relevant health and safety requirements specified in Annex II

For and on behalf of BSI, a Notified Body for the above Regulation (Notified Body Number 2797):

Previous Notified Body: BSI 0086

First Issued: 2019-09-10

Latest Issue: 2019-10-24



Drs. Dave Hagenaaers, Managing Director

Effective Date: 2019-10-24

Expiry Date: 2024-09-10

Page: 1 of 12



...making excellence a habit.™

# EU Type Examination Certificate

No. CE 709415

## Product Specification – Flat Fold Particulate Half Masks

**Technical specification:** EN 149:2001+A1:2009 Respiratory protective devices – Filtering half masks to protect against particles  
Classification – Single shift NR

**Description:** The particulate respirators are designed to protect against solid and non-volatile liquid particles. Half masks are horizontal flat fold design, comprising of PP iron wire nose bridge, a sponge nose pad and are held on the face by a pair of natural rubber headbands. All models are single shift devices (denoted by the classification symbol NR). Some models incorporate a single, PP plastic exhalation valve, see details on models denoted by letter V.

**FFP Class** **FFP2** (without clogging option D)

**Models:** E-820  
E-820V (mask with exhalation valve fitted)  
E-820C (mask with activated charcoal)  
E-820VC (mask with exhalation valve and activated charcoal)

## Product Specification – Flat Fold Particulate Half Masks

**Technical specification:** EN 149:2001+A1:2009 Respiratory protective devices – Filtering half masks to protect against particles  
Classification – Single shift NR

**Description:** The particulate respirator is designed to protect against solid and non-volatile liquid particles. Half mask is a horizontal flat fold design, comprising of PP plastic wire nose bridge, a sponge nose pad and is held on the face by a pair of elastic headbands. Model is a single shift device (denoted by the classification symbol NR), and incorporates a single, PP plastic exhalation valve.

**FFP Class** **FFP3** (without clogging option D)

**Model:** E-830V (mask with exhalation valve fitted)

First Issued: 2019-09-10

Latest Issue: 2019-10-24

Effective Date: 2019-10-24

Expiry Date: 2024-09-10

Page: 8 of 12

This certificate has been issued by and remains the property of BSI Group The Netherlands B.V., John M. Keynesplein 9, 1066 EP Amsterdam, The Netherlands and should be returned immediately upon request.  
To check its validity telephone +31 20 3460780. An electronic certificate can be authenticated [online](#).

BSI Group The Netherlands B.V., registered in the Netherlands under number 33264284, at John M. Keynesplein 9, 1066 EP Amsterdam, The Netherlands  
A member of BSI Group of Companies.