

## CERTIFICATE OF ANALYSES

Amphetamine(AMP),Ecstasy(MDMA),Morphine(MOR),Marijuana (THC), Cocaine  
(COC),Benzodiazepines(BZO),Barbiturates(BAR), Methamphetamine  
(MET),Methadone(MTD),Phencyclidine(PCP),Tricyclic  
Antidepressant(TCA),Opiate(OPI)  
Multi Panel Screen Test  
Batch Number: MDC23022701

**Description:** Multi Panel Screen Test

**Appearance:** Well sealed in foil pouch

<u>TEST SAMPLE</u>	<u>RESULTS</u>	<u>COMMENTS</u>
Negative Urine	(-) Negative	Pass
-50% cutoff AMP	(-) Negative	Pass
+50% cutoff AMP	(+) Positive	Pass
3X cutoff AMP	(+) Positive	Pass
-50% cutoff MDMA	(-) Negative	Pass
+50% cutoff MDMA	(+) Positive	Pass
3X cutoff MDMA	(+) Positive	Pass
-50% cutoff MOR	(-) Negative	Pass
+50% cutoff MOR	(+) Positive	Pass
3X cutoff MOR	(+) Positive	Pass
-50% cutoff THC	(-) Negative	Pass
+50% cutoff THC	(+) Positive	Pass
3X cutoff THC	(+) Positive	Pass
-50% cutoff COC	(+) Positive	Pass
+50% cutoff COC	(+) Positive	Pass
3X cutoff COC	(+) Positive	Pass
-50% cutoff BZO	(-) Negative	Pass
+50% cutoff BZO	(+) Positive	Pass
3X cutoff BZO	(+) Positive	Pass
-50% cutoff BAR	(-) Negative	Pass
+50% cutoff BAR	(+) Positive	Pass
3X cutoff BAR	(+) Positive	Pass
-50% cutoff MET	(-) Negative	Pass
+50% cutoff MET	(+) Positive	Pass
3X cutoff MET	(+) Positive	Pass
-50% cutoff MTD	(-) Negative	Pass
+50% cutoff MTD	(+) Positive	Pass
3X cutoff MTD	(+) Positive	Pass
-50% cutoff PCP	(-) Negative	Pass
+50% cutoff PCP	(+) Positive	Pass
3X cutoff PCP	(+) Positive	Pass

---

-50% cutoff	TCA	(-) Negative	Pass
+50% cutoff	TCA	(+) Positive	Pass
3X cutoff	TCA	(+) Positive	Pass
-50% cutoff	OPI	(-) Negative	Pass
+50% cutoff	OPI	(+) Positive	Pass
3X cutoff	OPI	(+) Positive	Pass

**Other Tests:**

Strip Length	80mm	Pass
Strip Width	4mm	Pass
Appearance of C Line	20-40 seconds	Pass
Background	Clear after 5 minutes	Pass
Visual Quality	Component firmly assembled, no flaw, well sealed.	Pass

This product was tested and approved in accordance with Quality Control testing performance criteria.

**Tested By:**

Chenyang Zheng

**Quality Control Approval:**

Wuxiang Xu

**Date Tested:** Feb.28, 2023

**Man. Date:** 2023.02.27

**EXP. Date:** 02-2025

