

Certificate of Analysis for
Povidone-iodine prep pad

Description: Povidone-iodine Prep pad, quantity200,000pcs Lot No.: 20240410

FINAL INSPECTION RESULT: Samples qty:150pcs

Test standard No.: SXZZ20182640193

No.	Inspection items	Standard requirements	Defects quantity	result	Testing Instruments No.
1	Packing	According to Contract.	0	OK!	By visual Inspection
2.	Printing quality	Nice print, very clear	0	OK!	By visual Inspection
3	Size of bag, non-woven.	$50\pm 2\text{mm}\times 50\pm 2\text{mm}$ $50\pm 2\text{g/M}^2$	0	OK!	Digital scale Vernier caliper
4.	Weight of 10% Povidone-iodine Prep pad	$0.55\pm 0.10\text{g}$	0	OK!	Digital scale
5	Cutting of sides	No obviously burr	0	OK	By visual Inspection

Final Result: acceptable

Inspection: Lu jiao

Approved by: Ding Yan Ping

May 16th, 2024

宁波柯泰医疗器械有限公司
NINGBO GREETMED MEDICAL INSTRUMENTS CO.,LTD.

李桂蓉