



Manual Sealer

Model: CT-FK-800

Technical Parameters:

Voltage: 220V

Power: 400W

Indentation Width: 10 mm

Sealing Width: 300mm

Outer Size: 370*300*190mm


Weight: 5kg

Main feature:

Automatically constant temperature, easy to operate

Adjustable temperature, Automatic self-safety for over temperature

Built-in Cutting Knife





2020/1/9 14:39

CERTIFICATE

Anbotek

Product Safety

CERTIFICATE

Of Conformity

Registration No.: AT17117858E

Report No.: 201711858E

Applicant : Tianjin Chaotuo Medical Equipment Co.,Ltd.

Tianjin Huayuan Industrial Park Road,
Zhu Yuan 6# 2-1-501, Tianjin,
China

Product : Sealer; Cutter

Identification : **Model No.:** CT-FK-820; CT-FK-830; CT-FK-800
CT-QF-320; CT-QF-420;
CT-QF-310; CT-QG-100

Trade mark : Chaotuo

Test standards : EN61000-6-1:2007, EN ISO12100:2010,
EN60204-1:2006+A1:2009+AC2010,
EN61000-3-2:2014, EN61000-3-3:2013

Based upon a review of worksheets and the Technical Construction File,
the apparatus is confirmed to meet the requirements of the above standards
and hence fulfils the requirements of bellow:

2006/42/EC Machinery Directive

2014/35/EU Low Voltage Directive

2014/30/EU Electromagnetic Compatibility Directive



Certified by

Tom Chen
Tom Chen
Manager

Date: Nov.25,2017

Shenzhen Anbotek Compliance Laboratory Limited

1/F., Building 1, SEC Industrial park, No.0409 Qianhai Road,
Nanshan District, Shenzhen, Guangdong, China
[Http://www.anbotek.com](http://www.anbotek.com)

Tel: (86) 755-260665533
Fax: (86) 755-26014772
Email: service@anbotek.com