


Test Report

– Eppendorf PCR clean –

Safe-Lock Tubes 2.0 mL



1000 EPPENDORF TUBES®
PCR clean
centrifugation up to 25 000 x g

LOT M213150N

2028-08-28

Order no.: 0030 123.344

Eppendorf SE
22331 Hamburg · Germany
Email: eppendorf@eppendorf.com

• Human DNA

Method: real-time PCR acc. to SOP 99880/V2
Limit: < 2 pg; less than one human cell

• DNase

Method: DNA Digestion acc. to SOP 100570/V1
Limit: Not detectable (LOD: 1.0×10^{-6} Kunitz units)

• PCR inhibition

Method: real-time PCR acc. to SOP 100559/V1
Limit: Less than 10 targets amplifiable

• RNase

Method: RNA Digestion acc. to SOP 100571/V1
Limit: Not detectable (LOD: 1.0×10^{-9} Kunitz units)

Customer: Eppendorf SE, Barkhausenweg 1, D-22339 Hamburg

Date of sample receipt / examination: 25.08.23

Testing of the above described lot showed conformity within the limits of detection.

23.08.23

Date of release

Epp 2308-7684

Registration number

i.v. P. Weigel

Dipl.-Biol. Thomas Weigel (Head)

The test results refer exclusively to the items tested as arrived. The test report must not be copied partially without the approval of LADR GmbH.

page 1 of 1